

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012221\
 Data File : VX020707.D
 Acq On : 22 Jan 2021 11:39
 Operator : JC/SP
 Sample : VX0122WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0122WBL01

Quant Time: Jan 23 01:40:11 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 19 14:09:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	217903	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.824	114	383563	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	368914	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	160798	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.025	65	160896	54.00	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 108.00%	
35) Dibromofluoromethane	5.458	113	133862	50.65	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 101.30%	
50) Toluene-d8	8.689	98	478314	50.20	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 100.40%	
62) 4-Bromofluorobenzene	11.122	95	166903	46.34	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 92.68%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012221\
Data File : VX020707.D
Acq On : 22 Jan 2021 11:39
Operator : JC/SP
Sample : VX0122WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0122WBL01

Quant Time: Jan 23 01:40:11 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011921W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 19 14:09:08 2021
Response via : Initial Calibration

