

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012224\
 Data File : VX039888.D
 Acq On : 22 Jan 2024 16:53
 Operator : JC/MD
 Sample : P1186-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-191-GW-358-360

Quant Time: Jan 23 00:19:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	81043	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	147465	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	140655	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	66824	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	64180	47.275	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.540%
35) Dibromofluoromethane	5.385	113	48781	49.113	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.220%
50) Toluene-d8	8.647	98	181928	48.565	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.140%
62) 4-Bromofluorobenzene	11.079	95	69313	47.316	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.640%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.947	59	1626	4.601	ug/l	98
16) Acetone	2.374	43	2175	2.849	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012224\
Data File : VX039888.D
Acq On : 22 Jan 2024 16:53
Operator : JC/MD
Sample : P1186-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-191-GW-358-360

Quant Time: Jan 23 00:19:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

