

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012224\
 Data File : VX039889.D
 Acq On : 22 Jan 2024 17:16
 Operator : JC/MD
 Sample : P1215-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-195-GW-360-362

Quant Time: Jan 23 00:19:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	77854	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	144782	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	138745	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	65910	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	62457	47.890	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.780%
35) Dibromofluoromethane	5.379	113	48054	49.278	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.560%
50) Toluene-d8	8.647	98	178972	48.661	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.320%
62) 4-Bromofluorobenzene	11.079	95	69069	48.023	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.040%
Target Compounds						
					Qvalue	
3) Chloromethane	1.294	50	1194	1.038	ug/l	96
16) Acetone	2.373	43	5820	7.935	ug/l	99
25) 2-Butanone	4.587	43	1007	0.955	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012224\
Data File : VX039889.D
Acq On : 22 Jan 2024 17:16
Operator : JC/MD
Sample : P1215-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-195-GW-360-362

Quant Time: Jan 23 00:19:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

