

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012224\
 Data File : VX039893.D
 Acq On : 22 Jan 2024 18:46
 Operator : JC/MD
 Sample : P1215-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-195-GW-380-382

Quant Time: Jan 23 00:21:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	78772	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	145735	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	141617	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	69382	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63276	47.953	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.900%
35) Dibromofluoromethane	5.385	113	49372	50.298	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.600%
50) Toluene-d8	8.646	98	184064	49.719	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.440%
62) 4-Bromofluorobenzene	11.079	95	71635	49.482	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.960%
Target Compounds						
16) Acetone	2.373	43	2226	3.000	ug/l	Qvalue # 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012224\
Data File : VX039893.D
Acq On : 22 Jan 2024 18:46
Operator : JC/MD
Sample : P1215-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-195-GW-380-382

Quant Time: Jan 23 00:21:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

