

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX012224\
 Data File : VX039894.D
 Acq On : 22 Jan 2024 19:09
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 23 00:21:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	69	0.00
2 T	Dichlorodifluoromethane	0.647	0.594	8.2	62	0.00
3 P	Chloromethane	0.739	0.653	11.6	63	0.00
4 C	Vinyl Chloride	0.697	0.668	4.2#	68	0.00
5 T	Bromomethane	0.398	0.428	-7.5	76	0.00
6 T	Chloroethane	0.396	0.428	-8.1	72	0.00
7 T	Trichlorofluoromethane	1.006	1.107	-10.0	79	0.00
8 T	Diethyl Ether	0.372	0.394	-5.9	76	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.624	0.674	-8.0	77	0.00
10 T	Methyl Iodide	0.658	0.720	-9.4	72	0.00
11 T	Tert butyl alcohol	0.218	0.240	-10.1	77	0.00
12 CM	1,1-Dichloroethene	0.602	0.643	-6.8#	75	0.00
13 T	Acrolein	0.147	0.150	-2.0	77	0.00
14 T	Allyl chloride	1.184	1.203	-1.6	71	0.00
15 T	Acrylonitrile	0.450	0.480	-6.7	74	0.00
16 T	Acetone	0.471	0.433	8.1	63	0.00
17 T	Carbon Disulfide	1.730	1.747	-1.0	72	0.00
18 T	Methyl Acetate	0.985	1.127	-14.4	80	0.00
19 T	Methyl tert-butyl Ether	2.136	2.299	-7.6	76	0.00
20 T	Methylene Chloride	0.713	0.743	-4.2	76	0.00
21 T	trans-1,2-Dichloroethene	0.633	0.692	-9.3	78	0.00
22 T	Diisopropyl ether	2.260	2.408	-6.5	76	0.00
23 T	Vinyl Acetate	2.101	2.272	-8.1	74	0.00
24 P	1,1-Dichloroethane	1.243	1.344	-8.1	75	0.00
25 T	2-Butanone	0.677	0.699	-3.2	71	0.00
26 T	2,2-Dichloropropane	0.998	0.956	4.2	67	0.00
27 T	cis-1,2-Dichloroethene	0.726	0.808	-11.3	77	0.00
28 T	Bromochloromethane	0.564	0.582	-3.2	72	0.00
29 T	Tetrahydrofuran	0.436	0.462	-6.0	74	0.00
30 C	Chloroform	1.228	1.412	-15.0#	81	0.00
31 T	Cyclohexane	1.105	1.123	-1.6	71	0.00
32 T	1,1,1-Trichloroethane	1.051	1.162	-10.6	76	0.00
33 S	1,2-Dichloroethane-d4	0.838	0.841	-0.4	73	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	69	0.00
35 S	Dibromofluoromethane	0.337	0.354	-5.0	75	0.00
36 T	1,1-Dichloropropene	0.524	0.554	-5.7	76	0.00
37 T	Ethyl Acetate	0.699	0.730	-4.4	75	0.00
38 T	Carbon Tetrachloride	0.480	0.544	-13.3	78	0.00
39 T	Methylcyclohexane	0.629	0.640	-1.7	71	0.00
40 TM	Benzene	1.484	1.615	-8.8	76	0.00
41 T	Methacrylonitrile	0.342	0.381	-11.4	73	0.00
42 TM	1,2-Dichloroethane	0.564	0.647	-14.7	79	0.00
43 T	Isopropyl Acetate	1.026	1.132	-10.3	74	0.00
44 TM	Trichloroethene	0.364	0.405	-11.3	77	0.00
45 C	1,2-Dichloropropane	0.398	0.432	-8.5#	77	0.00
46 T	Dibromomethane	0.275	0.314	-14.2	79	0.00
47 T	Bromodichloromethane	0.509	0.600	-17.9	80	0.00
48 T	Methyl methacrylate	0.504	0.552	-9.5	75	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012224\
 Data File : VX039894.D
 Acq On : 22 Jan 2024 19:09
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 23 00:21:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.010	0.010	0.0	72	0.00
50 S	Toluene-d8	1.270	1.291	-1.7	73	0.00
51 T	4-Methyl-2-Pentanone	0.707	0.783	-10.7	76	0.00
52 CM	Toluene	0.887	1.003	-13.1#	77	0.00
53 T	t-1,3-Dichloropropene	0.540	0.621	-15.0	76	0.00
54 T	cis-1,3-Dichloropropene	0.594	0.673	-13.3	76	0.00
55 T	1,1,2-Trichloroethane	0.377	0.418	-10.9	77	0.00
56 T	Ethyl methacrylate	0.611	0.687	-12.4	76	0.00
57 T	1,3-Dichloropropane	0.640	0.718	-12.2	78	0.00
58 T	2-Chloroethyl Vinyl ether	0.289	0.295	-2.1	68	0.00
59 T	2-Hexanone	0.559	0.632	-13.1	75	0.00
60 T	Dibromochloromethane	0.356	0.432	-21.3	81	0.00
61 T	1,2-Dibromoethane	0.391	0.461	-17.9	80	0.00
62 S	4-Bromofluorobenzene	0.497	0.538	-8.2	77	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	72	0.00
64 T	Tetrachloroethene	0.325	0.361	-11.1	85	0.00
65 PM	Chlorobenzene	1.041	1.126	-8.2	79	0.00
66 T	1,1,1,2-Tetrachloroethane	0.355	0.408	-14.9	80	0.00
67 C	Ethyl Benzene	1.933	2.107	-9.0#	78	0.00
68 T	m/p-Xylenes	0.711	0.778	-9.4	79	0.00
69 T	o-Xylene	0.704	0.757	-7.5	78	0.00
70 T	Styrene	1.150	1.283	-11.6	80	0.00
71 P	Bromoform	0.266	0.322	-21.1	81	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	79	0.00
73 T	Isopropylbenzene	3.981	3.973	0.2	80	0.00
74 T	N-amyl acetate	2.075	2.090	-0.7	76	0.00
75 P	1,1,2,2-Tetrachloroethane	1.520	1.492	1.8	79	0.00
76 T	1,2,3-Trichloropropane	1.296	1.273	1.8	79	0.00
77 T	Bromobenzene	0.886	0.880	0.7	81	0.00
78 T	n-propylbenzene	4.818	4.908	-1.9	80	0.00
79 T	2-Chlorotoluene	2.978	2.880	3.3	80	0.00
80 T	1,3,5-Trimethylbenzene	3.381	3.437	-1.7	81	0.00
81 T	trans-1,4-Dichloro-2-butene	0.427	0.397	7.0	71	0.00
82 T	4-Chlorotoluene	3.378	3.402	-0.7	82	0.00
83 T	tert-Butylbenzene	3.228	3.269	-1.3	81	0.00
84 T	1,2,4-Trimethylbenzene	3.373	3.440	-2.0	80	0.00
85 T	sec-Butylbenzene	4.190	4.362	-4.1	82	0.00
86 T	p-Isopropyltoluene	3.349	3.555	-6.2	82	0.00
87 T	1,3-Dichlorobenzene	1.699	1.718	-1.1	82	0.00
88 T	1,4-Dichlorobenzene	1.755	1.772	-1.0	83	0.00
89 T	n-Butylbenzene	3.234	3.440	-6.4	82	0.00
90 T	Hexachloroethane	0.565	0.608	-7.6	81	0.00
91 T	1,2-Dichlorobenzene	1.680	1.723	-2.6	85	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.372	0.394	-5.9	81	0.00
93 T	1,2,4-Trichlorobenzene	1.026	1.069	-4.2	84	0.00
94 T	Hexachlorobutadiene	0.427	0.442	-3.5	82	0.00
95 T	Naphthalene	3.779	3.738	1.1	79	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012224\
Data File : VX039894.D
Acq On : 22 Jan 2024 19:09
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jan 23 00:21:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.036	1.038	-0.2	82	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6