

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012323\
 Data File : VX033843.D
 Acq On : 23 Jan 2023 12:51
 Operator : JC/MD
 Sample : PB150369ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB150369ZHE#02

Quant Time: Jan 24 01:43:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	160866	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	237873	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	206439	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96421	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	78367	46.268	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.540%
35) Dibromofluoromethane	5.385	113	73153	49.526	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.060%
50) Toluene-d8	8.647	98	271179	50.389	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.780%
62) 4-Bromofluorobenzene	11.079	95	90062	45.073	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.140%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	15540	13.528	ug/l	92
37) Ethyl Acetate	4.715	43	22451	8.545	ug/l	100
43) Isopropyl Acetate	6.336	43	94643	27.810	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012323\
Data File : VX033843.D
Acq On : 23 Jan 2023 12:51
Operator : JC/MD
Sample : PB150369ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB150369ZHE#02

Quant Time: Jan 24 01:43:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 11 14:48:42 2023
Response via : Initial Calibration

