

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012323\
 Data File : VX033854.D
 Acq On : 23 Jan 2023 17:06
 Operator : JC/MD
 Sample : 01251-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TCLP-1

Quant Time: Jan 24 00:44:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	165046	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	242450	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	212330	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96104	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	79700	45.863	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.720%
35) Dibromofluoromethane	5.385	113	74595	49.549	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.100%
50) Toluene-d8	8.647	98	276130	50.340	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.680%
62) 4-Bromofluorobenzene	11.079	95	91624	44.989	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.980%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	15235	12.347	ug/l	94
18) Methyl Acetate	2.709	43	2851	1.176	ug/l	99
20) Methylene Chloride	2.788	84	11048	5.060	ug/l	94
37) Ethyl Acetate	4.721	43	20158	7.167	ug/l	99
43) Isopropyl Acetate	6.336	43	88610	25.546	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012323\
Data File : VX033854.D
Acq On : 23 Jan 2023 17:06
Operator : JC/MD
Sample : 01251-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TCPL-1

Quant Time: Jan 24 00:44:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 11 14:48:42 2023
Response via : Initial Calibration

