

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012324\
 Data File : VX039897.D
 Acq On : 23 Jan 2024 09:44
 Operator : JC/MD
 Sample : VX0123MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0123MBL01

Quant Time: Jan 24 02:27:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	78034	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	138210	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	133782	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	63783	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	62163	47.555	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.100%
35) Dibromofluoromethane	5.379	113	45742	49.137	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.280%
50) Toluene-d8	8.647	98	172483	49.127	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.260%
62) 4-Bromofluorobenzene	11.079	95	66552	48.474	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012324\
Data File : VX039897.D
Acq On : 23 Jan 2024 09:44
Operator : JC/MD
Sample : VX0123MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0123MBL01

Quant Time: Jan 24 02:27:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

