

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012324\
 Data File : VX039901.D
 Acq On : 23 Jan 2024 11:47
 Operator : JC/MD
 Sample : P1196-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE137-SP302-B-20240118

Quant Time: Jan 24 02:29:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	77976	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	141223	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	135197	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	64917	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	61436	47.034	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.060%
35) Dibromofluoromethane	5.385	113	46786	49.186	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.380%
50) Toluene-d8	8.647	98	171639	47.844	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.680%
62) 4-Bromofluorobenzene	11.079	95	66601	47.474	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.940%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.331	101	5290	5.440	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012324\
Data File : VX039901.D
Acq On : 23 Jan 2024 11:47
Operator : JC/MD
Sample : P1196-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE137-SP302-B-20240118

Quant Time: Jan 24 02:29:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

