

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012324\
 Data File : VX039907.D
 Acq On : 23 Jan 2024 14:02
 Operator : JC/MD
 Sample : P1222-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-191-GW-398-400

Quant Time: Jan 24 02:32:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	77319	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	143842	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	136959	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	67362	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	61462	47.453	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.900%
35) Dibromofluoromethane	5.379	113	47492	49.020	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.040%
50) Toluene-d8	8.647	98	176714	48.361	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.720%
62) 4-Bromofluorobenzene	11.079	95	67878	47.504	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.000%
Target Compounds						
16) Acetone	2.374	43	1803	2.475	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012324\
Data File : VX039907.D
Acq On : 23 Jan 2024 14:02
Operator : JC/MD
Sample : P1222-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-191-GW-398-400

Quant Time: Jan 24 02:32:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

