

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012324\
 Data File : VX039910.D
 Acq On : 23 Jan 2024 15:10
 Operator : JC/MD
 Sample : P1223-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-195-GW-440-442

Quant Time: Jan 24 02:33:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	78984	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	143980	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	138682	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	64938	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	62019	46.874	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.740%
35) Dibromofluoromethane	5.385	113	47387	48.864	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.720%
50) Toluene-d8	8.647	98	178030	48.675	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.340%
62) 4-Bromofluorobenzene	11.079	95	67677	47.318	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.640%
Target Compounds						
17) Carbon Disulfide	2.514	76	826	0.302	ug/l	# 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012324\
Data File : VX039910.D
Acq On : 23 Jan 2024 15:10
Operator : JC/MD
Sample : P1223-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-195-GW-440-442

Quant Time: Jan 24 02:33:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

