

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012324\
 Data File : VX039912.D
 Acq On : 23 Jan 2024 15:56
 Operator : JC/MD
 Sample : P1223-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-195-GW-460-462

Quant Time: Jan 24 02:33:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	76683	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	141498	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	135966	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	66656	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	61755	48.075	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.160%
35) Dibromofluoromethane	5.385	113	46576	48.870	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.740%
50) Toluene-d8	8.647	98	174863	48.648	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.300%
62) 4-Bromofluorobenzene	11.079	95	66455	47.278	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.560%
Target Compounds						
16) Acetone	2.374	43	3607	4.993	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012324\
Data File : VX039912.D
Acq On : 23 Jan 2024 15:56
Operator : JC/MD
Sample : P1223-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-195-GW-460-462

Quant Time: Jan 24 02:33:40 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

