

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012324\
 Data File : VX039916.D
 Acq On : 23 Jan 2024 17:49
 Operator : JC/MD
 Sample : P1197-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Jan 24 02:35:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	80098	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	143845	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	137814	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	63562	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63120	47.043	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.080%
35) Dibromofluoromethane	5.385	113	49092	50.670	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.340%
50) Toluene-d8	8.647	98	178780	48.926	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.860%
62) 4-Bromofluorobenzene	11.079	95	66550	46.573	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012324\
Data File : VX039916.D
Acq On : 23 Jan 2024 17:49
Operator : JC/MD
Sample : P1197-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Jan 24 02:35:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

