

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX012419\
 Data File : VX007206.D
 Acq On : 24 Jan 2019 16:22
 Operator : JC/SP
 Sample : K1016-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TR-05-RO-3-012319

Quant Time: Jan 25 00:04:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	546783	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	843827	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	801424	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	383240	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	298093	52.17	ug/l	0.00
Spiked Amount			Recovery	=	104.34%	
35) Dibromofluoromethane	5.51	113	262998	48.67	ug/l	0.00
Spiked Amount			Recovery	=	97.34%	
50) Toluene-d8	8.71	98	1014832	51.31	ug/l	0.00
Spiked Amount			Recovery	=	102.62%	
62) 4-Bromofluorobenzene	11.14	95	354761	51.45	ug/l	0.00
Spiked Amount			Recovery	=	102.90%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Chloromethane	1.33	50	1266	0.242	ug/l #	83
5) Bromomethane	1.65	94	2386	0.397	ug/l #	54
11) Tert butyl alcohol	3.07	59	1543	1.145	ug/l #	1
13) Acrolein	2.30	56	1109	4.968	ug/l #	77
16) Acetone	2.45	43	20346	7.202	ug/l #	85
18) Methyl Acetate	2.78	43	22878	2.796	ug/l	99
20) Methylene Chloride	2.86	84	10989	1.496	ug/l	93
25) 2-Butanone	4.71	43	6366	1.575	ug/l	97
29) Tetrahydrofuran	5.17	42	1555	0.592	ug/l #	87
36) 1,1-Dichloropropene	5.67	75	34482	4.700	ug/l #	49
38) Carbon Tetrachloride	5.67	117	49337	6.684	ug/l #	15
43) Isopropyl Acetate	6.46	43	28664	2.310	ug/l	98
48) Methyl methacrylate	7.68	41	16492	2.589	ug/l #	12
49) 1,4-Dioxane	7.74	88	89	Below Cal	#	1
51) 4-Methyl-2-Pentanone	8.62	43	134048	16.813	ug/l #	48
52) Toluene	8.79	92	3600	0.250	ug/l	97
54) cis-1,3-Dichloropropene	8.62	75	64549	7.830	ug/l #	63
57) 1,3-Dichloropropane	9.54	76	15322	1.624	ug/l #	43
59) 2-Hexanone	9.54	43	207271	34.113	ug/l #	51
68) m/p-Xylenes	10.36	106	2296	0.202	ug/l	98
71) Bromoform	10.86	173	121	4.748	ug/l #	28
76) 1,2,3-Trichloropropane	11.13	75	172538	22.217	ug/l #	40
80) 1,3,5-Trimethylbenzene	11.51	105	4536	0.201	ug/l	96
81) trans-1,4-Dichloro-2-buten	11.13	75	172538	65.757	ug/l #	10
84) 1,2,4-Trimethylbenzene	11.81	105	12109	0.533	ug/l	98
90) Hexachloroethane	12.61	117	710	3.082	ug/l #	14
95) Naphthalene	13.83	128	26527	0.961	ug/l	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX012419\
Data File : VX007206.D
Acq On : 24 Jan 2019 16:22
Operator : JC/SP
Sample : K1016-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TR-05-RO-3-012319

Quant Time: Jan 25 00:04:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.67 min Scan# 740

Delta R.T. -0.00 min

Lab File: VX007206.D

Acq: 24 Jan 2019 16:22

Instrument :

MSVOA_X

ClientSampleId :


TR-05-RO-3-012319

Tot Ion:168 Resp: 546783

Ion Ratio Lower Upper

168 100

99 48.2 41.8 62.8



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.67

200000



150000

100000

50000

0

Time-->



#3

Chloromethane

Concen: 0.242 ug/l

RT: 1.33 min Scan# 29

Delta R.T. -0.00 min

Lab File: VX007206.D


Acq: 24 Jan 2019 16:22

Tgt Ion: 50 Resp: 1266

Ion Ratio Lower Upper

50 100

52 41.4 25.4 38.2#



Abundance Ion 49.90 (49.60 to 50.60): VX0

Ion 51.90 (51.60 to 52.60): VX0


1.33


1000

500

0

Time-->





#5

Bromomethane

Concen: 0.397 ug/l

RT: 1.65 min Scan# 81

Delta R.T. 0.01 min

Lab File: VX007206.D

Acq: 24 Jan 2019 16:22

Instrument :

MSVOA_X

ClientSampleId :





TR-05-RO-3-012319

Tot Ion: 94 Resp: 2386

Ion Ratio Lower Upper

94 100

96 47.9 73.4 110.2#



#11

Tert butyl alcohol

Concen: 1.145 ug/l

RT: 3.07 min Scan# 314

Delta R.T. -0.01 min

Lab File: VX007206.D




Acq: 24 Jan 2019 16:22


Tgt Ion: 59 Resp: 1543

Ion Ratio Lower Upper

59 100

57 90.0 8.5 12.7#





#13

Acrolein

Concen: 4.968 ug/l

RT: 2.30 min Scan# 188

Delta R.T. -0.00 min

Lab File: VX007206.D

Acq: 24 Jan 2019 16:22

Instrument :

MSVOA_X

ClientSampleId :





TR-05-RO-3-012319

Tot Ion: 56 Resp: 1109

Ion Ratio Lower Upper

56 100

55 53.5 58.2 87.4#



#16

Acetone

Concen: 7.202 ug/l

RT: 2.45 min Scan# 213

Delta R.T. 0.01 min

Lab File: VX007206.D




Acq: 24 Jan 2019 16:22


Tgt Ion: 43 Resp: 20346

Ion Ratio Lower Upper

43 100

58 23.0 25.0 37.6#









#18
Methyl Acetate
Concen: 2.796 ug/l
RT: 2.78 min Scan# 267
Delta R.T. -0.00 min
Lab File: VX007206.D
Acq: 24 Jan 2019 16:22




Instrument :
MSVOA_X
ClientSampleId :
TR-05-RO-3-012319


Tot Ion: 43 Resp: 22878
Ion Ratio Lower Upper
43 100
74 23.6 18.6 28.0



#20
Methylene Chloride
Concen: 1.496 ug/l
RT: 2.86 min Scan# 280
Delta R.T. 0.00 min
Lab File: VX007206.D
Acq: 24 Jan 2019 16:22

Tgt Ion: 84 Resp: 10989
Ion Ratio Lower Upper
84 100
49 122.3 91.8 137.6
51 29.9 28.5 42.7
86 67.8 50.2 75.2





#25

2-Butanone

Concen: 1.575 ug/l

RT: 4.71 min Scan# 583

Delta R.T. 0.01 min

Lab File: VX007206.D

Acq: 24 Jan 2019 16:22

Instrument :

MSVOA_X

ClientSampleId :





TR-05-RO-3-012319

Tot Ion: 43 Resp: 6366

Ion Ratio Lower Upper

43 100

72 29.7 22.4 33.6



#29

Tetrahydrofuran

Concen: 0.592 ug/l

RT: 5.17 min Scan# 658

Delta R.T. 0.01 min

Lab File: VX007206.D

Acq: 24 Jan 2019 16:22




Tgt Ion: 42 Resp: 1555


Ion Ratio Lower Upper

42 100

72 61.1 38.9 58.3#

71 41.2 36.2 54.4









#33
 1,2-Dichloroethane-d4
 Concen: 52.166 ug/l
 RT: 6.07 min Scan# 806
 Delta R.T. -0.00 min
 Lab File: VX007206.D
 Acq: 24 Jan 2019 16:22




Instrument :
 MSVOA_X
 ClientSampleId :
 TR-05-RO-3-012319


Tot Ion: 65 Resp: 298093
 Ion Ratio Lower Upper
 65 100
 67 53.2 0.0 105.0



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.86 min Scan# 936
 Delta R.T. -0.00 min
 Lab File: VX007206.D
 Acq: 24 Jan 2019 16:22

Tgt Ion: 114 Resp: 843827
 Ion Ratio Lower Upper
 114 100
 63 18.1 0.0 34.8
 88 14.6 0.0 28.8





#35

Dibromofluoromethane

Concen: 48.669 ug/l

RT: 5.51 min Scan# 714

Delta R.T. 0.01 min

Lab File: VX007206.D

Acq: 24 Jan 2019 16:22

Instrument :

MSVOA_X

ClientSampleId :

TR-05-RO-3-012319


Tot Ion:113 Resp: 262998

Ion Ratio Lower Upper

113 100

111 102.1 81.6 122.4




192 22.3 16.7 25.1



Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V



#36

1,1-Dichloropropene

Concen: 4.700 ug/l

RT: 5.67 min Scan# 740

Delta R.T. -0.13 min

Lab File: VX007206.D

Acq: 24 Jan 2019 16:22


Tgt Ion: 75 Resp: 34482

Ion Ratio Lower Upper

75 100

110 10.8 20.2 60.5#



77 0.7 24.8 37.2#




Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 109.90 (109.60 to 110.60): V

Ion 76.90 (76.60 to 77.60): VX0





#38

Carbon Tetrachloride

Concen: 6.684 ug/l

RT: 5.67 min Scan# 741

Delta R.T. -0.11 min

Lab File: VX007206.D

Acq: 24 Jan 2019 16:22

Instrument :

MSVOA_X

ClientSampleId :

TR-05-RO-3-012319


Tot Ion:117 Resp: 49337

Ion Ratio Lower Upper

117 100

119 4.9 79.4 119.2#




121 0.5 24.0 36.0#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V



#43

Isopropyl Acetate

Concen: 2.310 ug/l

RT: 6.46 min Scan# 870

Delta R.T. -0.00 min

Lab File: VX007206.D

Acq: 24 Jan 2019 16:22


Tgt Ion: 43 Resp: 28664

Ion Ratio Lower Upper

43 100

61 22.0 16.8 25.2



87 16.6 12.4 18.6




Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 61.00 (60.70 to 61.70): VX0

Ion 87.00 (86.70 to 87.70): VX0









#48
Methyl methacrylate
Concen: 2.589 ug/l
RT: 7.68 min Scan# 1071
Delta R.T. -0.09 min
Lab File: VX007206.D
Acq: 24 Jan 2019 16:22




Instrument :
MSVOA_X
ClientSampleId :
TR-05-RO-3-012319


Tot Ion: 41 Resp: 16492
Ion Ratio Lower Upper
41 100
69 1.5 72.4 108.6#
39 0.0 45.8 68.8#



#49
1,4-Dioxane
Concen: Below Cal
RT: 7.74 min Scan# 1081
Delta R.T. -0.01 min
Lab File: VX007206.D
Acq: 24 Jan 2019 16:22

Tgt Ion: 88 Resp: 89
Ion Ratio Lower Upper
88 100
43 0.0 22.2 33.4#
58 1673.0 51.0 76.4#





#50

Toluene-d8

Concen: 51.307 ug/l

RT: 8.71 min Scan# 1240

Delta R.T. -0.00 min

Lab File: VX007206.D

Acq: 24 Jan 2019 16:22

Instrument :

MSVOA_X

ClientSampleId :





TR-05-RO-3-012319

Tot Ion: 98 Resp: 1014832

Ion Ratio Lower Upper

98 100

100 64.9 52.0 78.0



#51

4-Methyl-2-Pentanone

Concen: 16.813 ug/l

RT: 8.62 min Scan# 1224

Delta R.T. -0.03 min

Lab File: VX007206.D




Acq: 24 Jan 2019 16:22


Tgt Ion: 43 Resp: 134048

Ion Ratio Lower Upper

43 100

58 7.8 31.8 47.6#





#52

Toluene

Concen: 0.250 ug/l

RT: 8.79 min Scan# 1252

Delta R.T. -0.00 min

Lab File: VX007206.D

Acq: 24 Jan 2019 16:22

Instrument :

MSVOA_X

ClientSampleId :





TR-05-RO-3-012319

Tot Ion: 92 Resp: 3600

Ion Ratio Lower Upper

92 100

91 176.8 137.6 206.4



#54

cis-1,3-Dichloropropene

Concen: 7.830 ug/l

RT: 8.62 min Scan# 1225

Delta R.T. 0.19 min

Lab File: VX007206.D

Acq: 24 Jan 2019 16:22




Tgt Ion: 75 Resp: 64549


Ion Ratio Lower Upper

75 100

77 0.3 26.2 39.2#

39 31.1 36.7 55.1#





#57

1,3-Dichloropropane

Concen: 1.624 ug/l

RT: 9.54 min Scan# 1376

Delta R.T. 0.17 min

Lab File: VX007206.D

Acq: 24 Jan 2019 16:22

Instrument :

MSVOA_X

ClientSampleId :





TR-05-RO-3-012319

Tot Ion: 76 Resp: 15322

Ion Ratio Lower Upper

76 100

78 0.7 26.0 39.0#



#59

2-Hexanone

Concen: 34.113 ug/l

RT: 9.54 min Scan# 1376

Delta R.T. 0.05 min

Lab File: VX007206.D




Acq: 24 Jan 2019 16:22


Tgt Ion: 43 Resp: 207271

Ion Ratio Lower Upper

43 100

58 19.9 27.7 83.1#





#62

4-Bromofluorobenzene

Concen: 51.447 ug/l

RT: 11.14 min Scan# 1638

Delta R.T. 0.01 min

Lab File: VX007206.D

Acq: 24 Jan 2019 16:22

Instrument :

MSVOA_X

ClientSampleId :

TR-05-RO-3-012319


Tot Ion: 95 Resp: 354761

Ion Ratio Lower Upper

95 100

174 93.1 0.0 182.2




176 90.8 0.0 178.8



Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.12 min Scan# 1470

Delta R.T. -0.00 min

Lab File: VX007206.D

Acq: 24 Jan 2019 16:22


Tgt Ion: 117 Resp: 801424

Ion Ratio Lower Upper

117 100

82 49.8 41.0 61.6



119 32.7 25.4 38.0




Abundance Ion 116.90 (116.60 to 117.60): V

Ion 82.00 (81.70 to 82.70): VX0

Ion 118.90 (118.60 to 119.60): V





#68

m/p-Xvlenes

Concen: 0.202 ug/l

RT: 10.36 min Scan# 1510

Delta R.T. -0.00 min

Lab File: VX007206.D

Acq: 24 Jan 2019 16:22

Instrument :

MSVOA_X

ClientSampleId :



TR-05-RO-3-012319

Tot Ion:106 Resp: 2296

Ion Ratio Lower Upper



106 100

91 192.5 155.8 233.6



Abundance Ion 106.00 (105.70 to 106.70): V

Ion 91.00 (90.70 to 91.70): VX0



#71

Bromoform

Concen: 4.748 ug/l

RT: 10.86 min Scan# 1592

Delta R.T. 0.00 min

Lab File: VX007206.D

Acq: 24 Jan 2019 16:22



Tgt Ion:173 Resp: 121

Ion Ratio Lower Upper

173 100

175 0.0 24.7 74.1#


254 0.0 0.2 0.2#




Abundance Ion 172.70 (172.40 to 173.40): V

Ion 174.70 (174.40 to 175.40): V

Ion 254.40 (254.10 to 255.10): V









#72

1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.08 min Scan# 1792
Delta R.T. -0.00 min
Lab File: VX007206.D
Acq: 24 Jan 2019 16:22

Instrument :
MSVOA_X
ClientSampleId :
TR-05-RO-3-012319




Tot Ion	Ratio	Lower	Upper
152	100		
115	55.1	39.1	117.2
150	155.7	0.0	344.8




#76

1,2,3-Trichloropropane
Concen: 22.217 ug/l
RT: 11.13 min Scan# 1637
Delta R.T. -0.16 min
Lab File: VX007206.D
Acq: 24 Jan 2019 16:22

Tgt Ion	Ratio	Lower	Upper
75	100		
77	1.2	18.5	55.5#









#80
 1,3,5-Trimethylbenzene
 Concen: 0.201 ug/l
 RT: 11.51 min Scan# 1698
 Delta R.T. -0.00 min
 Lab File: VX007206.D
 Acq: 24 Jan 2019 16:22




Instrument :
 MSVOA_X
 ClientSampleId :
 TR-05-RO-3-012319


Tot Ion:105 Resp: 4536
 Ion Ratio Lower Upper
 105 100
 120 53.3 25.1 75.3



#81
 trans-1,4-Dichloro-2-butene
 Concen: 65.757 ug/l
 RT: 11.13 min Scan# 1637
 Delta R.T. 0.06 min
 Lab File: VX007206.D
 Acq: 24 Jan 2019 16:22

Tgt Ion: 75 Resp: 172538
 Ion Ratio Lower Upper
 75 100
 53 0.0 79.7 119.5#
 89 0.1 38.0 57.0#





#84

1,2,4-Trimethylbenzene

Concen: 0.533 ug/l

RT: 11.81 min Scan# 1748

Delta R.T. 0.01 min

Lab File: VX007206.D

Acq: 24 Jan 2019 16:22

Instrument :

MSVOA_X

ClientSampleId :


TR-05-RO-3-012319

Tot Ion:105 Resp: 12109

Ion Ratio Lower Upper




105 100

120 47.4 23.2 69.6



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V



#90

Hexachloroethane

Concen: 3.082 ug/l

RT: 12.61 min Scan# 1879

Delta R.T. 0.01 min

Lab File: VX007206.D


Acq: 24 Jan 2019 16:22

Tgt Ion:117 Resp: 710

Ion Ratio Lower Upper



117 100


201 7.0 42.9 128.7#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 200.70 (200.40 to 201.40): V






#95
Naphthalene
Concen: 0.961 ug/l
RT: 13.83 min Scan# 2080
Delta R.T. -0.00 min
Lab File: VX007206.D
Acq: 24 Jan 2019 16:22

Instrument :
MSVOA_X
ClientSampleId :
TR-05-RO-3-012319

Tot Ion:128 Resp: 26527
Ion Ratio Lower Upper
128 100
127 14.6 10.1 15.1
129 12.7 8.6 13.0



Abundance Ion 128.00 (127.70 to 128.70): V
Ion 127.00 (126.70 to 127.70): V
Ion 129.00 (128.70 to 129.70): V

