

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX012419\
 Data File : VX007203.D
 Acq On : 24 Jan 2019 15:02
 Operator : JC/SP
 Sample : K1003-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BU-02-012319

Quant Time: Jan 25 01:23:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	540940	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	828226	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	789877	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	379248	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	287438	50.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.68%	
35) Dibromofluoromethane	5.51	113	259726	48.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.94%	
50) Toluene-d8	8.71	98	1009879	52.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.04%	
62) 4-Bromofluorobenzene	11.13	95	351956	52.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.00%	
Target Compounds						
16) Acetone	2.45	43	106026	37.937	ug/l	98
20) Methylene Chloride	2.85	84	15272	2.333	ug/l	99
25) 2-Butanone	4.71	43	7339	1.835	ug/l	98
43) Isopropyl Acetate	6.46	43	25401	2.086	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX012419\
Data File : VX007203.D
Acq On : 24 Jan 2019 15:02
Operator : JC/SP
Sample : K1003-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BU-02-012319

Quant Time: Jan 25 01:23:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

