

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX012420\
 Data File : VX014695.D
 Acq On : 24 Jan 2020 14:21
 Operator : JC/SP
 Sample : L1222-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Jan 24 22:52:13 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X011520W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Jan 15 12:41:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.01	128	67567	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.85	114	359341	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.11	117	321082	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	6.05	65	148371	31.24	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	104.13%
60) 4-Bromofluorobenzene	11.14	95	136928	26.65	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	88.83%
63) Toluene-d8	8.71	98	433204	29.87	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	99.57%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX012420\
Data File : VX014695.D
Acq On : 24 Jan 2020 14:21
Operator : JC/SP
Sample : L1222-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Jan 24 22:52:13 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X011520W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Jan 15 12:41:12 2020
Response via : Initial Calibration

