

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX012420\
 Data File : VX014697.D
 Acq On : 24 Jan 2020 15:08
 Operator : JC/SP
 Sample : L1232-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 55-ST-25

Quant Time: Jan 24 22:56:30 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X011520W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Jan 15 12:41:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.01	128	67830	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.85	114	362359	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.11	117	326193	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	6.05	65	153378	32.17	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	107.23%
60) 4-Bromofluorobenzene	11.14	95	141552	27.12	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	90.40%
63) Toluene-d8	8.71	98	440229	29.88	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	99.60%

Target Compounds

					Ovalue
15) Acetone	2.44	58	5964	7.665	ug/l 99
25) Chloroform	5.20	83	6325	0.874	ug/l 92
30) 2-Butanone	4.67	43	16809	4.993	ug/l # 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX012420\
Data File : VX014697.D
Acq On : 24 Jan 2020 15:08
Operator : JC/SP
Sample : L1232-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
55-ST-25

Quant Time: Jan 24 22:56:30 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X011520W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Jan 15 12:41:12 2020
Response via : Initial Calibration

