

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX012420\
 Data File : VX014707.D
 Acq On : 24 Jan 2020 19:16
 Operator : JC/SP
 Sample : L1173-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 0048-SITE-WATER

Quant Time: Jan 24 23:44:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 22 12:35:15 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	426902	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	668318	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	601967	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	261001	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	250974	50.34	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.68%	
35) Dibromofluoromethane	5.49	113	205580	48.75	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.50%	
50) Toluene-d8	8.71	98	795040	49.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.08%	
62) 4-Bromofluorobenzene	11.13	95	260242	45.06	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.12%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX012420\
Data File : VX014707.D
Acq On : 24 Jan 2020 19:16
Operator : JC/SP
Sample : L1173-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
0048-SITE-WATER

Quant Time: Jan 24 23:44:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X012220W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 22 12:35:15 2020
Response via : Initial Calibration

