

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX012420\
 Data File : VX014710.D
 Acq On : 24 Jan 2020 20:26
 Operator : JC/SP
 Sample : PB126303ZHE#11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB126303ZHE#11

Quant Time: Jan 24 23:50:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 22 12:35:15 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	445193	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	702405	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	655745	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	296019	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	255172	49.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.16%	
35) Dibromofluoromethane	5.49	113	217716	49.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.24%	
50) Toluene-d8	8.71	98	851398	49.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.94%	
62) 4-Bromofluorobenzene	11.14	95	292748	48.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.46%	
Target Compounds						
16) Acetone	2.44	43	13876	5.017	ug/l	100
20) Methylene Chloride	2.85	84	11787	2.236	ug/l	93
43) Isopropyl Acetate	6.43	43	105521	8.620	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX012420\
Data File : VX014710.D
Acq On : 24 Jan 2020 20:26
Operator : JC/SP
Sample : PB126303ZHE#11
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB126303ZHE#11

Quant Time: Jan 24 23:50:44 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X012220W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 22 12:35:15 2020
Response via : Initial Calibration

