

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012422\
 Data File : VX026555.D
 Acq On : 24 Jan 2022 12:51
 Operator : JC/MD
 Sample : N1245-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB

Quant Time: Jan 25 03:25:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	136319	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	227046	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	202427	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	82295	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	88905	46.687	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.380%
35) Dibromofluoromethane	5.391	113	71754	48.939	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.880%
50) Toluene-d8	8.653	98	265031	50.162	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.320%
62) 4-Bromofluorobenzene	11.079	95	94729	50.471	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.940%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	2705	2.785	ug/l	95
20) Methylene Chloride	2.794	84	931	0.580	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012422\
Data File : VX026555.D
Acq On : 24 Jan 2022 12:51
Operator : JC/MD
Sample : N1245-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB

Quant Time: Jan 25 03:25:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

