

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012423\
 Data File : VX033873.D
 Acq On : 24 Jan 2023 17:30
 Operator : JC/MD
 Sample : 01266-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VNJ-219

Quant Time: Jan 25 03:19:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	161293	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	236331	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	205706	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	94040	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	77867	45.851	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.700%
35) Dibromofluoromethane	5.385	113	71814	48.937	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.880%
50) Toluene-d8	8.647	98	268252	50.170	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.340%
62) 4-Bromofluorobenzene	11.079	95	89756	45.213	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.420%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	22018	24.501	ug/l	96
20) Methylene Chloride	2.788	84	8045	3.098	ug/l	92
37) Ethyl Acetate	4.721	43	18979	6.819	ug/l	98
43) Isopropyl Acetate	6.336	43	86804	25.673	ug/l	100
95) Naphthalene	13.774	128	16822	2.831	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012423\
Data File : VX033873.D
Acq On : 24 Jan 2023 17:30
Operator : JC/MD
Sample : 01266-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VNJ-219

Quant Time: Jan 25 03:19:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 11 14:48:42 2023
Response via : Initial Calibration

