

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012423\
 Data File : VX033874.D
 Acq On : 24 Jan 2023 17:53
 Operator : JC/MD
 Sample : PB150399TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB150399TB

Quant Time: Jan 25 03:20:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	158746	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	231383	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	204523	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96070	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	77609	46.432	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.860%
35) Dibromofluoromethane	5.385	113	70964	49.392	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.780%
50) Toluene-d8	8.647	98	266661	50.939	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.880%
62) 4-Bromofluorobenzene	11.079	95	90945	46.792	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.580%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	18650	19.269	ug/l	94
20) Methylene Chloride	2.788	84	6084	1.770	ug/l	93
37) Ethyl Acetate	4.721	43	22377	8.831	ug/l	99
43) Isopropyl Acetate	6.336	43	100206	30.271	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012423\
Data File : VX033874.D
Acq On : 24 Jan 2023 17:53
Operator : JC/MD
Sample : PB150399TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB150399TB

Quant Time: Jan 25 03:20:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 11 14:48:42 2023
Response via : Initial Calibration

