

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012423\
 Data File : VX033875.D
 Acq On : 24 Jan 2023 18:16
 Operator : JC/MD
 Sample : 01260-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-24

Quant Time: Jan 25 03:20:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	156688	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	228990	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	204093	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	98010	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	75873	45.990	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.980%
35) Dibromofluoromethane	5.385	113	69756	49.058	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.120%
50) Toluene-d8	8.647	98	262869	50.739	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.480%
62) 4-Bromofluorobenzene	11.079	95	90910	47.262	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.520%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	18283	19.050	ug/l	92
20) Methylene Chloride	2.788	84	5326	1.271	ug/l	93
37) Ethyl Acetate	4.715	43	21342	8.400	ug/l	98
43) Isopropyl Acetate	6.336	43	93798	28.631	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012423\
Data File : VX033875.D
Acq On : 24 Jan 2023 18:16
Operator : JC/MD
Sample : 01260-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-24

Quant Time: Jan 25 03:20:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 11 14:48:42 2023
Response via : Initial Calibration

