

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012424\
 Data File : VX039929.D
 Acq On : 24 Jan 2024 12:21
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jan 25 01:36:57 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	76577	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	140725	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	135464	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	64106	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63206	49.273	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.540%
35) Dibromofluoromethane	5.385	113	46242	48.787	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.580%
50) Toluene-d8	8.647	98	174429	48.793	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.580%
62) 4-Bromofluorobenzene	11.079	95	68256	48.826	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012424\
Data File : VX039929.D
Acq On : 24 Jan 2024 12:21
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jan 25 01:36:57 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

