

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012424\
 Data File : VX039923.D
 Acq On : 24 Jan 2024 09:56
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 25 01:33:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	80	0.00
2 T	Dichlorodifluoromethane	0.647	0.693	-7.1	84	0.00
3 P	Chloromethane	0.739	0.700	5.3	78	0.00
4 C	Vinyl Chloride	0.697	0.688	1.3#	81	0.00
5 T	Bromomethane	0.398	0.446	-12.1	92	0.00
6 T	Chloroethane	0.396	0.410	-3.5	81	0.00
7 T	Trichlorofluoromethane	1.006	1.087	-8.1	90	0.00
8 T	Diethyl Ether	0.372	0.369	0.8	83	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.624	0.640	-2.6	85	0.00
10 T	Methyl Iodide	0.658	0.716	-8.8	83	0.00
11 T	Tert butyl alcohol	0.218	0.185	15.1	69	0.00
12 CM	1,1-Dichloroethene	0.602	0.615	-2.2#	83	0.00
13 T	Acrolein	0.147	0.116	21.1	70	0.00
14 T	Allyl chloride	1.184	1.111	6.2	76	0.00
15 T	Acrylonitrile	0.450	0.393	12.7	70	0.00
16 T	Acetone	0.471	0.472	-0.2	80	0.00
17 T	Carbon Disulfide	1.730	1.729	0.1	83	0.00
18 T	Methyl Acetate	0.985	1.012	-2.7	84	0.00
19 T	Methyl tert-butyl Ether	2.136	2.062	3.5	80	0.00
20 T	Methylene Chloride	0.713	0.679	4.8	80	0.00
21 T	trans-1,2-Dichloroethene	0.633	0.644	-1.7	84	0.00
22 T	Diisopropyl ether	2.260	2.189	3.1	80	0.00
23 T	Vinyl Acetate	2.101	2.030	3.4	77	0.00
24 P	1,1-Dichloroethane	1.243	1.213	2.4	79	0.00
25 T	2-Butanone	0.677	0.612	9.6	73	0.00
26 T	2,2-Dichloropropane	0.998	0.983	1.5	80	0.00
27 T	cis-1,2-Dichloroethene	0.726	0.720	0.8	80	0.00
28 T	Bromochloromethane	0.564	0.552	2.1	79	0.00
29 T	Tetrahydrofuran	0.436	0.379	13.1	70	0.00
30 C	Chloroform	1.228	1.283	-4.5#	86	0.00
31 T	Cyclohexane	1.105	1.086	1.7	80	0.00
32 T	1,1,1-Trichloroethane	1.051	1.095	-4.2	84	0.00
33 S	1,2-Dichloroethane-d4	0.838	0.787	6.1	80	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	77	0.00
35 S	Dibromofluoromethane	0.337	0.333	1.2	79	0.00
36 T	1,1-Dichloropropene	0.524	0.549	-4.8	84	0.00
37 T	Ethyl Acetate	0.699	0.651	6.9	75	0.00
38 T	Carbon Tetrachloride	0.480	0.533	-11.0	85	0.00
39 T	Methylcyclohexane	0.629	0.680	-8.1	84	0.00
40 TM	Benzene	1.484	1.555	-4.8	82	0.00
41 T	Methacrylonitrile	0.342	0.349	-2.0	74	0.00
42 TM	1,2-Dichloroethane	0.564	0.638	-13.1	88	0.00
43 T	Isopropyl Acetate	1.026	1.034	-0.8	76	0.00
44 TM	Trichloroethene	0.364	0.390	-7.1	83	0.00
45 C	1,2-Dichloropropane	0.398	0.418	-5.0#	83	0.00
46 T	Dibromomethane	0.275	0.306	-11.3	86	0.00
47 T	Bromodichloromethane	0.509	0.579	-13.8	87	0.00
48 T	Methyl methacrylate	0.504	0.518	-2.8	78	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012424\
 Data File : VX039923.D
 Acq On : 24 Jan 2024 09:56
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 25 01:33:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.010	0.008	20.0	63	0.00
50 S	Toluene-d8	1.270	1.224	3.6	77	0.00
51 T	4-Methyl-2-Pentanone	0.707	0.698	1.3	75	0.00
52 CM	Toluene	0.887	0.976	-10.0#	84	0.00
53 T	t-1,3-Dichloropropene	0.540	0.617	-14.3	84	0.00
54 T	cis-1,3-Dichloropropene	0.594	0.658	-10.8	82	0.00
55 T	1,1,2-Trichloroethane	0.377	0.402	-6.6	83	0.00
56 T	Ethyl methacrylate	0.611	0.656	-7.4	81	0.00
57 T	1,3-Dichloropropane	0.640	0.685	-7.0	83	0.00
58 T	2-Chloroethyl Vinyl ether	0.289	0.275	4.8	71	0.00
59 T	2-Hexanone	0.559	0.575	-2.9	76	0.00
60 T	Dibromochloromethane	0.356	0.409	-14.9	86	0.00
61 T	1,2-Dibromoethane	0.391	0.432	-10.5	84	0.00
62 S	4-Bromofluorobenzene	0.497	0.519	-4.4	83	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	81	0.00
64 T	Tetrachloroethene	0.325	0.346	-6.5	92	0.00
65 PM	Chlorobenzene	1.041	1.082	-3.9	86	0.00
66 T	1,1,1,2-Tetrachloroethane	0.355	0.386	-8.7	85	0.00
67 C	Ethyl Benzene	1.933	2.056	-6.4#	86	0.00
68 T	m/p-Xylenes	0.711	0.768	-8.0	87	0.00
69 T	o-Xylene	0.704	0.740	-5.1	86	0.00
70 T	Styrene	1.150	1.254	-9.0	88	0.00
71 P	Bromoform	0.266	0.302	-13.5	86	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87	0.00
73 T	Isopropylbenzene	3.981	3.928	1.3	87	0.00
74 T	N-amyl acetate	2.075	1.984	4.4	80	0.00
75 P	1,1,2,2-Tetrachloroethane	1.520	1.402	7.8	82	0.00
76 T	1,2,3-Trichloropropane	1.296	1.220	5.9	84	0.00
77 T	Bromobenzene	0.886	0.880	0.7	90	0.00
78 T	n-propylbenzene	4.818	4.938	-2.5	89	0.00
79 T	2-Chlorotoluene	2.978	2.892	2.9	89	0.00
80 T	1,3,5-Trimethylbenzene	3.381	3.429	-1.4	89	0.00
81 T	trans-1,4-Dichloro-2-butene	0.427	0.384	10.1	76	0.00
82 T	4-Chlorotoluene	3.378	3.434	-1.7	91	0.00
83 T	tert-Butylbenzene	3.228	3.253	-0.8	89	0.00
84 T	1,2,4-Trimethylbenzene	3.373	3.451	-2.3	89	0.00
85 T	sec-Butylbenzene	4.190	4.378	-4.5	91	0.00
86 T	p-Isopropyltoluene	3.349	3.570	-6.6	91	0.00
87 T	1,3-Dichlorobenzene	1.699	1.745	-2.7	92	0.00
88 T	1,4-Dichlorobenzene	1.755	1.764	-0.5	92	0.00
89 T	n-Butylbenzene	3.234	3.565	-10.2	94	0.00
90 T	Hexachloroethane	0.565	0.597	-5.7	88	0.00
91 T	1,2-Dichlorobenzene	1.680	1.667	0.8	91	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.372	0.358	3.8	82	0.00
93 T	1,2,4-Trichlorobenzene	1.026	1.068	-4.1	93	0.00
94 T	Hexachlorobutadiene	0.427	0.477	-11.7	98	0.00
95 T	Naphthalene	3.779	3.652	3.4	85	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012424\
Data File : VX039923.D
Acq On : 24 Jan 2024 09:56
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jan 25 01:33:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.036	1.041	-0.5	90	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6