

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012424\
 Data File : VX039937.D
 Acq On : 24 Jan 2024 15:22
 Operator : JC/MD
 Sample : P1237-05MS
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SUMOGW-2-10-13MS

Quant Time: Jan 25 01:39:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 01/25/2024
 Supervised By :Semsettin Yesilyurt 01/25/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	84158	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	149845	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	141308	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	72912	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	65059	46.149	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	92.300%		
35) Dibromofluoromethane	5.385	113	46807	46.377	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	92.760%		
50) Toluene-d8	8.647	98	169541	44.540	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	89.080%#		
62) 4-Bromofluorobenzene	11.079	95	73174	49.158	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	98.320%		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	57696	52.987	ug/l	95
3) Chloromethane	1.295	50	57539	46.271	ug/l	99
4) Vinyl Chloride	1.374	62	55778	47.529	ug/l	99
5) Bromomethane	1.599	94	36045	53.830	ug/l	96
6) Chloroethane	1.679	64	33698	50.511	ug/l	97
7) Trichlorofluoromethane	1.886	101	89526	52.882	ug/l	99
8) Diethyl Ether	2.130	74	30127	48.155	ug/l	98
9) 1,1,2-Trichlorotrifluo...	2.325	101	52852	50.357	ug/l	98
10) Methyl Iodide	2.453	142	57657	52.073	ug/l	97
11) Tert butyl alcohol	2.959	59	87671	238.901	ug/l #	91
12) 1,1-Dichloroethene	2.319	96	49923	49.287	ug/l	92
13) Acrolein	2.233	56	52427	231.116	ug/l	99
14) Allyl chloride	2.660	41	92906	46.602	ug/l	98
15) Acrylonitrile	3.062	53	178588	235.591	ug/l	99
16) Acetone	2.374	43	154912	195.388	ug/l	96
17) Carbon Disulfide	2.508	76	143284	49.216	ug/l	98
18) Methyl Acetate	2.703	43	91490	55.157	ug/l	100
19) Methyl tert-butyl Ether	3.111	73	172625	48.005	ug/l	97
20) Methylene Chloride	2.788	84	57000	47.530	ug/l	95
21) trans-1,2-Dichloroethene	3.093	96	53553	50.241	ug/l	96
22) Diisopropyl ether	3.757	45	184474	48.489	ug/l	99
23) Vinyl Acetate	3.721	43	869890	245.978	ug/l	99
24) 1,1-Dichloroethane	3.611	63	102824	49.128	ug/l	98
25) 2-Butanone	4.550	43	263204	231.008	ug/l	99
26) 2,2-Dichloropropane	4.471	77	77272	46.001	ug/l	97
27) cis-1,2-Dichloroethene	4.489	96	61205	50.063	ug/l	100
28) Bromochloromethane	4.898	49	41206	43.433	ug/l	95
29) Tetrahydrofuran	4.995	42	176247	240.331	ug/l	99
30) Chloroform	5.093	83	108228	52.381	ug/l	95
31) Cyclohexane	5.471	56	90417	48.597	ug/l	99
32) 1,1,1-Trichloroethane	5.379	97	93645	52.941	ug/l	98
36) 1,1-Dichloropropene	5.690	75	78611	50.039	ug/l	100
37) Ethyl Acetate	4.709	43	98153	46.833	ug/l	99
38) Carbon Tetrachloride	5.678	117	77064	53.560	ug/l	99
39) Methylcyclohexane	7.379	83	94761	50.232	ug/l	98
40) Benzene	6.038	78	223700	50.287	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012424\
 Data File : VX039937.D
 Acq On : 24 Jan 2024 15:22
 Operator : JC/MD
 Sample : P1237-05MS
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SUMOGW-2-10-13MS

Quant Time: Jan 25 01:39:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 01/25/2024
 Supervised By :Semsettin Yesilyurt 01/25/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	52958	51.673	ug/l	95
42) 1,2-Dichloroethane	6.086	62	93116	55.059	ug/l	98
43) Isopropyl Acetate	6.336	43	155102	50.449	ug/l	98
44) Trichloroethene	7.123	130	54779	50.246	ug/l	91
45) 1,2-Dichloropropane	7.428	63	58040	48.637	ug/l	97
46) Dibromomethane	7.580	93	43915	53.280	ug/l	95
47) Bromodichloromethane	7.818	83	83338	54.593	ug/l	96
48) Methyl methacrylate	7.690	41	78414	51.889	ug/l	94
49) 1,4-Dioxane	7.659	88	26089	892.206	ug/l #	97
51) 4-Methyl-2-Pentanone	8.574	43	545680	257.712	ug/l	96
52) Toluene	8.720	92	140109	52.732	ug/l	94
53) t-1,3-Dichloropropene	8.976	75	85421	52.735	ug/l	97
54) cis-1,3-Dichloropropene	8.366	75	91236	51.224	ug/l	99
55) 1,1,2-Trichloroethane	9.153	97	58390	51.616	ug/l	94
56) Ethyl methacrylate	9.116	69	96317	52.607	ug/l	95
57) 1,3-Dichloropropane	9.305	76	100079	52.204	ug/l	99
59) 2-Hexanone	9.427	43	439061	262.000	ug/l	96
60) Dibromochloromethane	9.519	129	59931	56.177	ug/l	100
61) 1,2-Dibromoethane	9.610	107	63916	54.560	ug/l	100
64) Tetrachloroethene	9.275	164	60777	66.139	ug/l	96
65) Chlorobenzene	10.080	112	147703	50.186	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	54305	54.188	ug/l	99
67) Ethyl Benzene	10.195	91	283915	51.960	ug/l	98
68) m/p-Xylenes	10.299	106	208066	103.540	ug/l	94
69) o-Xylene	10.640	106	99477	49.987	ug/l	91
70) Styrene	10.653	104	169755	52.219	ug/l	98
71) Bromoform	10.799	173	42166	56.034	ug/l #	100
73) Isopropylbenzene	10.964	105	273657	47.142	ug/l	100
74) N-amyl acetate	10.842	43	142117	46.967	ug/l	97
75) 1,1,2,2-Tetrachloroethane	11.213	83	105565	47.623	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	89487m	47.340	ug/l	
77) Bromobenzene	11.195	156	61105	47.276	ug/l	95
78) n-propylbenzene	11.305	91	341430	48.595	ug/l	98
79) 2-Chlorotoluene	11.360	91	202684	46.675	ug/l	98
80) 1,3,5-Trimethylbenzene	11.451	105	241231	48.929	ug/l	95
81) trans-1,4-Dichloro-2-b...	11.018	75	27991	44.741	ug/l	97
82) 4-Chlorotoluene	11.451	91	239138	48.550	ug/l	98
83) tert-Butylbenzene	11.713	119	225695	47.949	ug/l	95
84) 1,2,4-Trimethylbenzene	11.750	105	241118	49.020	ug/l	100
85) sec-Butylbenzene	11.890	105	301856	49.402	ug/l	98
86) p-Isopropyltoluene	12.006	119	247367	50.654	ug/l	97
87) 1,3-Dichlorobenzene	11.969	146	119657	48.296	ug/l	98
88) 1,4-Dichlorobenzene	12.036	146	122489	47.868	ug/l	99
89) n-Butylbenzene	12.329	91	241874	51.281	ug/l	95
90) Hexachloroethane	12.536	117	41262	50.085	ug/l	95
91) 1,2-Dichlorobenzene	12.335	146	117139	47.823	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.939	75	27543	50.707	ug/l	95
93) 1,2,4-Trichlorobenzene	13.585	180	72314	48.311	ug/l	98
94) Hexachlorobutadiene	13.725	225	30641	49.243	ug/l	98
95) Naphthalene	13.774	128	259528	47.097	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	70856	46.903	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012424\
Data File : VX039937.D
Acq On : 24 Jan 2024 15:22
Operator : JC/MD
Sample : P1237-05MS
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SUMOGW-2-10-13MS


Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024

Quant Time: Jan 25 01:39:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#)=qualifier out of range (m)=manual integration (+)=signals summed						



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22




Instrument :
MSVOA_X
ClientSampleId :
SUMOGW-2-10-13MS



Tgt Ion:168 Resp: 84158
Ion Ratio Lower Upper
168 100
99 66.3 56.6 85.0




Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024






#2
Dichlorodifluoromethane
Concen: 52.987 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22

Tgt Ion: 85 Resp: 57696
Ion Ratio Lower Upper
85 100
87 30.7 16.7 50.1



Abundance Scan 34 (1.294 min): VX039697.D\data.ms (-26)



#3

Chloromethane

Concen: 46.271 ug/l

RT: 1.295 min Scan# 34

Delta R.T. 0.000 min

Lab File: VX039937.D

Acq: 24 Jan 2024 15:22

Tgt Ion: 50 Resp: 57539

Ion Ratio Lower Upper

50 100

52 32.2 25.5 38.3

Instrument :

MSVOA_X

ClientSampleId :


SUMOGW-2-10-13MS

Manual Integrations




APPROVED

Reviewed By :Mahesh Dadoda 01/25/2024

Supervised By :Semsettin Yesilyurt 01/25/2024



Abundance Scan 47 (1.374 min): VX039697.D\data.ms (-42)



#4

Vinyl Chloride

Concen: 47.529 ug/l

RT: 1.374 min Scan# 47

Delta R.T. 0.000 min

Lab File: VX039937.D


Acq: 24 Jan 2024 15:22


Tgt Ion: 62 Resp: 55778

Ion Ratio Lower Upper

62 100


64 32.1 25.4 38.0





#5
Bromomethane
Concen: 53.830 ug/l
RT: 1.599 min Scan# 84
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22




Instrument :
MSVOA_X
ClientSampleId :
SUMOGW-2-10-13MS



Tgt Ion: 94 Resp: 3604
Ion Ratio Lower Upper
94 100
96 97.0 74.6 111.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024



#6
Chloroethane
Concen: 50.511 ug/l
RT: 1.679 min Scan# 97
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22

Tgt Ion: 64 Resp: 33698
Ion Ratio Lower Upper
64 100
66 33.1 25.1 37.7









#7
Trichlorofluoromethane
Concen: 52.882 ug/l
RT: 1.886 min Scan# 131
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22

Instrument :
MSVOA_X
ClientSampleId :
SUMOGW-2-10-13MS

Tgt Ion:101 Resp: 89520
Ion Ratio Lower Upper
101 100
103 63.8 51.4 77.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024



#8
Diethyl Ether
Concen: 48.155 ug/l
RT: 2.130 min Scan# 171
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22

Tgt Ion: 74 Resp: 30127
Ion Ratio Lower Upper
74 100
45 104.9 53.3 159.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 50.357 ug/l
RT: 2.325 min Scan# 204
Delta R.T. -0.006 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22

Instrument :
MSVOA_X
ClientSampleId :
SUMOGW-2-10-13MS

Tgt Ion:101 Resp: 5285
Ion Ratio Lower Upper
101 100
85 47.8 37.0 55.6
151 71.4 58.6 87.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024



#10
Methyl Iodide
Concen: 52.073 ug/l
RT: 2.453 min Scan# 224
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22

Tgt Ion:142 Resp: 57657
Ion Ratio Lower Upper
142 100
127 48.9 41.6 62.4
141 14.4 11.6 17.4





#11

Tert butyl alcohol

Concen: 238.901 ug/l

RT: 2.959 min Scan# 307

Delta R.T. 0.000 min

Lab File: VX039937.D

Acq: 24 Jan 2024 15:22

Instrument :

MSVOA_X

ClientSampleId :

SUMOGW-2-10-13MS

Tgt Ion: 59 Resp: 8767

Ion Ratio Lower Upper

59 100





57 10.7 6.0 9.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 01/25/2024

Supervised By :Semsettin Yesilyurt 01/25/2024



#12

1,1-Dichloroethene

Concen: 49.287 ug/l

RT: 2.319 min Scan# 202

Delta R.T. 0.000 min

Lab File: VX039937.D

Acq: 24 Jan 2024 15:22




Tgt Ion: 96 Resp: 49923


Ion Ratio Lower Upper

96 100

61 166.3 144.6 216.8


98 62.7 50.4 75.6





#13
Acrolein
Concen: 231.116 ug/l
RT: 2.233 min Scan# 188
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22




Instrument :
MSVOA_X
ClientSampleId :
SUMOGW-2-10-13MS



Tgt Ion: 56 Resp: 5242
Ion Ratio Lower Upper
56 100
55 72.5 57.1 85.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024



#14
Allyl chloride
Concen: 46.602 ug/l
RT: 2.660 min Scan# 258
Delta R.T. -0.006 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22


Tgt Ion: 41 Resp: 92906
Ion Ratio Lower Upper
41 100
39 71.2 55.8 83.6
76 31.6 24.3 36.5





#15
Acrylonitrile
Concen: 235.591 ug/l
RT: 3.062 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22




Instrument :
MSVOA_X
ClientSampleId :
SUMOGW-2-10-13MS



Tgt Ion: 53 Resp: 17858
Ion Ratio Lower Upper
53 100
52 83.5 66.0 99.0
51 37.1 29.8 44.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024



#16
Acetone
Concen: 195.388 ug/l
RT: 2.374 min Scan# 211
Delta R.T. -0.006 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22

Tgt Ion: 43 Resp: 154912
Ion Ratio Lower Upper
43 100
58 29.5 21.9 32.9









#17
Carbon Disulfide
Concen: 49.216 ug/l
RT: 2.508 min Scan# 21
Delta R.T. -0.006 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22

Instrument :
MSVOA_X
ClientSampleId :
SUMOGW-2-10-13MS

Tgt Ion: 76 Resp: 143284
Ion Ratio Lower Upper
76 100
78 8.1 7.0 10.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024



#18
Methyl Acetate
Concen: 55.157 ug/l
RT: 2.703 min Scan# 265
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22

Tgt Ion: 43 Resp: 91490
Ion Ratio Lower Upper
43 100
74 22.2 17.8 26.8









#19
Methyl tert-butyl Ether
Concen: 48.005 ug/l
RT: 3.111 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22

Instrument :
MSVOA_X
ClientSampleId :
SUMOGW-2-10-13MS

Tgt Ion: 73 Resp: 172625
Ion Ratio Lower Upper
73 100
57 23.3 17.3 25.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024



#20
Methylene Chloride
Concen: 47.530 ug/l
RT: 2.788 min Scan# 279
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22


Tgt Ion: 84 Resp: 57000
Ion Ratio Lower Upper
84 100
49 136.4 103.6 155.4
51 40.7 33.8 50.6
86 64.9 49.4 74.0





#21
trans-1,2-Dichloroethene
Concen: 50.241 ug/l
RT: 3.093 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22




Instrument :
MSVOA_X
ClientSampleId :
SUMOGW-2-10-13MS



Tgt Ion: 96 Resp: 5355
Ion Ratio Lower Upper
96 100
61 143.9 119.2 178.8
98 64.3 48.7 73.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024



#22
Diisopropyl ether
Concen: 48.489 ug/l
RT: 3.757 min Scan# 438
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22


Tgt Ion: 45 Resp: 184474
Ion Ratio Lower Upper
45 100
43 88.9 69.7 104.5
87 25.8 20.8 31.2
59 11.0 9.0 13.4





#23
 Vinyl Acetate
 Concen: 245.978 ug/l
 RT: 3.721 min Scan# 41
 Delta R.T. 0.000 min
 Lab File: VX039937.D
 Acq: 24 Jan 2024 15:22




Instrument :
 MSVOA_X
 ClientSampleId :
 SUMOGW-2-10-13MS




Tgt Ion: 43 Resp: 869890
 Ion Ratio Lower Upper
 43 100
 86 9.6 7.8 11.8

Manual Integrations
 APPROVED



Reviewed By :Mahesh Dadoda 01/25/2024
 Supervised By :Semsettin Yesilyurt 01/25/2024




#24
 1,1-Dichloroethane
 Concen: 49.128 ug/l
 RT: 3.611 min Scan# 414
 Delta R.T. 0.000 min
 Lab File: VX039937.D
 Acq: 24 Jan 2024 15:22




Tgt Ion: 63 Resp: 102824
 Ion Ratio Lower Upper
 63 100
 98 5.3 3.3 9.8
 100 4.6 2.1 6.5





#25
2-Butanone
Concen: 231.008 ug/l
RT: 4.550 min Scan# 568
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22




Instrument :
MSVOA_X
ClientSampleId :
SUMOGW-2-10-13MS



Tgt Ion: 43 Resp: 263204
Ion Ratio Lower Upper
43 100
72 24.7 19.4 29.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024



#26
2,2-Dichloropropane
Concen: 46.001 ug/l
RT: 4.471 min Scan# 555
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22


Tgt Ion: 77 Resp: 77272
Ion Ratio Lower Upper
77 100
97 22.8 10.8 32.3





#27
cis-1,2-Dichloroethene
Concen: 50.063 ug/l
RT: 4.489 min Scan# 557
Delta R.T. 0.006 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22




Instrument :
MSVOA_X
ClientSampleId :
SUMOGW-2-10-13MS




Tgt Ion: 96 Resp: 6120
Ion Ratio Lower Upper
96 100
61 155.3 0.0 312.2
98 64.3 0.0 129.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024




#28
Bromochloromethane
Concen: 43.433 ug/l
RT: 4.898 min Scan# 625
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22



Tgt Ion: 49 Resp: 41206
Ion Ratio Lower Upper
49 100
129 1.8 0.0 3.8
130 67.9 57.5 86.3





#29

Tetrahydrofuran

Concen: 240.331 ug/l

RT: 4.995 min Scan# 642

Delta R.T. -0.006 min

Lab File: VX039937.D

Acq: 24 Jan 2024 15:22

Instrument :

MSVOA_X

ClientSampleId :

SUMOGW-2-10-13MS





Tgt Ion: 42 Resp: 17624

Ion	Ratio	Lower	Upper
42	100		
72	43.0	34.2	51.4
71	39.8	32.2	48.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024



#30

Chloroform

Concen: 52.381 ug/l

RT: 5.093 min Scan# 657




Delta R.T. 0.000 min


Lab File: VX039937.D

Acq: 24 Jan 2024 15:22

Tgt Ion: 83 Resp: 108228


Ion	Ratio	Lower	Upper
83	100		
85	66.8	50.3	75.5





#31
Cyclohexane
Concen: 48.597 ug/l
RT: 5.471 min Scan# 719
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22




Instrument :
MSVOA_X
ClientSampleId :
SUMOGW-2-10-13MS



Tgt Ion: 56 Resp: 9041
Ion Ratio Lower Upper
56 100
69 30.3 23.4 35.2
84 81.6 65.9 98.9




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024



#32
1,1,1-Trichloroethane
Concen: 52.941 ug/l
RT: 5.379 min Scan# 704
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22

Tgt Ion: 97 Resp: 93645
Ion Ratio Lower Upper
97 100
99 64.3 51.1 76.7
61 48.1 36.4 54.6









#33
1,2-Dichloroethane-d4
Concen: 46.149 ug/l
RT: 5.952 min Scan# 798
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22

Instrument :
MSVOA_X
ClientSampleId :
SUMOGW-2-10-13MS

Tgt Ion: 65 Resp: 65059
Ion Ratio Lower Upper
65 100
67 50.2 0.0 100.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 931
Delta R.T. 0.006 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22

Tgt Ion:114 Resp: 149845
Ion Ratio Lower Upper
114 100
63 21.7 0.0 45.0
88 17.6 0.0 36.6





#35

Dibromofluoromethane

Concen: 46.377 ug/l

RT: 5.385 min Scan# 705

Delta R.T. 0.000 min

Lab File: VX039937.D

Acq: 24 Jan 2024 15:22

Instrument :

MSVOA_X

ClientSampleId :





SUMOGW-2-10-13MS

Tgt Ion:	113	Resp:	4680
Ion	Ratio	Lower	Upper
113	100		
111	105.2	82.9	124.3
192	16.7	13.3	19.9

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024



#36

1,1-Dichloropropene

Concen: 50.039 ug/l




RT: 5.690 min Scan# 755


Delta R.T. 0.000 min

Lab File: VX039937.D

Acq: 24 Jan 2024 15:22


Tgt Ion:	75	Resp:	78611
Ion	Ratio	Lower	Upper
75	100		
110	33.8	17.0	50.9
77	30.5	24.7	37.1





#37
Ethyl Acetate
Concen: 46.833 ug/l
RT: 4.709 min Scan# 594
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22




Instrument :
MSVOA_X
ClientSampleId :
SUMOGW-2-10-13MS



Tgt Ion: 43 Resp: 9815
Ion Ratio Lower Upper
43 100
61 13.6 10.6 16.0
70 10.0 8.3 12.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024



#38
Carbon Tetrachloride
Concen: 53.560 ug/l
RT: 5.678 min Scan# 753
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22


Tgt Ion:117 Resp: 77064
Ion Ratio Lower Upper
117 100
119 99.2 78.2 117.2
121 31.0 24.4 36.6





#39
Methylcyclohexane
Concen: 50.232 ug/l
RT: 7.379 min Scan# 1032
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22




Instrument :
MSVOA_X
ClientSampleId :
SUMOGW-2-10-13MS




Tgt Ion: 83 Resp: 9476
Ion Ratio Lower Upper
83 100
55 78.9 64.6 97.0
98 44.5 36.7 55.1

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024




#40
Benzene
Concen: 50.287 ug/l
RT: 6.038 min Scan# 812
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22




Tgt Ion: 78 Resp: 223700
Ion Ratio Lower Upper
78 100
77 22.9 19.3 28.9





#41
Methacrylonitrile
Concen: 51.673 ug/l
RT: 4.922 min Scan# 61
Delta R.T. 0.006 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22




Instrument :
MSVOA_X
ClientSampleId :
SUMOGW-2-10-13MS



Tgt Ion: 41 Resp: 52958
Ion Ratio Lower Upper
41 100
39 60.0 45.2 67.8
67 69.7 52.6 78.8
52 31.6 22.7 34.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024



#42
1,2-Dichloroethane
Concen: 55.059 ug/l
RT: 6.086 min Scan# 820
Delta R.T. 0.006 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22


Tgt Ion: 62 Resp: 93116
Ion Ratio Lower Upper
62 100
98 7.4 0.0 16.0





#43
Isopropyl Acetate
Concen: 50.449 ug/l
RT: 6.336 min Scan# 861
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22




Instrument :
MSVOA_X
ClientSampleId :
SUMOGW-2-10-13MS



Tgt Ion: 43 Resp: 15510
Ion Ratio Lower Upper
43 100
61 18.3 15.2 22.8
87 11.6 9.8 14.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024



#44
Trichloroethene
Concen: 50.246 ug/l
RT: 7.123 min Scan# 990
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22

Tgt Ion:130 Resp: 54779
Ion Ratio Lower Upper
130 100
95 111.2 0.0 204.4





#45

1,2-Dichloropropane

Concen: 48.637 ug/l

RT: 7.428 min Scan# 1040

Delta R.T. 0.000 min

Lab File: VX039937.D

Acq: 24 Jan 2024 15:22

Instrument :

MSVOA_X

ClientSampleId :

SUMOGW-2-10-13MS

Tgt Ion: 63 Resp: 5804

Ion Ratio Lower Upper

63 100





65 31.9 24.3 36.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 01/25/2024

Supervised By :Semsettin Yesilyurt 01/25/2024



#46

Dibromomethane

Concen: 53.280 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. 0.000 min

Lab File: VX039937.D

Acq: 24 Jan 2024 15:22




Tgt Ion: 93 Resp: 43915


Ion Ratio Lower Upper

93 100

95 81.8 66.6 100.0

174 79.5 69.4 104.0









#47
Bromodichloromethane
Concen: 54.593 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22

Instrument :
MSVOA_X
ClientSampleId :
SUMOGW-2-10-13MS

Tgt Ion: 83 Resp: 83338
Ion Ratio Lower Upper
83 100
85 65.2 49.8 74.6
127 8.1 7.1 10.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024



#48
Methyl methacrylate
Concen: 51.889 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22

Tgt Ion: 41 Resp: 78414
Ion Ratio Lower Upper
41 100
69 76.9 59.2 88.8
39 66.8 47.7 71.5





#49

1,4-Dioxane

Concen: 892.206 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. 0.000 min

Lab File: VX039937.D

Acq: 24 Jan 2024 15:22

Instrument :

MSVOA_X

ClientSampleId :

SUMOGW-2-10-13MS





Tgt Ion: 88 Resp: 26089

Ion	Ratio	Lower	Upper
88	100		
43	38.6	0.0	0.0
58	72.7	60.1	90.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024



#50

Toluene-d8

Concen: 44.540 ug/l

RT: 8.647 min Scan# 1240




Delta R.T. 0.000 min


Lab File: VX039937.D

Acq: 24 Jan 2024 15:22

Tgt Ion: 98 Resp: 169541


Ion	Ratio	Lower	Upper
98	100		
100	63.7	53.1	79.7





#51
4-Methyl-2-Pentanone
Concen: 257.712 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22




Instrument :
MSVOA_X
ClientSampleId :
SUMOGW-2-10-13MS




Tgt Ion: 43 Resp: 545680
Ion Ratio Lower Upper
43 100
58 35.7 30.4 45.6

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024




#52
Toluene
Concen: 52.732 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. 0.006 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22



Tgt Ion: 92 Resp: 140109
Ion Ratio Lower Upper
92 100
91 175.2 134.2 201.2





#53

t-1,3-Dichloropropene

Concen: 52.735 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX039937.D

Acq: 24 Jan 2024 15:22

Instrument :

MSVOA_X

ClientSampleId :

SUMOGW-2-10-13MS

Tgt Ion: 75 Resp: 8542

Ion Ratio Lower Upper





75 100

77 30.5 25.7 38.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024



#54

cis-1,3-Dichloropropene

Concen: 51.224 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX039937.D

Acq: 24 Jan 2024 15:22




Tgt Ion: 75 Resp: 91236


Ion Ratio Lower Upper

75 100

77 31.5 25.3 37.9


39 59.2 46.7 70.1





#55
1,1,2-Trichloroethane
Concen: 51.616 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22




Instrument :
MSVOA_X
ClientSampleId :
SUMOGW-2-10-13MS




Tgt Ion:	97	Resp:	58390
Ion Ratio	100	Lower	Upper
97	100		
83	87.8	65.8	98.8
85	57.9	41.6	62.4
99	62.0	47.8	71.8

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024




#56
Ethyl methacrylate
Concen: 52.607 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22




Tgt Ion:	69	Resp:	96317
Ion Ratio	100	Lower	Upper
69	100		
41	72.5	56.8	85.2
39	47.3	32.0	48.0





#57
1,3-Dichloropropane
Concen: 52.204 ug/l
RT: 9.305 min Scan# 1348
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22




Instrument :
MSVOA_X
ClientSampleId :
SUMOGW-2-10-13MS



Tgt Ion: 76 Resp: 100079
Ion Ratio Lower Upper
76 100
78 32.0 26.1 39.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024



#59
2-Hexanone
Concen: 262.000 ug/l
RT: 9.427 min Scan# 1368
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22


Tgt Ion: 43 Resp: 439061
Ion Ratio Lower Upper
43 100
58 49.6 26.2 78.5





#60
Dibromochloromethane
Concen: 56.177 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22




Instrument :
MSVOA_X
ClientSampleId :
SUMOGW-2-10-13MS



Tgt Ion:129 Resp: 5993
Ion Ratio Lower Upper
129 100
127 77.7 38.8 116.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024



#61
1,2-Dibromoethane
Concen: 54.560 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22


Tgt Ion:107 Resp: 63916
Ion Ratio Lower Upper
107 100
109 94.1 75.2 112.8





#62
4-Bromofluorobenzene
Concen: 49.158 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22




Instrument :
MSVOA_X
ClientSampleId :
SUMOGW-2-10-13MS




Tgt Ion: 95 Resp: 73174
Ion Ratio Lower Upper
95 100
174 65.2 0.0 135.6
176 62.3 0.0 131.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024




#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22




Tgt Ion:117 Resp: 141308
Ion Ratio Lower Upper
117 100
82 61.0 46.2 69.4
119 32.4 25.5 38.3





#64
Tetrachloroethene
Concen: 66.139 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22




Instrument :
MSVOA_X
ClientSampleId :
SUMOGW-2-10-13MS



Tgt Ion:164 Resp: 60772
Ion Ratio Lower Upper
164 100
166 124.3 102.7 154.1
129 102.2 80.4 120.6
131 101.6 76.5 114.7




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024



#65
Chlorobenzene
Concen: 50.186 ug/l
RT: 10.080 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22


Tgt Ion:112 Resp: 147703
Ion Ratio Lower Upper
112 100
114 32.7 25.7 38.5





#66
1,1,1,2-Tetrachloroethane
Concen: 54.188 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22




Instrument :
MSVOA_X
ClientSampleId :
SUMOGW-2-10-13MS



Tgt Ion:131 Resp: 54308
Ion Ratio Lower Upper
131 100
133 93.7 47.1 141.4
119 64.9 32.2 96.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024



#67
Ethyl Benzene
Concen: 51.960 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22


Tgt Ion: 91 Resp: 283915
Ion Ratio Lower Upper
91 100
106 29.8 24.9 37.3





#68
m/p-Xylenes
Concen: 103.540 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22




Instrument :
MSVOA_X
ClientSampleId :
SUMOGW-2-10-13MS



Tgt Ion:106 Resp: 208060
Ion Ratio Lower Upper
106 100
91 216.4 165.5 248.3




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024



#69
o-Xylene
Concen: 49.987 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22

Tgt Ion:106 Resp: 99477
Ion Ratio Lower Upper
106 100
91 232.6 108.9 326.8





#70

Styrene

Concen: 52.219 ug/l

RT: 10.653 min Scan# 1569

Delta R.T. 0.000 min

Lab File: VX039937.D

Acq: 24 Jan 2024 15:22

Instrument :

MSVOA_X

ClientSampleId :

SUMOGW-2-10-13MS





Tgt Ion:104 Resp: 16975

Ion	Ratio	Lower	Upper
104	100		
78	56.9	43.6	65.4
103	55.1	43.6	65.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024



#71

Bromoform

Concen: 56.034 ug/l

RT: 10.799 min Scan# 1593




Delta R.T. 0.000 min


Lab File: VX039937.D

Acq: 24 Jan 2024 15:22

Tgt Ion:173 Resp: 42166


Ion	Ratio	Lower	Upper
173	100		
175	48.1	24.1	72.4
254	0.0	0.0	0.0





#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1794
Delta R.T. -0.006 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22




Instrument :
MSVOA_X
ClientSampleId :
SUMOGW-2-10-13MS




Tgt Ion:152 Resp: 72912
Ion Ratio Lower Upper
152 100
115 93.6 43.3 129.9
150 178.5 0.0 345.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024




#73
Isopropylbenzene
Concen: 47.142 ug/l
RT: 10.964 min Scan# 1620
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22



Tgt Ion:105 Resp: 273657
Ion Ratio Lower Upper
105 100
120 25.9 12.9 38.6





#74

N-amy1 acetate

Concen: 46.967 ug/l

RT: 10.842 min Scan# 1600

Delta R.T. 0.000 min

Lab File: VX039937.D

Acq: 24 Jan 2024 15:22

Instrument :

MSVOA_X

ClientSampleId :

SUMOGW-2-10-13MS

Tgt Ion: 43 Resp: 14211

Ion Ratio Lower Upper

43 100

70 38.2 32.9 49.3

55 23.4 19.3 28.9





61 21.9 18.8 28.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 01/25/2024

Supervised By :Semsettin Yesilyurt 01/25/2024



#75

1,1,2,2-Tetrachloroethane

Concen: 47.623 ug/l

RT: 11.213 min Scan# 1661

Delta R.T. 0.000 min

Lab File: VX039937.D

Acq: 24 Jan 2024 15:22




Tgt Ion: 83 Resp: 105565


Ion Ratio Lower Upper

83 100

131 9.3 5.3 15.8

85 63.7 32.0 96.2









#76
1,2,3-Trichloropropane
Concen: 47.340 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22

Instrument :
MSVOA_X
ClientSampleId :
SUMOGW-2-10-13MS

Tgt Ion: 75 Resp: 8948
Ion Ratio Lower Upper
75 100
77 40.2 19.1 57.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024



#77
Bromobenzene
Concen: 47.276 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22


Tgt Ion:156 Resp: 61105
Ion Ratio Lower Upper
156 100
77 187.7 88.4 265.3
158 95.5 48.0 144.0





#78
n-propylbenzene
Concen: 48.595 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22




Instrument :
MSVOA_X
ClientSampleId :
SUMOGW-2-10-13MS




Tgt Ion: 91 Resp: 341430
Ion Ratio Lower Upper
91 100
120 21.8 11.5 34.4

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024




#79
2-Chlorotoluene
Concen: 46.675 ug/l
RT: 11.360 min Scan# 1685
Delta R.T. -0.006 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22



Tgt Ion: 91 Resp: 202684
Ion Ratio Lower Upper
91 100
126 31.5 16.3 48.8





#80

1,3,5-Trimethylbenzene

Concen: 48.929 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. 0.000 min

Lab File: VX039937.D

Acq: 24 Jan 2024 15:22

Instrument :

MSVOA_X

ClientSampleId :

SUMOGW-2-10-13MS

Tgt Ion:105 Resp: 24123

Ion Ratio Lower Upper

105 100





120 45.0 24.1 72.3

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 01/25/2024

Supervised By :Semsettin Yesilyurt 01/25/2024



#81

trans-1,4-Dichloro-2-butene

Concen: 44.741 ug/l

RT: 11.018 min Scan# 1629

Delta R.T. 0.000 min

Lab File: VX039937.D

Acq: 24 Jan 2024 15:22




Tgt Ion: 75 Resp: 27991


Ion Ratio Lower Upper

75 100

53 115.5 90.5 135.7


89 44.0 36.9 55.3





#82
4-Chlorotoluene
Concen: 48.550 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22




Instrument :
MSVOA_X
ClientSampleId :
SUMOGW-2-10-13MS




Tgt Ion: 91 Resp: 239138
Ion Ratio Lower Upper
91 100
126 27.3 14.2 42.6

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024




#83
tert-Butylbenzene
Concen: 47.949 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22




Tgt Ion:119 Resp: 225695
Ion Ratio Lower Upper
119 100
91 66.4 31.2 93.6
134 22.0 11.7 35.0





#84
1,2,4-Trimethylbenzene
Concen: 49.020 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22




Instrument :
MSVOA_X
ClientSampleId :
SUMOGW-2-10-13MS




Tgt Ion:105 Resp: 241118
Ion Ratio Lower Upper
105 100
120 42.1 21.2 63.6

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024




#85
sec-Butylbenzene
Concen: 49.402 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22




Tgt Ion:105 Resp: 301856
Ion Ratio Lower Upper
105 100
134 18.8 9.8 29.4





#86
p-Isopropyltoluene
Concen: 50.654 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22




Instrument :
MSVOA_X
ClientSampleId :
SUMOGW-2-10-13MS




Tgt Ion:119 Resp: 24736
Ion Ratio Lower Upper
119 100
134 25.4 12.8 38.3
91 28.8 13.1 39.3

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024




#87
1,3-Dichlorobenzene
Concen: 48.296 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22




Tgt Ion:146 Resp: 119657
Ion Ratio Lower Upper
146 100
111 46.3 22.1 66.1
148 64.0 31.8 95.3





#88
1,4-Dichlorobenzene
Concen: 47.868 ug/l
RT: 12.036 min Scan# 1797
Delta R.T. -0.006 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22




Instrument :
MSVOA_X
ClientSampleId :
SUMOGW-2-10-13MS



Tgt Ion:146 Resp: 122489
Ion Ratio Lower Upper
146 100
111 43.6 21.3 63.7
148 64.3 31.8 95.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024



#89
n-Butylbenzene
Concen: 51.281 ug/l
RT: 12.329 min Scan# 1844
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22

Tgt Ion: 91 Resp: 241874
Ion Ratio Lower Upper
91 100
92 51.2 27.4 82.0
134 23.0 12.6 37.8









#90
Hexachloroethane
Concen: 50.085 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22

Instrument :
MSVOA_X
ClientSampleId :
SUMOGW-2-10-13MS

Tgt Ion:117 Resp: 41262
Ion Ratio Lower Upper
117 100
201 61.7 32.9 98.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024



#91
1,2-Dichlorobenzene
Concen: 47.823 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX039937.D
Acq: 24 Jan 2024 15:22

Tgt Ion:146 Resp: 117139
Ion Ratio Lower Upper
146 100
111 46.5 22.4 67.0
148 63.9 31.6 95.0





#92

1,2-Dibromo-3-Chloropropane

Concen: 50.707 ug/l

RT: 12.939 min Scan# 1944

Delta R.T. -0.006 min

Lab File: VX039937.D

Acq: 24 Jan 2024 15:22

Instrument :

MSVOA_X

ClientSampleId :





SUMOGW-2-10-13MS

Tgt Ion: 75	Resp: 2754
Ion Ratio	Lower Upper
75	100
155	66.0 36.1 108.5
157	87.1 44.9 134.7

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024



#93

1,2,4-Trichlorobenzene

Concen: 48.311 ug/l




RT: 13.585 min Scan# 2050


Delta R.T. 0.000 min

Lab File: VX039937.D

Acq: 24 Jan 2024 15:22

Tgt Ion:180	Resp: 72314
Ion Ratio	Lower Upper
180	100
182	93.9 48.3 144.8
145	34.9 18.0 54.0





#94

Hexachlorobutadiene

Concen: 49.243 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. 0.000 min

Lab File: VX039937.D

Acq: 24 Jan 2024 15:22

Instrument :

MSVOA_X

ClientSampleId :





SUMOGW-2-10-13MS

Tgt Ion:	225	Resp:	3064
Ion Ratio	Lower	Upper	
225	100		
223	61.8	31.1	93.2
227	62.9	32.5	97.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 01/25/2024
Supervised By :Semsettin Yesilyurt 01/25/2024



#95

Naphthalene

Concen: 47.097 ug/l




RT: 13.774 min Scan# 2081


Delta R.T. 0.000 min

Lab File: VX039937.D

Acq: 24 Jan 2024 15:22


Tgt Ion:	128	Resp:	259528
Ion Ratio	Lower	Upper	
128	100		
127	13.1	10.3	15.5
129	11.0	8.8	13.2





#96
 1,2,3-Trichlorobenzene
 Concen: 46.903 ug/l
 RT: 13.957 min Scan# 2111
 Delta R.T. -0.006 min
 Lab File: VX039937.D
 Acq: 24 Jan 2024 15:22

Instrument :
 MSVOA_X
 ClientSampleId :
 SUMOGW-2-10-13MS



Tgt Ion:180 Resp: 70850
 Ion Ratio Lower Upper
 180 100
 182 93.1 46.9 140.6
 145 36.1 17.8 53.5

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 01/25/2024
 Supervised By :Semsettin Yesilyurt 01/25/2024

