

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX012519\
 Data File : VX007223.D
 Acq On : 25 Jan 2019 12:20
 Operator : JC/SP
 Sample : K1221-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 2-AVE-30

Quant Time: Jan 28 10:43:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X012519W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Jan 28 10:35:22 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.02	128	83447	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.86	114	439629	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.11	117	392633	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	6.07	65	165614	30.23	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	100.77%
60) 4-Bromofluorobenzene	11.13	95	169866	28.03	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	93.43%
63) Toluene-d8	8.71	98	522558	29.53	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	98.43%

Target Compounds

					Ovalue
4) Vinyl Chloride	1.41	62	5994	1.261	ug/l 93
22) cis-1,2-Dichloroethene	4.60	96	88881	17.330	ug/l 88
37) Trichloroethene	7.21	130	16110	3.091	ug/l 99
61) Tetrachloroethene	9.34	164	45873	8.382	ug/l 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX012519\
Data File : VX007223.D
Acq On : 25 Jan 2019 12:20
Operator : JC/SP
Sample : K1221-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
2-AVE-30

Quant Time: Jan 28 10:43:54 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X012519W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Mon Jan 28 10:35:22 2019
Response via : Initial Calibration

