

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012521\
 Data File : VX020726.D
 Acq On : 25 Jan 2021 15:49
 Operator : JC/SP
 Sample : VX0125WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Jan 26 03:08:07 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:38:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	314013	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.824	114	498256	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	473360	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	241084	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.025	65	212434	49.60	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.20%
35) Dibromofluoromethane	5.458	113	171588	50.28	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.56%
50) Toluene-d8	8.696	98	624321	49.90	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.80%
62) 4-Bromofluorobenzene	11.116	95	245422	50.51	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	101.02%

Target Compounds					Qvalue
2) Dichlorodifluoromethane	1.185	85	93669	19.61	ug/l 100
3) Chloromethane	1.313	50	85433	19.34	ug/l 99
4) Vinyl Chloride	1.392	62	84517	19.19	ug/l 99
5) Bromomethane	1.624	94	50184	23.72	ug/l 95
6) Chloroethane	1.709	64	48954	19.64	ug/l 98
7) Trichlorofluoromethane	1.916	101	116708	19.54	ug/l 96
8) Diethyl Ether	2.166	74	42798	19.96	ug/l 96
9) 1,1,2-Trichlorotrifluo...	2.361	101	58559	19.20	ug/l 98
10) Methyl Iodide	2.489	142	78723	18.41	ug/l 92
11) Tert butyl alcohol	3.014	59	81928	102.46	ug/l 99
12) 1,1-Dichloroethene	2.355	96	57893	19.20	ug/l 94
13) Acrolein	2.270	56	56316	93.39	ug/l 99
14) Allyl chloride	2.703	41	105036	18.70	ug/l 95
15) Acrylonitrile	3.111	53	198727	98.19	ug/l 100
16) Acetone	2.416	43	146932	90.48	ug/l 99
17) Carbon Disulfide	2.550	76	178457	18.93	ug/l 99
18) Methyl Acetate	2.745	43	83243	20.06	ug/l 96
19) Methyl tert-butyl Ether	3.166	73	230298	20.41	ug/l 98
20) Methylene Chloride	2.831	84	77339	18.87	ug/l 96
21) trans-1,2-Dichloroethene	3.142	96	70964	18.84	ug/l 96
22) Diisopropyl ether	3.825	45	226061	20.48	ug/l 92
23) Vinyl Acetate	3.782	43	929185	102.31	ug/l 99
24) 1,1-Dichloroethane	3.666	63	129909	19.47	ug/l 99
25) 2-Butanone	4.629	43	244380	96.08	ug/l 98
26) 2,2-Dichloropropane	4.550	77	101894	19.03	ug/l 99
27) cis-1,2-Dichloroethene	4.556	96	81260	19.67	ug/l 99
28) Bromochloromethane	4.971	49	52523	18.36	ug/l 94
29) Tetrahydrofuran	5.080	42	161553	98.62	ug/l 95
30) Chloroform	5.172	83	135142	19.70	ug/l 99
31) Cyclohexane	5.544	56	109687	19.05	ug/l 100
32) 1,1,1-Trichloroethane	5.458	97	117527	19.47	ug/l 98
36) 1,1-Dichloropropene	5.763	75	97655	19.53	ug/l 98
37) Ethyl Acetate	4.788	43	95689	20.18	ug/l 100
38) Carbon Tetrachloride	5.745	117	104364	19.85	ug/l 96
39) Methylcyclohexane	7.434	83	112371	19.72	ug/l 98
40) Benzene	6.105	78	294167	19.97	ug/l 98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012521\
 Data File : VX020726.D
 Acq On : 25 Jan 2021 15:49
 Operator : JC/SP
 Sample : VX0125WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Jan 26 03:08:07 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:38:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.995	41	54458	20.52	ug/l	97
42) 1,2-Dichloroethane	6.153	62	106140	20.21	ug/l	98
43) Isopropyl Acetate	6.403	43	148596	19.66	ug/l	100
44) Trichloroethene	7.184	130	76492	19.52	ug/l	96
45) 1,2-Dichloropropane	7.482	63	77929	20.30	ug/l	98
46) Dibromomethane	7.635	93	52925	20.54	ug/l	96
47) Bromodichloromethane	7.873	83	108131	19.87	ug/l	99
48) Methyl methacrylate	7.739	41	75976	19.90	ug/l	97
49) 1,4-Dioxane	7.708	88	37239	415.77	ug/l	98
51) 4-Methyl-2-Pentanone	8.616	43	485846	101.86	ug/l	99
52) Toluene	8.763	92	183782	19.85	ug/l	99
53) t-1,3-Dichloropropene	9.019	75	113117	19.75	ug/l	97
54) cis-1,3-Dichloropropene	8.409	75	122947	19.91	ug/l	97
55) 1,1,2-Trichloroethane	9.196	97	76179	19.81	ug/l	98
56) Ethyl methacrylate	9.159	69	107465	19.27	ug/l	98
57) 1,3-Dichloropropane	9.348	76	130651	20.25	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.287	63	265623	99.84	ug/l	99
59) 2-Hexanone	9.470	43	362506	102.09	ug/l	98
60) Dibromochloromethane	9.561	129	84896	19.97	ug/l	100
61) 1,2-Dibromoethane	9.653	107	78874	19.82	ug/l	98
64) Tetrachloroethene	9.317	164	74441	20.51	ug/l	93
65) Chlorobenzene	10.116	112	197928	19.33	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.201	131	75950	19.98	ug/l	98
67) Ethyl Benzene	10.232	91	349442	19.68	ug/l	99
68) m/p-Xylenes	10.342	106	264714	39.98	ug/l	99
69) o-Xylene	10.683	106	124027	19.91	ug/l	95
70) Styrene	10.695	104	216140	20.21	ug/l	99
71) Bromoform	10.836	173	59888	19.62	ug/l #	99
73) Isopropylbenzene	11.000	105	341813	19.62	ug/l	99
74) N-amyl acetate	10.878	43	127833	18.74	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.250	83	116737	19.88	ug/l	98
76) 1,2,3-Trichloropropane	11.281	75	134652	24.08	ug/l	88
77) Bromobenzene	11.238	156	86050	19.51	ug/l	98
78) n-propylbenzene	11.342	91	391447	19.46	ug/l	100
79) 2-Chlorotoluene	11.402	91	243257	19.80	ug/l	99
80) 1,3,5-Trimethylbenzene	11.488	105	298582	20.05	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.055	75	36301	18.19	ug/l	95
82) 4-Chlorotoluene	11.494	91	283109	19.47	ug/l	98
83) tert-Butylbenzene	11.756	119	281580	19.58	ug/l	97
84) 1,2,4-Trimethylbenzene	11.793	105	303077	20.18	ug/l	98
85) sec-Butylbenzene	11.927	105	332643	20.01	ug/l	99
86) p-Isopropyltoluene	12.049	119	305939	19.83	ug/l	99
87) 1,3-Dichlorobenzene	12.006	146	155233	19.32	ug/l	98
88) 1,4-Dichlorobenzene	12.079	146	159825	19.15	ug/l	98
89) n-Butylbenzene	12.372	91	261386	18.92	ug/l	99
90) Hexachloroethane	12.579	117	52012	18.72	ug/l	95
91) 1,2-Dichlorobenzene	12.372	146	152466	19.69	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.981	75	24861	19.10	ug/l	87
93) 1,2,4-Trichlorobenzene	13.628	180	95134	18.94	ug/l	98
94) Hexachlorobutadiene	13.762	225	44339	18.74	ug/l	100
95) Naphthalene	13.817	128	301358	18.23	ug/l	100
96) 1,2,3-Trichlorobenzene	14.000	180	96431	19.72	ug/l	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012521\
Data File : VX020726.D
Acq On : 25 Jan 2021 15:49
Operator : JC/SP
Sample : VX0125WBSD01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Jan 26 03:08:07 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:38:15 2021
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012521\
 Data File : VX020726.D
 Acq On : 25 Jan 2021 15:49
 Operator : JC/SP
 Sample : VX0125WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Jan 26 03:08:07 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:38:15 2021
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 5.623 min Scan# 744
Delta R.T. -0.006 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:168 Resp: 314013
Ion Ratio Lower Upper
168 100
99 61.4 45.8 68.6



#2
Dichlorodifluoromethane
Concen: 19.61 ug/l
RT: 1.185 min Scan# 16
Delta R.T. -0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Tgt Ion: 85 Resp: 93669
Ion Ratio Lower Upper
85 100
87 32.6 16.4 49.2









#3
Chloromethane
Concen: 19.34 ug/l
RT: 1.313 min Scan# 37
Delta R.T. -0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 50 Resp: 85433
Ion Ratio Lower Upper
50 100
52 32.5 26.4 39.6



#4
Vinyl Chloride
Concen: 19.19 ug/l
RT: 1.392 min Scan# 50
Delta R.T. 0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Tgt Ion: 62 Resp: 84517
Ion Ratio Lower Upper
62 100
64 31.5 25.7 38.5









#5
Bromomethane
Concen: 23.72 ug/l
RT: 1.624 min Scan# 88
Delta R.T. -0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 94 Resp: 50184
Ion Ratio Lower Upper
94 100
96 91.0 76.4 114.6



#6
Chloroethane
Concen: 19.64 ug/l
RT: 1.709 min Scan# 102
Delta R.T. 0.006 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Tgt Ion: 64 Resp: 48954
Ion Ratio Lower Upper
64 100
66 32.9 25.4 38.0









#7
Trichlorofluoromethane
Concen: 19.54 ug/l
RT: 1.916 min Scan# 136
Delta R.T. 0.006 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:101 Resp: 116708
Ion Ratio Lower Upper
101 100
103 62.9 53.2 79.8



#8
Diethyl Ether
Concen: 19.96 ug/l
RT: 2.166 min Scan# 177
Delta R.T. 0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Tgt Ion: 74 Resp: 42798
Ion Ratio Lower Upper
74 100
45 92.6 44.5 133.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 19.20 ug/l
RT: 2.361 min Scan# 209
Delta R.T. 0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:101 Resp: 58559
Ion Ratio Lower Upper
101 100
85 44.6 34.9 52.3
151 75.4 61.4 92.2



#10
Methyl Iodide
Concen: 18.41 ug/l
RT: 2.489 min Scan# 230
Delta R.T. 0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Tgt Ion:142 Resp: 78723
Ion Ratio Lower Upper
142 100
127 47.7 33.3 49.9
141 14.8 11.4 17.0









#11
Tert butyl alcohol
Concen: 102.46 ug/l
RT: 3.014 min Scan# 316
Delta R.T. -0.006 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 59 Resp: 81928
Ion Ratio Lower Upper
59 100
57 9.6 7.8 11.8



#12
1,1-Dichloroethene
Concen: 19.20 ug/l
RT: 2.355 min Scan# 208
Delta R.T. 0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Tgt Ion: 96 Resp: 57893
Ion Ratio Lower Upper
96 100
61 164.1 124.6 187.0
98 66.6 50.6 76.0









#13
Acrolein
Concen: 93.39 ug/l
RT: 2.270 min Scan# 194
Delta R.T. -0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 56 Resp: 56316
Ion Ratio Lower Upper
56 100
55 69.6 56.4 84.6



#14
Allyl chloride
Concen: 18.70 ug/l
RT: 2.703 min Scan# 265
Delta R.T. -0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Tgt Ion: 41 Resp: 105036
Ion Ratio Lower Upper
41 100
39 58.9 51.6 77.4
76 36.5 30.5 45.7









#15
Acrylonitrile
Concen: 98.19 ug/l
RT: 3.111 min Scan# 332
Delta R.T. -0.006 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 53 Resp: 198727
Ion Ratio Lower Upper
53 100
52 81.4 65.3 97.9
51 34.8 28.2 42.2



#16
Acetone
Concen: 90.48 ug/l
RT: 2.416 min Scan# 218
Delta R.T. -0.006 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Tgt Ion: 43 Resp: 146932
Ion Ratio Lower Upper
43 100
58 35.5 27.8 41.8









#17
Carbon Disulfide
Concen: 18.93 ug/l
RT: 2.550 min Scan# 240
Delta R.T. 0.006 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 76 Resp: 178457
Ion Ratio Lower Upper
76 100
78 8.6 7.2 10.8



#18
Methyl Acetate
Concen: 20.06 ug/l
RT: 2.745 min Scan# 272
Delta R.T. 0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Tgt Ion: 43 Resp: 83243
Ion Ratio Lower Upper
43 100
74 25.1 21.9 32.9









#19
Methyl tert-butyl Ether
Concen: 20.41 ug/l
RT: 3.166 min Scan# 341
Delta R.T. 0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 73 Resp: 230298
Ion Ratio Lower Upper
73 100
57 20.6 17.1 25.7



#20
Methylene Chloride
Concen: 18.87 ug/l
RT: 2.831 min Scan# 286
Delta R.T. -0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49


Tgt Ion: 84 Resp: 77339
Ion Ratio Lower Upper
84 100
49 114.9 88.6 132.8
51 34.9 27.0 40.6
86 60.1 50.3 75.5








#21
trans-1,2-Dichloroethene
Concen: 18.84 ug/l
RT: 3.142 min Scan# 337
Delta R.T. -0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Instrument :
MSVOA_X
ClientSampled :







Tgt Ion: 96 Resp: 70964
Ion Ratio Lower Upper
96 100
61 129.6 107.7 161.5
98 59.9 50.4 75.6



#22
Diisopropyl ether
Concen: 20.48 ug/l
RT: 3.825 min Scan# 449
Delta R.T. -0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Tgt Ion: 45 Resp: 226061
Ion Ratio Lower Upper
45 100
43 63.0 58.0 87.0
87 29.0 25.4 38.0
59 12.3 10.2 15.2









#23
 Vinyl Acetate
 Concen: 102.31 ug/l
 RT: 3.782 min Scan# 442
 Delta R.T. -0.000 min
 Lab File: VX020726.D
 Acq: 25 Jan 2021 15:49




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion: 43 Resp: 929185
 Ion Ratio Lower Upper
 43 100
 86 12.0 9.9 14.9



#24
 1,1-Dichloroethane
 Concen: 19.47 ug/l
 RT: 3.666 min Scan# 423
 Delta R.T. 0.000 min
 Lab File: VX020726.D
 Acq: 25 Jan 2021 15:49

Tgt Ion: 63 Resp: 129909
 Ion Ratio Lower Upper
 63 100
 98 7.0 3.8 11.2
 100 4.7 2.5 7.5









#25
2-Butanone
Concen: 96.08 ug/l
RT: 4.629 min Scan# 581
Delta R.T. -0.006 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 43 Resp: 244380
Ion Ratio Lower Upper
43 100
72 28.5 23.7 35.5



#26
2,2-Dichloropropane
Concen: 19.03 ug/l
RT: 4.550 min Scan# 568
Delta R.T. 0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Tgt Ion: 77 Resp: 101894
Ion Ratio Lower Upper
77 100
97 23.1 11.7 35.1





#27

cis-1,2-Dichloroethene

Concen: 19.67 ug/l

RT: 4.556 min Scan# 569





Delta R.T. -0.006 min

Lab File: VX020726.D

Acq: 25 Jan 2021 15:49

Tgt Ion:	96	Resp:	81260
Ion	Ratio	Lower	Upper
96	100		
61	143.0	0.0	285.6
98	63.6	0.0	129.6

Instrument :
MSVOA_X
ClientSampled :



#28

Bromochloromethane

Concen: 18.36 ug/l




RT: 4.971 min Scan# 637


Delta R.T. -0.006 min

Lab File: VX020726.D

Acq: 25 Jan 2021 15:49

Tgt Ion:	49	Resp:	52523
Ion	Ratio	Lower	Upper
49	100		
129	2.2	0.0	4.0
130	79.9	68.2	102.2





#29

Tetrahydrofuran

Concen: 98.62 ug/l

RT: 5.080 min Scan# 655

Delta R.T. -0.006 min

Lab File: VX020726.D

Acq: 25 Jan 2021 15:49





Tgt Ion: 42 Resp: 161553

Ion Ratio Lower Upper

42 100

72 48.4 41.7 62.5

71 45.3 38.4 57.6



#30

Chloroform

Concen: 19.70 ug/l

RT: 5.172 min Scan# 670

Delta R.T. -0.006 min

Lab File: VX020726.D




Acq: 25 Jan 2021 15:49


Tgt Ion: 83 Resp: 135142

Ion Ratio Lower Upper

83 100


85 63.8 52.0 78.0








#31
Cyclohexane
Concen: 19.05 ug/l
RT: 5.544 min Scan# 731
Delta R.T. -0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 56 Resp: 109687
Ion Ratio Lower Upper
56 100
69 31.9 25.3 37.9
84 88.8 71.4 107.0



#32
1,1,1-Trichloroethane
Concen: 19.47 ug/l
RT: 5.458 min Scan# 717
Delta R.T. 0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Tgt Ion: 97 Resp: 117527
Ion Ratio Lower Upper
97 100
99 65.9 51.5 77.3
61 41.9 34.2 51.4





#33

1,2-Dichloroethane-d4

Concen: 49.60 ug/l

RT: 6.025 min Scan# 810

Delta R.T. 0.000 min

Lab File: VX020726.D





Acq: 25 Jan 2021 15:49

Tgt Ion: 65 Resp: 212434

Ion Ratio Lower Upper

65 100

67 52.0 0.0 105.4



#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 6.824 min Scan# 941

Delta R.T. -0.006 min

Lab File: VX020726.D

Acq: 25 Jan 2021 15:49




Tgt Ion: 114 Resp: 498256


Ion Ratio Lower Upper

114 100

63 21.6 0.0 40.8

88 17.0 0.0 32.8









#35
Dibromofluoromethane
Concen: 50.28 ug/l
RT: 5.458 min Scan# 717
Delta R.T. -0.006 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:113 Resp: 171588
Ion Ratio Lower Upper
113 100
111 103.8 81.9 122.9
192 19.6 16.7 25.1



#36
1,1-Dichloropropene
Concen: 19.53 ug/l
RT: 5.763 min Scan# 767
Delta R.T. -0.006 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49


Tgt Ion: 75 Resp: 97655
Ion Ratio Lower Upper
75 100
110 34.7 18.1 54.4
77 31.4 25.3 37.9








#37
Ethyl Acetate
Concen: 20.18 ug/l
RT: 4.788 min Scan# 607
Delta R.T. -0.006 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49


Instrument :
MSVOA_X
ClientSampled :





Tgt Ion: 43 Resp: 95689
Ion Ratio Lower Upper
43 100
61 15.8 12.8 19.2
70 12.9 10.5 15.7




#38
Carbon Tetrachloride
Concen: 19.85 ug/l
RT: 5.745 min Scan# 764
Delta R.T. -0.006 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49



Tgt Ion: 117 Resp: 104364
Ion Ratio Lower Upper
117 100
119 98.6 76.2 114.2
121 28.8 24.8 37.2









#39
Methylcyclohexane
Concen: 19.72 ug/l
RT: 7.434 min Scan# 1041
Delta R.T. -0.006 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 83 Resp: 112371
Ion Ratio Lower Upper
83 100
55 74.2 58.9 88.3
98 45.5 38.1 57.1



#40
Benzene
Concen: 19.97 ug/l
RT: 6.105 min Scan# 823
Delta R.T. -0.006 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Tgt Ion: 78 Resp: 294167
Ion Ratio Lower Upper
78 100
77 22.7 18.9 28.3









#41
Methacrylonitrile
Concen: 20.52 ug/l
RT: 4.995 min Scan# 641
Delta R.T. -0.006 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:	41	Resp:	54458
Ion Ratio	Lower	Upper	
41	100		
39	49.6	42.5	63.7
67	78.3	65.0	97.6
52	31.0	25.3	37.9



#42
1,2-Dichloroethane
Concen: 20.21 ug/l
RT: 6.153 min Scan# 831
Delta R.T. -0.006 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Tgt Ion:	62	Resp:	106140
Ion Ratio	Lower	Upper	
62	100		
98	9.5	0.0	20.2









#43
Isopropyl Acetate
Concen: 19.66 ug/l
RT: 6.403 min Scan# 872
Delta R.T. -0.006 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 43 Resp: 148596
Ion Ratio Lower Upper
43 100
61 22.5 18.2 27.4
87 15.5 12.5 18.7



#44
Trichloroethene
Concen: 19.52 ug/l
RT: 7.184 min Scan# 1000
Delta R.T. -0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Tgt Ion:130 Resp: 76492
Ion Ratio Lower Upper
130 100
95 94.9 0.0 198.0





#45

1,2-Dichloropropane

Concen: 20.30 ug/l

RT: 7.482 min Scan# 1049

Delta R.T. -0.006 min

Lab File: VX020726.D





Acq: 25 Jan 2021 15:49

Tgt Ion: 63 Resp: 77929

Ion Ratio Lower Upper

63 100

65 30.7 25.7 38.5



#46

Dibromomethane

Concen: 20.54 ug/l

RT: 7.635 min Scan# 1074

Delta R.T. -0.000 min

Lab File: VX020726.D

Acq: 25 Jan 2021 15:49




Tgt Ion: 93 Resp: 52925


Ion Ratio Lower Upper

93 100

95 82.4 66.8 100.2

174 96.7 83.1 124.7









#47
Bromodichloromethane
Concen: 19.87 ug/l
RT: 7.873 min Scan# 1113
Delta R.T. -0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 83 Resp: 108131
Ion Ratio Lower Upper
83 100
85 63.6 51.4 77.0
127 9.3 6.8 10.2



#48
Methyl methacrylate
Concen: 19.90 ug/l
RT: 7.739 min Scan# 1091
Delta R.T. -0.006 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Tgt Ion: 41 Resp: 75976
Ion Ratio Lower Upper
41 100
69 94.5 76.0 114.0
39 51.4 44.6 67.0









#49
1,4-Dioxane
Concen: 415.77 ug/l
RT: 7.708 min Scan# 1086
Delta R.T. -0.006 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 88 Resp: 37239
Ion Ratio Lower Upper
88 100
43 30.3 23.2 34.8
58 70.5 55.0 82.4



#50
Toluene-d8
Concen: 49.90 ug/l
RT: 8.696 min Scan# 1248
Delta R.T. -0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Tgt Ion: 98 Resp: 624321
Ion Ratio Lower Upper
98 100
100 66.8 51.9 77.9









#51
4-Methyl-2-Pentanone
Concen: 101.86 ug/l
RT: 8.616 min Scan# 1235
Delta R.T. 0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 43 Resp: 485846
Ion Ratio Lower Upper
43 100
58 43.6 35.4 53.0



#52
Toluene
Concen: 19.85 ug/l
RT: 8.763 min Scan# 1259
Delta R.T. -0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Tgt Ion: 92 Resp: 183782
Ion Ratio Lower Upper
92 100
91 172.2 137.1 205.7









#53
t-1,3-Dichloropropene
Concen: 19.75 ug/l
RT: 9.019 min Scan# 1301
Delta R.T. -0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 75 Resp: 113117
Ion Ratio Lower Upper
75 100
77 32.7 24.9 37.3



#54
cis-1,3-Dichloropropene
Concen: 19.91 ug/l
RT: 8.409 min Scan# 1201
Delta R.T. -0.006 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Tgt Ion: 75 Resp: 122947
Ion Ratio Lower Upper
75 100
77 31.2 25.7 38.5
39 41.8 35.1 52.7









#55
1,1,2-Trichloroethane
Concen: 19.81 ug/l
RT: 9.196 min Scan# 1330
Delta R.T. 0.001 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 97	Resp: 76179
Ion Ratio	Lower Upper
97 100	
83 84.6	69.3 103.9
85 53.1	44.3 66.5
99 62.5	50.3 75.5



#56
Ethyl methacrylate
Concen: 19.27 ug/l
RT: 9.159 min Scan# 1324
Delta R.T. -0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Tgt Ion: 69	Resp: 107465
Ion Ratio	Lower Upper
69 100	
41 59.9	46.8 70.2
39 29.8	24.7 37.1









#57
1,3-Dichloropropane
Concen: 20.25 ug/l
RT: 9.348 min Scan# 1355
Delta R.T. -0.006 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 76 Resp: 130651
Ion Ratio Lower Upper
76 100
78 32.8 26.2 39.2



#58
2-Chloroethyl Vinyl ether
Concen: 99.84 ug/l
RT: 8.287 min Scan# 1181
Delta R.T. 0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Tgt Ion: 63 Resp: 265623
Ion Ratio Lower Upper
63 100
106 28.6 23.5 35.3









#59
2-Hexanone
Concen: 102.09 ug/l
RT: 9.470 min Scan# 1375
Delta R.T. -0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 43 Resp: 362506
Ion Ratio Lower Upper
43 100
58 60.5 30.9 92.8



#60
Dibromochloromethane
Concen: 19.97 ug/l
RT: 9.561 min Scan# 1390
Delta R.T. 0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Tgt Ion: 129 Resp: 84896
Ion Ratio Lower Upper
129 100
127 76.9 38.4 115.2









#61
1,2-Dibromoethane
Concen: 19.82 ug/l
RT: 9.653 min Scan# 1405
Delta R.T. -0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:107 Resp: 78874
Ion Ratio Lower Upper
107 100
109 96.4 75.8 113.6



#62
4-Bromofluorobenzene
Concen: 50.51 ug/l
RT: 11.116 min Scan# 1645
Delta R.T. -0.006 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Tgt Ion: 95 Resp: 245422
Ion Ratio Lower Upper
95 100
174 75.5 0.0 154.6
176 74.1 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.092 min Scan# 1477
Delta R.T. -0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:117 Resp: 473360
Ion Ratio Lower Upper
117 100
82 58.0 47.4 71.2
119 31.2 26.6 39.8



#64
Tetrachloroethene
Concen: 20.51 ug/l
RT: 9.317 min Scan# 1350
Delta R.T. 0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Tgt Ion:164 Resp: 74441
Ion Ratio Lower Upper
164 100
166 116.5 103.4 155.2
129 94.5 70.8 106.2
131 89.8 72.5 108.7









#65
Chlorobenzene
Concen: 19.33 ug/l
RT: 10.116 min Scan# 1481
Delta R.T. -0.006 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:112 Resp: 197928
Ion Ratio Lower Upper
112 100
114 33.0 25.8 38.6



#66
1,1,1,2-Tetrachloroethane
Concen: 19.98 ug/l
RT: 10.201 min Scan# 1495
Delta R.T. 0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Tgt Ion:131 Resp: 75950
Ion Ratio Lower Upper
131 100
133 93.5 47.7 143.1
119 62.7 32.1 96.5









#67
Ethyl Benzene
Concen: 19.68 ug/l
RT: 10.232 min Scan# 1500
Delta R.T. -0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 349442
Ion Ratio Lower Upper
91 100
106 30.0 24.6 37.0



#68
m/p-Xylenes
Concen: 39.98 ug/l
RT: 10.342 min Scan# 1518
Delta R.T. -0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Tgt Ion:106 Resp: 264714
Ion Ratio Lower Upper
106 100
91 205.8 162.8 244.2









#69
o-Xylene
Concen: 19.91 ug/l
RT: 10.683 min Scan# 1574
Delta R.T. 0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:106 Resp: 124027
Ion Ratio Lower Upper
106 100
91 224.0 108.1 324.2



#70
Styrene
Concen: 20.21 ug/l
RT: 10.695 min Scan# 1576
Delta R.T. 0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Tgt Ion:104 Resp: 216140
Ion Ratio Lower Upper
104 100
78 50.7 40.0 60.0
103 53.6 43.4 65.0









#71
Bromoform
Concen: 19.62 ug/l
RT: 10.836 min Scan# 1599
Delta R.T. 0.001 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:173 Resp: 59888
Ion Ratio Lower Upper
173 100
175 48.4 23.9 71.8
254 0.0 0.1 0.1#



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.061 min Scan# 1800
Delta R.T. -0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Tgt Ion:152 Resp: 241084
Ion Ratio Lower Upper
152 100
115 70.4 41.1 123.3
150 163.2 0.0 349.0









#73
Isopropylbenzene
Concen: 19.62 ug/l
RT: 11.000 min Scan# 1626
Delta R.T. 0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:105 Resp: 341813
Ion Ratio Lower Upper
105 100
120 26.1 13.2 39.6



#74
N-ethyl acetate
Concen: 18.74 ug/l
RT: 10.878 min Scan# 1606
Delta R.T. 0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Tgt Ion: 43 Resp: 127833
Ion Ratio Lower Upper
43 100
70 48.7 39.0 58.6
55 28.3 21.5 32.3
61 27.0 22.1 33.1









#75
1,1,2,2-Tetrachloroethane
Concen: 19.88 ug/l
RT: 11.250 min Scan# 1667
Delta R.T. 0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 83 Resp: 116737
Ion Ratio Lower Upper
83 100
131 10.1 5.1 15.2
85 64.9 31.8 95.4



#76
1,2,3-Trichloropropane
Concen: 24.08 ug/l
RT: 11.281 min Scan# 1672
Delta R.T. 0.001 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Tgt Ion: 75 Resp: 134652
Ion Ratio Lower Upper
75 100
77 31.1 19.1 57.5









#77
Bromobenzene
Concen: 19.51 ug/l
RT: 11.238 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:156 Resp: 86050
Ion Ratio Lower Upper
156 100
77 166.7 82.8 248.6
158 94.5 49.1 147.4



#78
n-propylbenzene
Concen: 19.46 ug/l
RT: 11.342 min Scan# 1682
Delta R.T. 0.001 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Tgt Ion: 91 Resp: 391447
Ion Ratio Lower Upper
91 100
120 22.8 11.5 34.5









#79
2-Chlorotoluene
Concen: 19.80 ug/l
RT: 11.402 min Scan# 1692
Delta R.T. 0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 91 Resp: 243257
Ion Ratio Lower Upper
91 100
126 32.5 16.6 49.8



#80
1,3,5-Trimethylbenzene
Concen: 20.05 ug/l
RT: 11.488 min Scan# 1706
Delta R.T. -0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Tgt Ion: 105 Resp: 298582
Ion Ratio Lower Upper
105 100
120 46.9 23.8 71.5









#81
trans-1,4-Dichloro-2-butene
Concen: 18.19 ug/l
RT: 11.055 min Scan# 1635
Delta R.T. -0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 75 Resp: 36301
Ion Ratio Lower Upper
75 100
53 99.2 74.5 111.7
89 45.5 37.9 56.9



#82
4-Chlorotoluene
Concen: 19.47 ug/l
RT: 11.494 min Scan# 1707
Delta R.T. -0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Tgt Ion: 91 Resp: 283109
Ion Ratio Lower Upper
91 100
126 28.4 14.9 44.5









#83
tert-Butylbenzene
Concen: 19.58 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. 0.006 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:119 Resp: 281580
Ion Ratio Lower Upper
119 100
91 58.5 27.9 83.7
134 22.6 11.3 33.9



#84
1,2,4-Trimethylbenzene
Concen: 20.18 ug/l
RT: 11.793 min Scan# 1756
Delta R.T. -0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Tgt Ion:105 Resp: 303077
Ion Ratio Lower Upper
105 100
120 43.4 22.4 67.3









#85
 sec-Butylbenzene
 Concen: 20.01 ug/l
 RT: 11.927 min Scan# 1778
 Delta R.T. -0.000 min
 Lab File: VX020726.D
 Acq: 25 Jan 2021 15:49




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion:105 Resp: 332643
 Ion Ratio Lower Upper
 105 100
 134 19.3 10.0 29.8



#86
 p-Isopropyltoluene
 Concen: 19.83 ug/l
 RT: 12.049 min Scan# 1798
 Delta R.T. -0.000 min
 Lab File: VX020726.D
 Acq: 25 Jan 2021 15:49

Tgt Ion:119 Resp: 305939
 Ion Ratio Lower Upper
 119 100
 134 25.7 12.9 38.7
 91 24.3 11.7 35.0









#87
1,3-Dichlorobenzene
Concen: 19.32 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:146 Resp: 155233
Ion Ratio Lower Upper
146 100
111 41.6 20.2 60.5
148 62.8 31.9 95.7



#88
1,4-Dichlorobenzene
Concen: 19.15 ug/l
RT: 12.079 min Scan# 1803
Delta R.T. 0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Tgt Ion:146 Resp: 159825
Ion Ratio Lower Upper
146 100
111 41.7 19.6 58.8
148 62.2 31.6 95.0









#89
n-Butylbenzene
Concen: 18.92 ug/l
RT: 12.372 min Scan# 1851
Delta R.T. -0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 261386
Ion Ratio Lower Upper
91 100
92 54.6 27.4 82.0
134 25.2 13.2 39.5



#90
Hexachloroethane
Concen: 18.72 ug/l
RT: 12.579 min Scan# 1885
Delta R.T. 0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Tgt Ion: 117 Resp: 52012
Ion Ratio Lower Upper
117 100
201 76.0 40.4 121.2









#91
1,2-Dichlorobenzene
Concen: 19.69 ug/l
RT: 12.372 min Scan# 1851
Delta R.T. -0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:146 Resp: 152466
Ion Ratio Lower Upper
146 100
111 43.8 20.8 62.2
148 62.4 31.9 95.5



#92
1,2-Dibromo-3-Chloropropane
Concen: 19.10 ug/l
RT: 12.981 min Scan# 1951
Delta R.T. 0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Tgt Ion: 75 Resp: 24861
Ion Ratio Lower Upper
75 100
155 81.3 45.3 135.9
157 98.0 57.7 173.1









#93
1,2,4-Trichlorobenzene
Concen: 18.94 ug/l
RT: 13.628 min Scan# 2057
Delta R.T. -0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:180 Resp: 95134
Ion Ratio Lower Upper
180 100
182 93.8 47.7 143.1
145 32.3 15.0 45.0



#94
Hexachlorobutadiene
Concen: 18.74 ug/l
RT: 13.762 min Scan# 2079
Delta R.T. -0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Tgt Ion:225 Resp: 44339
Ion Ratio Lower Upper
225 100
223 62.6 31.1 93.2
227 63.5 31.6 95.0









#95
Naphthalene
Concen: 18.23 ug/l
RT: 13.817 min Scan# 2088
Delta R.T. -0.000 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Instrument :
MSVOA_X
ClientSampled :

Tgt Ion:128 Resp: 301358
Ion Ratio Lower Upper
128 100
127 12.8 10.3 15.5
129 10.9 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 19.72 ug/l
RT: 14.000 min Scan# 2118
Delta R.T. 0.001 min
Lab File: VX020726.D
Acq: 25 Jan 2021 15:49

Tgt Ion:180 Resp: 96431
Ion Ratio Lower Upper
180 100
182 90.3 47.6 142.9
145 34.0 16.1 48.3

