


Dichlorodifluoromethane


Response Ratio



Response = 8.099e-001 * Amt - 1.942e-002
 Coef of Det (r^2) = 0.999797 Curve Fit: Linear
 Method Name: Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\82X012521W.M
 Calibration Table Last Updated: Mon Jan 25 13:38:15 2021

Carbon Disulfide

Response Ratio



$$\text{Response} = 1.576\text{e}+000 * \text{Amt} - 2.840\text{e}-002$$


Coef of Det (r^2) = 0.999620 Curve Fit: Linear

Method Name: Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\82X012521W.M

Calibration Table Last Updated: Mon Jan 25 13:38:15 2021

Ethyl methacrylate

Response Ratio



$$\text{Response} = 6.539\text{e-}001 * \text{Amt} - 3.641\text{e-}002$$


Coef of Det (r^2) = 0.998641 Curve Fit: Linear

Method Name: Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\82X012521W.M

Calibration Table Last Updated: Mon Jan 25 13:38:15 2021

N-amyl acetate

Response Ratio



$$\text{Response} = 1.585\text{e}+000 * \text{Amt} - 6.385\text{e}-002$$


Coef of Det (r^2) = 0.999046 Curve Fit: Linear

Method Name: Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\82X012521W.M

Calibration Table Last Updated: Mon Jan 25 13:38:15 2021

Naphthalene

Response Ratio



$$\text{Response} = 4.070\text{e}+000 * \text{Amt} - 2.338\text{e}-001$$

Coef of Det (r^2) = 0.998425 Curve Fit: Linear

Method Name: Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\82X012521W.M

Calibration Table Last Updated: Mon Jan 25 13:38:15 2021