

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012521\
 Data File : VX020723.D
 Acq On : 25 Jan 2021 13:41
 Operator : JC/SP
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX012521

Quant Time: Jan 26 03:06:34 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:38:15 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	108	0.00
2 T	Dichlorodifluoromethane	50.000	49.554	0.9	107	0.00
3 P	Chloromethane	50.000	50.797	-1.6	105	0.00
4 C	Vinyl Chloride	50.000	50.180	-0.4#	103	0.00
5 T	Bromomethane	50.000	48.273	3.5	111	0.00
6 T	Chloroethane	50.000	50.896	-1.8	103	0.00
7 T	Trichlorofluoromethane	50.000	50.385	-0.8	106	0.00
8 T	Diethyl Ether	50.000	50.460	-0.9	106	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	50.264	-0.5	107	0.00
10 T	Methyl Iodide	50.000	47.420	5.2	97	0.00
11 T	Tert butyl alcohol	250.000	227.904	8.8	97	0.00
12 CM	1,1-Dichloroethene	50.000	49.525	1.0#	106	0.00
13 T	Acrolein	250.000	231.681	7.3	97	0.00
14 T	Allyl chloride	50.000	49.384	1.2	104	0.00
15 T	Acrylonitrile	250.000	238.973	4.4	99	0.00
16 T	Acetone	250.000	230.771	7.7	100	0.00
17 T	Carbon Disulfide	50.000	48.104	3.8	106	0.00
18 T	Methyl Acetate	50.000	48.809	2.4	99	0.00
19 T	Methyl tert-butyl Ether	50.000	51.400	-2.8	104	0.00
20 T	Methylene Chloride	50.000	47.469	5.1	104	0.00
21 T	trans-1,2-Dichloroethene	50.000	48.019	4.0	103	0.00
22 T	Diisopropyl ether	50.000	50.904	-1.8	103	0.00
23 T	Vinyl Acetate	250.000	258.085	-3.2	103	0.00
24 P	1,1-Dichloroethane	50.000	49.539	0.9	103	0.00
25 T	2-Butanone	250.000	240.817	3.7	101	0.00
26 T	2,2-Dichloropropane	50.000	50.463	-0.9	107	0.00
27 T	cis-1,2-Dichloroethene	50.000	49.828	0.3	104	0.00
28 T	Bromochloromethane	50.000	44.407	11.2	103	0.00
29 T	Tetrahydrofuran	250.000	237.745	4.9	97	0.00
30 C	Chloroform	50.000	49.581	0.8#	103	0.00
31 T	Cyclohexane	50.000	50.526	-1.1	106	0.00
32 T	1,1,1-Trichloroethane	50.000	49.839	0.3	103	0.00
33 S	1,2-Dichloroethane-d4	50.000	44.997	10.0	100	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	104	0.00
35 S	Dibromofluoromethane	50.000	47.028	5.9	100	0.00
36 T	1,1-Dichloropropene	50.000	50.385	-0.8	105	0.00
37 T	Ethyl Acetate	50.000	49.317	1.4	99	0.00
38 T	Carbon Tetrachloride	50.000	51.151	-2.3	104	0.00
39 T	Methylcyclohexane	50.000	53.609	-7.2	108	0.00
40 TM	Benzene	50.000	50.575	-1.2	103	0.00
41 T	Methacrylonitrile	50.000	49.207	1.6	99	0.00
42 TM	1,2-Dichloroethane	50.000	50.115	-0.2	102	0.00
43 T	Isopropyl Acetate	50.000	50.684	-1.4	102	0.00
44 TM	Trichloroethene	50.000	50.293	-0.6	104	0.00
45 C	1,2-Dichloropropane	50.000	50.274	-0.5#	102	0.00
46 T	Dibromomethane	50.000	50.013	-0.0	102	0.00
47 T	Bromodichloromethane	50.000	50.212	-0.4	103	0.00
48 T	Methyl methacrylate	50.000	51.573	-3.1	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012521\
 Data File : VX020723.D
 Acq On : 25 Jan 2021 13:41
 Operator : JC/SP
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX012521

Quant Time: Jan 26 03:06:34 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:38:15 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	992.085	0.8	96	0.00
50 S	Toluene-d8	50.000	47.429	5.1	102	0.00
51 T	4-Methyl-2-Pentanone	250.000	255.724	-2.3	99	0.00
52 CM	Toluene	50.000	51.324	-2.6#	105	0.00
53 T	t-1,3-Dichloropropene	50.000	51.831	-3.7	105	0.00
54 T	cis-1,3-Dichloropropene	50.000	52.282	-4.6	105	0.00
55 T	1,1,2-Trichloroethane	50.000	50.959	-1.9	105	0.00
56 T	Ethyl methacrylate	50.000	47.304	5.4	105	0.00
57 T	1,3-Dichloropropane	50.000	50.971	-1.9	104	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	255.507	-2.2	101	0.00
59 T	2-Hexanone	250.000	255.648	-2.3	97	0.00
60 T	Dibromochloromethane	50.000	51.122	-2.2	104	0.00
61 T	1,2-Dibromoethane	50.000	50.957	-1.9	104	0.00
62 S	4-Bromofluorobenzene	50.000	47.658	4.7	100	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	107	0.00
64 T	Tetrachloroethene	50.000	49.549	0.9	102	0.00
65 PM	Chlorobenzene	50.000	50.551	-1.1	107	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	51.352	-2.7	106	0.00
67 C	Ethyl Benzene	50.000	52.420	-4.8#	106	0.00
68 T	m/p-Xylenes	100.000	105.975	-6.0	107	0.00
69 T	o-Xylene	50.000	52.905	-5.8	106	0.00
70 T	Styrene	50.000	53.651	-7.3	104	0.00
71 P	Bromoform	50.000	50.607	-1.2	102	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	104	0.00
73 T	Isopropylbenzene	50.000	53.297	-6.6	105	0.00
74 T	N-amyl acetate	50.000	47.005	6.0	101	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	48.391	3.2	99	0.00
76 T	1,2,3-Trichloropropane	50.000	49.801	0.4	99	0.00
77 T	Bromobenzene	50.000	51.497	-3.0	107	0.00
78 T	n-propylbenzene	50.000	53.874	-7.7	106	0.00
79 T	2-Chlorotoluene	50.000	52.937	-5.9	105	0.00
80 T	1,3,5-Trimethylbenzene	50.000	54.155	-8.3	104	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	50.039	-0.1	102	0.00
82 T	4-Chlorotoluene	50.000	52.361	-4.7	104	0.00
83 T	tert-Butylbenzene	50.000	52.754	-5.5	102	0.00
84 T	1,2,4-Trimethylbenzene	50.000	54.614	-9.2	105	0.00
85 T	sec-Butylbenzene	50.000	55.270	-10.5	106	0.00
86 T	p-Isopropyltoluene	50.000	54.743	-9.5	105	0.00
87 T	1,3-Dichlorobenzene	50.000	49.514	1.0	100	0.00
88 T	1,4-Dichlorobenzene	50.000	48.315	3.4	103	0.00
89 T	n-Butylbenzene	50.000	54.708	-9.4	105	0.00
90 T	Hexachloroethane	50.000	52.265	-4.5	104	0.00
91 T	1,2-Dichlorobenzene	50.000	50.238	-0.5	103	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	46.937	6.1	99	0.00
93 T	1,2,4-Trichlorobenzene	50.000	53.058	-6.1	107	0.00
94 T	Hexachlorobutadiene	50.000	49.692	0.6	105	0.00
95 T	Naphthalene	50.000	46.787	6.4	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012521\
Data File : VX020723.D
Acq On : 25 Jan 2021 13:41
Operator : JC/SP
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX012521

Quant Time: Jan 26 03:06:34 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:38:15 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	53.495	-7.0	108	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6