

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012521\
 Data File : VX020739.D
 Acq On : 25 Jan 2021 20:53
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 26 03:13:07 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:38:15 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	105	0.00
2 T	Dichlorodifluoromethane	50.000	47.328	5.3	99	0.00
3 P	Chloromethane	50.000	50.953	-1.9	103	0.00
4 C	Vinyl Chloride	50.000	51.851	-3.7#	103	0.00
5 T	Bromomethane	50.000	46.358	7.3	104	0.00
6 T	Chloroethane	50.000	52.090	-4.2	103	0.00
7 T	Trichlorofluoromethane	50.000	50.296	-0.6	103	0.00
8 T	Diethyl Ether	50.000	52.658	-5.3	107	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	48.206	3.6	100	0.00
10 T	Methyl Iodide	50.000	50.054	-0.1	100	0.00
11 T	Tert butyl alcohol	250.000	255.377	-2.2	105	0.00
12 CM	1,1-Dichloroethene	50.000	50.660	-1.3#	105	0.00
13 T	Acrolein	250.000	273.257	-9.3	111	0.00
14 T	Allyl chloride	50.000	50.906	-1.8	105	0.00
15 T	Acrylonitrile	250.000	264.811	-5.9	107	0.00
16 T	Acetone	250.000	251.731	-0.7	106	0.00
17 T	Carbon Disulfide	50.000	47.279	5.4	101	0.00
18 T	Methyl Acetate	50.000	54.832	-9.7	109	0.00
19 T	Methyl tert-butyl Ether	50.000	53.970	-7.9	107	0.00
20 T	Methylene Chloride	50.000	48.878	2.2	105	0.00
21 T	trans-1,2-Dichloroethene	50.000	49.123	1.8	103	0.00
22 T	Diisopropyl ether	50.000	52.956	-5.9	105	0.00
23 T	Vinyl Acetate	250.000	271.503	-8.6	106	0.00
24 P	1,1-Dichloroethane	50.000	51.402	-2.8	105	0.00
25 T	2-Butanone	250.000	264.466	-5.8	108	0.00
26 T	2,2-Dichloropropane	50.000	46.942	6.1	97	0.00
27 T	cis-1,2-Dichloroethene	50.000	51.343	-2.7	104	0.00
28 T	Bromochloromethane	50.000	47.388	5.2	107	0.00
29 T	Tetrahydrofuran	250.000	270.381	-8.2	107	0.00
30 C	Chloroform	50.000	50.575	-1.2#	103	0.00
31 T	Cyclohexane	50.000	51.782	-3.6	106	0.00
32 T	1,1,1-Trichloroethane	50.000	51.583	-3.2	104	0.00
33 S	1,2-Dichloroethane-d4	50.000	48.827	2.3	105	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	108	0.00
35 S	Dibromofluoromethane	50.000	48.051	3.9	106	0.00
36 T	1,1-Dichloropropene	50.000	48.641	2.7	105	0.00
37 T	Ethyl Acetate	50.000	51.990	-4.0	108	0.00
38 T	Carbon Tetrachloride	50.000	48.925	2.2	103	0.00
39 T	Methylcyclohexane	50.000	49.431	1.1	103	0.00
40 TM	Benzene	50.000	50.020	-0.0	106	0.00
41 T	Methacrylonitrile	50.000	52.510	-5.0	110	0.00
42 TM	1,2-Dichloroethane	50.000	49.269	1.5	104	0.00
43 T	Isopropyl Acetate	50.000	52.380	-4.8	110	0.00
44 TM	Trichloroethene	50.000	48.388	3.2	103	0.00
45 C	1,2-Dichloropropane	50.000	49.467	1.1#	104	0.00
46 T	Dibromomethane	50.000	48.851	2.3	103	0.00
47 T	Bromodichloromethane	50.000	48.247	3.5	102	0.00
48 T	Methyl methacrylate	50.000	52.676	-5.4	108	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012521\
 Data File : VX020739.D
 Acq On : 25 Jan 2021 20:53
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 26 03:13:07 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 25 13:38:15 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	952.495	4.8	95	0.00
50 S	Toluene-d8	50.000	48.740	2.5	108	0.00
51 T	4-Methyl-2-Pentanone	250.000	266.718	-6.7	107	0.00
52 CM	Toluene	50.000	50.901	-1.8#	108	0.00
53 T	t-1,3-Dichloropropene	50.000	49.544	0.9	104	0.00
54 T	cis-1,3-Dichloropropene	50.000	49.947	0.1	104	0.00
55 T	1,1,2-Trichloroethane	50.000	50.269	-0.5	107	0.00
56 T	Ethyl methacrylate	50.000	47.700	4.6	110	0.00
57 T	1,3-Dichloropropane	50.000	50.484	-1.0	106	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	258.127	-3.3	105	0.00
59 T	2-Hexanone	250.000	273.904	-9.6	107	0.00
60 T	Dibromochloromethane	50.000	48.939	2.1	103	0.00
61 T	1,2-Dibromoethane	50.000	49.845	0.3	105	0.00
62 S	4-Bromofluorobenzene	50.000	49.363	1.3	107	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	110	0.00
64 T	Tetrachloroethene	50.000	48.705	2.6	103	0.00
65 PM	Chlorobenzene	50.000	48.629	2.7	106	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	49.435	1.1	104	0.00
67 C	Ethyl Benzene	50.000	51.227	-2.5#	106	0.00
68 T	m/p-Xylenes	100.000	103.289	-3.3	107	0.00
69 T	o-Xylene	50.000	53.168	-6.3	109	0.00
70 T	Styrene	50.000	52.806	-5.6	105	0.00
71 P	Bromoform	50.000	49.641	0.7	103	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	106	0.00
73 T	Isopropylbenzene	50.000	52.303	-4.6	105	0.00
74 T	N-amyl acetate	50.000	48.946	2.1	108	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	50.666	-1.3	106	0.00
76 T	1,2,3-Trichloropropane	50.000	50.868	-1.7	104	0.00
77 T	Bromobenzene	50.000	50.863	-1.7	108	0.00
78 T	n-propylbenzene	50.000	51.945	-3.9	104	0.00
79 T	2-Chlorotoluene	50.000	51.249	-2.5	104	0.00
80 T	1,3,5-Trimethylbenzene	50.000	52.469	-4.9	104	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	49.733	0.5	105	0.00
82 T	4-Chlorotoluene	50.000	50.483	-1.0	103	0.00
83 T	tert-Butylbenzene	50.000	51.649	-3.3	103	0.00
84 T	1,2,4-Trimethylbenzene	50.000	53.366	-6.7	105	0.00
85 T	sec-Butylbenzene	50.000	53.025	-6.0	105	0.00
86 T	p-Isopropyltoluene	50.000	53.207	-6.4	105	0.00
87 T	1,3-Dichlorobenzene	50.000	49.221	1.6	102	0.00
88 T	1,4-Dichlorobenzene	50.000	47.121	5.8	103	0.00
89 T	n-Butylbenzene	50.000	51.108	-2.2	101	0.00
90 T	Hexachloroethane	50.000	50.344	-0.7	102	0.00
91 T	1,2-Dichlorobenzene	50.000	50.404	-0.8	106	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	50.754	-1.5	110	0.00
93 T	1,2,4-Trichlorobenzene	50.000	49.603	0.8	103	0.00
94 T	Hexachlorobutadiene	50.000	47.850	4.3	103	0.00
95 T	Naphthalene	50.000	47.881	4.2	107	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012521\
Data File : VX020739.D
Acq On : 25 Jan 2021 20:53
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jan 26 03:13:07 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X012521W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 25 13:38:15 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	52.534	-5.1	109	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6