

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012522\
 Data File : VX026571.D
 Acq On : 25 Jan 2022 15:59
 Operator : JC/MD
 Sample : N1252-07MS
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Jan 26 06:13:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	98010	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	161985	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	140403	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	65199	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	62341	45.533	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.060%
35) Dibromofluoromethane	5.391	113	48836	46.686	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.380%
50) Toluene-d8	8.653	98	174835	46.381	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.760%
62) 4-Bromofluorobenzene	11.079	95	63553	47.460	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.920%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.173	85	50180	43.358	ug/l	100
3) Chloromethane	1.295	50	46795	44.107	ug/l	100
4) Vinyl Chloride	1.374	62	52864	47.458	ug/l	99
5) Bromomethane	1.612	94	20037	38.994	ug/l	97
6) Chloroethane	1.691	64	30335	53.854	ug/l	99
7) Trichlorofluoromethane	1.892	101	78737	44.771	ug/l	97
8) Diethyl Ether	2.136	74	29054	46.687	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.337	101	51911	48.436	ug/l	99
10) Methyl Iodide	2.459	142	66380	56.865	ug/l	99
11) Tert butyl alcohol	2.983	59	47044	247.001	ug/l #	95
12) 1,1-Dichloroethene	2.325	96	61819	60.394	ug/l	96
13) Acrolein	2.239	56	50143	247.761	ug/l	99
14) Allyl chloride	2.666	41	90530	48.094	ug/l	97
15) Acrylonitrile	3.069	53	155710	265.584	ug/l	99
16) Acetone	2.386	43	131852	188.845	ug/l	98
17) Carbon Disulfide	2.514	76	126338	42.276	ug/l	98
18) Methyl Acetate	2.709	43	70896	51.076	ug/l	99
19) Methyl tert-butyl Ether	3.117	73	178686	48.111	ug/l	99
20) Methylene Chloride	2.794	84	54788	47.494	ug/l	97
21) trans-1,2-Dichloroethene	3.099	96	57486	51.878	ug/l	98
22) Diisopropyl ether	3.764	45	187103	48.933	ug/l	98
23) Vinyl Acetate	3.727	43	785633	249.611	ug/l	99
24) 1,1-Dichloroethane	3.617	63	107876	52.240	ug/l	99
25) 2-Butanone	4.562	43	215536	237.377	ug/l	100
26) 2,2-Dichloropropane	4.483	77	70315	45.938	ug/l #	36
27) cis-1,2-Dichloroethene	4.495	96	611818	483.776	ug/l	90
28) Bromochloromethane	4.904	49	40788	48.104	ug/l	100
29) Tetrahydrofuran	5.013	42	143912	266.864	ug/l	99
30) Chloroform	5.099	83	104634	48.767	ug/l	99
31) Cyclohexane	5.477	56	91261	45.192	ug/l	100
32) 1,1,1-Trichloroethane	5.391	97	91218	48.333	ug/l	99
36) 1,1-Dichloropropene	5.696	75	76815	46.517	ug/l	99
37) Ethyl Acetate	4.721	43	84228	50.422	ug/l #	81
38) Carbon Tetrachloride	5.684	117	81192	47.799	ug/l	95
39) Methylcyclohexane	7.385	83	92414	46.396	ug/l	98
40) Benzene	6.044	78	222401	48.726	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012522\
 Data File : VX026571.D
 Acq On : 25 Jan 2022 15:59
 Operator : JC/MD
 Sample : N1252-07MS
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Jan 26 06:13:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.928	41	48408	51.157	ug/l	97
42) 1,2-Dichloroethane	6.092	62	87289	48.880	ug/l	98
43) Isopropyl Acetate	6.342	43	135639	50.267	ug/l	100
44) Trichloroethene	7.129	130	135300	104.466	ug/l	97
45) 1,2-Dichloropropane	7.434	63	57307	50.300	ug/l	98
46) Dibromomethane	7.586	93	39485	49.163	ug/l	99
47) Bromodichloromethane	7.824	83	80591	49.663	ug/l	100
48) Methyl methacrylate	7.696	41	67847	50.736	ug/l	99
49) 1,4-Dioxane	7.678	88	22036	1004.136	ug/l	98
51) 4-Methyl-2-Pentanone	8.580	43	424847	265.448	ug/l	98
52) Toluene	8.720	92	529566	189.200	ug/l	100
53) t-1,3-Dichloropropene	8.982	75	77394	44.817	ug/l	98
54) cis-1,3-Dichloropropene	8.366	75	87722	50.330	ug/l	97
55) 1,1,2-Trichloroethane	9.153	97	55047	50.397	ug/l	99
56) Ethyl methacrylate	9.116	69	86759	51.830	ug/l	97
57) 1,3-Dichloropropane	9.311	76	95636	50.503	ug/l	100
59) 2-Hexanone	9.433	43	310806	253.725	ug/l	98
60) Dibromochloromethane	9.525	129	58956	51.797	ug/l	100
61) 1,2-Dibromoethane	9.610	107	57785	50.770	ug/l	98
64) Tetrachloroethene	9.275	164	64991	49.001	ug/l	96
65) Chlorobenzene	10.080	112	141691	47.922	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.165	131	53308	48.523	ug/l	99
67) Ethyl Benzene	10.195	91	258553	48.121	ug/l	99
68) m/p-Xylenes	10.305	106	196848	97.899	ug/l	99
69) o-Xylene	10.646	106	95619	49.283	ug/l	98
70) Styrene	10.659	104	158857	50.177	ug/l	98
71) Bromoform	10.799	173	39980	47.655	ug/l #	99
73) Isopropylbenzene	10.964	105	253622	47.457	ug/l	99
74) N-amyl acetate	10.848	43	100696	48.806	ug/l	98
75) 1,1,2,2-Tetrachloroethane	11.213	83	71144	49.596	ug/l	100
76) 1,2,3-Trichloropropane	11.244	75	86811	60.473	ug/l	91
77) Bromobenzene	11.201	156	58648	47.902	ug/l	99
78) n-propylbenzene	11.305	91	289269	47.465	ug/l	100
79) 2-Chlorotoluene	11.366	91	173393	46.918	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	209242	47.491	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	21354	47.229	ug/l	95
82) 4-Chlorotoluene	11.457	91	198775	46.799	ug/l	98
83) tert-Butylbenzene	11.719	119	199128	48.414	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	210658	47.802	ug/l	99
85) sec-Butylbenzene	11.890	105	256335	48.182	ug/l	100
86) p-Isopropyltoluene	12.012	119	217773	48.790	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	109111	47.545	ug/l	99
88) 1,4-Dichlorobenzene	12.043	146	110368	47.514	ug/l	99
89) n-Butylbenzene	12.335	91	184994	48.297	ug/l	100
90) Hexachloroethane	12.542	117	33735	45.495	ug/l	98
91) 1,2-Dichlorobenzene	12.335	146	106808	49.250	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.945	75	18004	51.930	ug/l	97
93) 1,2,4-Trichlorobenzene	13.591	180	67706	48.695	ug/l	100
94) Hexachlorobutadiene	13.725	225	27101	46.173	ug/l	99
95) Naphthalene	13.780	128	239823	51.711	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	68206	49.880	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012522\
Data File : VX026571.D
Acq On : 25 Jan 2022 15:59
Operator : JC/MD
Sample : N1252-07MS
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :


Quant Time: Jan 26 06:13:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed						

Quant Time: Jan 26 06:13:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 71
Delta R.T. 0.006 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:168 Resp: 98010
Ion Ratio Lower Upper
168 100
99 61.6 50.3 75.5



#2
Dichlorodifluoromethane
Concen: 43.358 ug/l
RT: 1.173 min Scan# 14
Delta R.T. 0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59

Tgt Ion: 85 Resp: 50180
Ion Ratio Lower Upper
85 100
87 32.4 16.4 49.1









#3
Chloromethane
Concen: 44.107 ug/l
RT: 1.295 min Scan# 34
Delta R.T. 0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 50 Resp: 46795
Ion Ratio Lower Upper
50 100
52 32.7 26.1 39.1



#4
Vinyl Chloride
Concen: 47.458 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.006 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59

Tgt Ion: 62 Resp: 52864
Ion Ratio Lower Upper
62 100
64 32.2 26.2 39.2









#5
Bromomethane
Concen: 38.994 ug/l
RT: 1.612 min Scan# 80
Delta R.T. -0.006 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 94 Resp: 20037
Ion Ratio Lower Upper
94 100
96 97.6 75.9 113.9



#6
Chloroethane
Concen: 53.854 ug/l
RT: 1.691 min Scan# 99
Delta R.T. 0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59


Tgt Ion: 64 Resp: 30335
Ion Ratio Lower Upper
64 100
66 32.5 25.4 38.2








#7
Trichlorofluoromethane
Concen: 44.771 ug/l
RT: 1.892 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:101 Resp: 78737
Ion Ratio Lower Upper
101 100
103 66.1 51.3 76.9




#8
Diethyl Ether
Concen: 46.687 ug/l
RT: 2.136 min Scan# 172
Delta R.T. 0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59



Tgt Ion: 74 Resp: 29054
Ion Ratio Lower Upper
74 100
45 101.4 50.7 152.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 48.436 ug/l
RT: 2.337 min Scan# 205
Delta R.T. 0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:101 Resp: 51911
Ion Ratio Lower Upper
101 100
85 43.5 35.3 52.9
151 76.4 61.4 92.2



#10
Methyl Iodide
Concen: 56.865 ug/l
RT: 2.459 min Scan# 225
Delta R.T. -0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59

Tgt Ion:142 Resp: 66380
Ion Ratio Lower Upper
142 100
127 47.0 38.1 57.1
141 14.4 11.2 16.8









#11
Tert butyl alcohol
Concen: 247.001 ug/l
RT: 2.983 min Scan# 311
Delta R.T. -0.024 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 59 Resp: 47044
Ion Ratio Lower Upper
59 100
57 11.0 7.2 10.8#



#12
1,1-Dichloroethene
Concen: 60.394 ug/l
RT: 2.325 min Scan# 203
Delta R.T. 0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59

Tgt Ion: 96 Resp: 61819
Ion Ratio Lower Upper
96 100
61 166.2 138.5 207.7
98 64.9 51.8 77.6





#13

Acrolein

Concen: 247.761 ug/l

RT: 2.239 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX026571.D

Acq: 25 Jan 2022 15:59

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 56 Resp: 50143

Ion Ratio Lower Upper

56 100

55 70.3 56.9 85.3



#14

Allyl chloride

Concen: 48.094 ug/l

RT: 2.666 min Scan# 259

Delta R.T. -0.006 min

Lab File: VX026571.D

Acq: 25 Jan 2022 15:59




Tgt Ion: 41 Resp: 90530


Ion Ratio Lower Upper

41 100

39 64.2 54.2 81.4


76 32.1 25.2 37.8








#15
Acrylonitrile
Concen: 265.584 ug/l
RT: 3.069 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 53 Resp: 155710
Ion Ratio Lower Upper
53 100
52 83.1 66.6 100.0
51 35.8 29.5 44.3



#16
Acetone
Concen: 188.845 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.006 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59

Tgt Ion: 43 Resp: 131852
Ion Ratio Lower Upper
43 100
58 30.0 23.0 34.6





#17

Carbon Disulfide

Concen: 42.276 ug/l

RT: 2.514 min Scan# 21

Delta R.T. -0.006 min

Lab File: VX026571.D

Acq: 25 Jan 2022 15:59

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 76 Resp: 126338

Ion Ratio Lower Upper

76 100

78 8.8 7.5 11.3



#18

Methyl Acetate

Concen: 51.076 ug/l

RT: 2.709 min Scan# 266

Delta R.T. 0.000 min

Lab File: VX026571.D




Acq: 25 Jan 2022 15:59


Tgt Ion: 43 Resp: 70896

Ion Ratio Lower Upper

43 100

74 23.2 18.3 27.5





#19

Methyl tert-butyl Ether

Concen: 48.111 ug/l

RT: 3.117 min Scan# 31

Delta R.T. 0.000 min

Lab File: VX026571.D

Acq: 25 Jan 2022 15:59

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 73 Resp: 178686

Ion Ratio Lower Upper

73 100

57 21.4 17.4 26.2



#20

Methylene Chloride

Concen: 47.494 ug/l

RT: 2.794 min Scan# 280

Delta R.T. 0.000 min

Lab File: VX026571.D

Acq: 25 Jan 2022 15:59

Tgt Ion: 84 Resp: 54788




Ion Ratio Lower Upper


84 100

49 122.8 101.5 152.3

51 37.0 31.2 46.8

86 63.0 51.3 76.9









#21
trans-1,2-Dichloroethene
Concen: 51.878 ug/l
RT: 3.099 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 96 Resp: 57486
Ion Ratio Lower Upper
96 100
61 139.0 112.9 169.3
98 64.9 49.5 74.3



#22
Diisopropyl ether
Concen: 48.933 ug/l
RT: 3.764 min Scan# 439
Delta R.T. 0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59

Tgt Ion: 45 Resp: 187103
Ion Ratio Lower Upper
45 100
43 78.9 60.6 91.0
87 27.5 22.0 33.0
59 11.2 9.0 13.4









#23
 Vinyl Acetate
 Concen: 249.611 ug/l
 RT: 3.727 min Scan# 411
 Delta R.T. 0.000 min
 Lab File: VX026571.D
 Acq: 25 Jan 2022 15:59




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 43 Resp: 785633
 Ion Ratio Lower Upper
 43 100
 86 10.4 8.1 12.1



#24
 1,1-Dichloroethane
 Concen: 52.240 ug/l
 RT: 3.617 min Scan# 415
 Delta R.T. 0.000 min
 Lab File: VX026571.D
 Acq: 25 Jan 2022 15:59

Tgt Ion: 63 Resp: 107876
 Ion Ratio Lower Upper
 63 100
 98 7.5 3.7 11.1
 100 4.5 2.4 7.2









#25
2-Butanone
Concen: 237.377 ug/l
RT: 4.562 min Scan# 51
Delta R.T. 0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 215536
Ion Ratio Lower Upper
43 100
72 25.7 20.4 30.6



#26
2,2-Dichloropropane
Concen: 45.938 ug/l
RT: 4.483 min Scan# 557
Delta R.T. 0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59

Tgt Ion: 77 Resp: 70315
Ion Ratio Lower Upper
77 100
97 52.6 11.0 33.0#





#27

cis-1,2-Dichloroethene

Concen: 483.776 ug/l

RT: 4.495 min Scan# 51

Delta R.T. 0.000 min


Lab File: VX026571.D

Acq: 25 Jan 2022 15:59

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion: 96 Resp: 611818

Ion Ratio Lower Upper

96 100

61 130.3 0.0 296.6

98 64.2 0.0 128.8



#28

Bromochloromethane

Concen: 48.104 ug/l

RT: 4.904 min Scan# 626

Delta R.T. 0.000 min

Lab File: VX026571.D

Acq: 25 Jan 2022 15:59




Tgt Ion: 49 Resp: 40788


Ion Ratio Lower Upper

49 100

129 2.0 0.0 4.2

130 78.4 62.4 93.6





#29

Tetrahydrofuran

Concen: 266.864 ug/l

RT: 5.013 min Scan# 644

Delta R.T. 0.000 min


Lab File: VX026571.D

Acq: 25 Jan 2022 15:59

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion: 42 Resp: 143912

Ion Ratio Lower Upper

42 100

72 43.5 34.6 52.0

71 40.3 31.8 47.8



#30

Chloroform

Concen: 48.767 ug/l

RT: 5.099 min Scan# 658

Delta R.T. 0.000 min

Lab File: VX026571.D




Acq: 25 Jan 2022 15:59


Tgt Ion: 83 Resp: 104634

Ion Ratio Lower Upper

83 100


85 65.5 51.6 77.4








#31
Cyclohexane
Concen: 45.192 ug/l
RT: 5.477 min Scan# 720
Delta R.T. 0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion: 56 Resp: 91261
Ion Ratio Lower Upper
56 100
69 32.1 25.8 38.6
84 86.4 68.9 103.3




#32
1,1,1-Trichloroethane
Concen: 48.333 ug/l
RT: 5.391 min Scan# 706
Delta R.T. 0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59



Tgt Ion: 97 Resp: 91218
Ion Ratio Lower Upper
97 100
99 64.6 51.3 76.9
61 45.1 35.0 52.6





#33

1,2-Dichloroethane-d4

Concen: 45.533 ug/l

RT: 5.958 min Scan# 799

Delta R.T. 0.000 min

Lab File: VX026571.D

Acq: 25 Jan 2022 15:59

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 65 Resp: 62341

Ion Ratio Lower Upper

65 100

67 52.5 0.0 100.8



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.769 min Scan# 932

Delta R.T. 0.006 min

Lab File: VX026571.D

Acq: 25 Jan 2022 15:59




Tgt Ion: 114 Resp: 161985


Ion Ratio Lower Upper

114 100

63 20.5 0.0 42.8

88 16.3 0.0 33.0





#35

Dibromofluoromethane

Concen: 46.686 ug/l

RT: 5.391 min Scan# 706

Delta R.T. 0.000 min

Lab File: VX026571.D

Acq: 25 Jan 2022 15:59

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:113 Resp: 48836

Ion Ratio Lower Upper

113 100

111 103.3 81.2 121.8

192 19.4 15.0 22.4



#36

1,1-Dichloropropene

Concen: 46.517 ug/l

RT: 5.696 min Scan# 756

Delta R.T. 0.000 min

Lab File: VX026571.D

Acq: 25 Jan 2022 15:59




Tgt Ion: 75 Resp: 76815


Ion Ratio Lower Upper

75 100

110 35.3 17.2 51.6

77 30.3 24.6 36.8









#37
Ethyl Acetate
Concen: 50.422 ug/l
RT: 4.721 min Scan# 596
Delta R.T. 0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 84228
Ion Ratio Lower Upper
43 100
61 0.0 10.6 16.0#
70 10.7 8.5 12.7



#38
Carbon Tetrachloride
Concen: 47.799 ug/l
RT: 5.684 min Scan# 754
Delta R.T. 0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59

Tgt Ion: 117 Resp: 81192
Ion Ratio Lower Upper
117 100
119 98.9 74.7 112.1
121 30.3 25.4 38.0





#39

Methylcyclohexane

Concen: 46.396 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. 0.000 min

Lab File: VX026571.D

Acq: 25 Jan 2022 15:59

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 83 Resp: 92414

Ion Ratio Lower Upper

83 100

55 80.0 63.6 95.4

98 50.6 38.4 57.6



#40

Benzene

Concen: 48.726 ug/l

RT: 6.044 min Scan# 813

Delta R.T. 0.000 min

Lab File: VX026571.D




Acq: 25 Jan 2022 15:59


Tgt Ion: 78 Resp: 222401

Ion Ratio Lower Upper

78 100


77 23.0 18.9 28.3








#41
Methacrylonitrile
Concen: 51.157 ug/l
RT: 4.928 min Scan# 61
Delta R.T. 0.006 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion: 41 Resp: 48408
Ion Ratio Lower Upper
41 100
39 54.9 45.4 68.0
67 69.3 53.2 79.8
52 28.6 23.0 34.6




#42
1,2-Dichloroethane
Concen: 48.880 ug/l
RT: 6.092 min Scan# 821
Delta R.T. 0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59



Tgt Ion: 62 Resp: 87289
Ion Ratio Lower Upper
62 100
98 9.5 0.0 17.8









#43
Isopropyl Acetate
Concen: 50.267 ug/l
RT: 6.342 min Scan# 862
Delta R.T. 0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 135639
Ion Ratio Lower Upper
43 100
61 19.4 15.4 23.0
87 12.8 10.2 15.2



#44
Trichloroethene
Concen: 104.466 ug/l
RT: 7.129 min Scan# 991
Delta R.T. 0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59

Tgt Ion:130 Resp: 135300
Ion Ratio Lower Upper
130 100
95 98.3 0.0 202.6





#45

1,2-Dichloropropane

Concen: 50.300 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. 0.000 min

Lab File: VX026571.D

Acq: 25 Jan 2022 15:59

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 63 Resp: 57307

Ion Ratio Lower Upper

63 100

65 30.7 23.8 35.6



#46

Dibromomethane

Concen: 49.163 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. 0.000 min

Lab File: VX026571.D

Acq: 25 Jan 2022 15:59




Tgt Ion: 93 Resp: 39485


Ion Ratio Lower Upper

93 100

95 83.2 66.1 99.1

174 99.8 78.7 118.1









#47
Bromodichloromethane
Concen: 49.663 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. 0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 80591
Ion Ratio Lower Upper
83 100
85 63.9 51.4 77.0
127 8.6 6.9 10.3



#48
Methyl methacrylate
Concen: 50.736 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. 0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59

Tgt Ion: 41 Resp: 67847
Ion Ratio Lower Upper
41 100
69 78.7 62.2 93.4
39 55.4 45.4 68.0





#49

1,4-Dioxane

Concen: 1004.136 ug/l

RT: 7.678 min Scan# 1084

Delta R.T. -0.018 min


Lab File: VX026571.D

Acq: 25 Jan 2022 15:59

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion: 88 Resp: 22036

Ion Ratio Lower Upper

88 100

43 36.0 29.9 44.9

58 70.7 57.8 86.6



#50

Toluene-d8

Concen: 46.381 ug/l

RT: 8.653 min Scan# 1241

Delta R.T. 0.000 min

Lab File: VX026571.D




Acq: 25 Jan 2022 15:59


Tgt Ion: 98 Resp: 174835

Ion Ratio Lower Upper

98 100

100 66.8 53.2 79.8









#51
4-Methyl-2-Pentanone
Concen: 265.448 ug/l
RT: 8.580 min Scan# 1229
Delta R.T. 0.006 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 424847
Ion Ratio Lower Upper
43 100
58 39.0 30.5 45.7



#52
Toluene
Concen: 189.200 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. 0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59

Tgt Ion: 92 Resp: 529566
Ion Ratio Lower Upper
92 100
91 169.3 135.7 203.5





#53

t-1,3-Dichloropropene

Concen: 44.817 ug/l

RT: 8.982 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX026571.D

Acq: 25 Jan 2022 15:59

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 75 Resp: 77394

Ion Ratio Lower Upper

75 100

77 31.1 25.7 38.5



#54

cis-1,3-Dichloropropene

Concen: 50.330 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX026571.D

Acq: 25 Jan 2022 15:59




Tgt Ion: 75 Resp: 87722


Ion Ratio Lower Upper

75 100

77 31.2 26.2 39.2

39 52.2 43.5 65.3





#55

1,1,2-Trichloroethane

Concen: 50.397 ug/l

RT: 9.153 min Scan# 1323

Delta R.T. 0.000 min

Lab File: VX026571.D

Acq: 25 Jan 2022 15:59

Instrument :

MSVOA_X

ClientSampleId :

Tgt Ion: 97 Resp: 55047





Ion Ratio Lower Upper

97 100

83 84.3 68.4 102.6

85 54.5 44.0 66.0

99 62.4 49.8 74.6



#56

Ethyl methacrylate

Concen: 51.830 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. 0.000 min

Lab File: VX026571.D

Acq: 25 Jan 2022 15:59




Tgt Ion: 69 Resp: 86759


Ion Ratio Lower Upper

69 100

41 69.2 57.2 85.8

39 37.3 30.9 46.3





#57

1,3-Dichloropropane

Concen: 50.503 ug/l

RT: 9.311 min Scan# 1349

Delta R.T. 0.000 min

Lab File: VX026571.D

Acq: 25 Jan 2022 15:59

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 76 Resp: 95636

Ion Ratio Lower Upper

76 100

78 32.2 25.7 38.5



#59

2-Hexanone

Concen: 253.725 ug/l

RT: 9.433 min Scan# 1369

Delta R.T. 0.000 min

Lab File: VX026571.D




Acq: 25 Jan 2022 15:59


Tgt Ion: 43 Resp: 310806

Ion Ratio Lower Upper

43 100

58 53.9 26.4 79.2





#60

Dibromochloromethane

Concen: 51.797 ug/l

RT: 9.525 min Scan# 1384

Delta R.T. 0.000 min

Lab File: VX026571.D

Acq: 25 Jan 2022 15:59

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion:129 Resp: 58956

Ion Ratio Lower Upper

129 100

127 77.0 38.5 115.5



#61

1,2-Dibromoethane

Concen: 50.770 ug/l

RT: 9.610 min Scan# 1398

Delta R.T. 0.000 min

Lab File: VX026571.D




Acq: 25 Jan 2022 15:59


Tgt Ion:107 Resp: 57785

Ion Ratio Lower Upper

107 100

109 92.8 75.5 113.3









#62
4-Bromofluorobenzene
Concen: 47.460 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 95 Resp: 63553
Ion Ratio Lower Upper
95 100
174 75.2 0.0 148.6
176 72.8 0.0 141.4



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59


Tgt Ion: 117 Resp: 140403
Ion Ratio Lower Upper
117 100
82 56.6 45.2 67.8
119 32.1 26.2 39.2








#64
Tetrachloroethene
Concen: 49.001 ug/l
RT: 9.275 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:164 Resp: 64991
Ion Ratio Lower Upper
164 100
166 126.2 101.8 152.6
129 97.3 83.3 124.9
131 93.1 78.0 117.0



#65
Chlorobenzene
Concen: 47.922 ug/l
RT: 10.080 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59

Tgt Ion:112 Resp: 141691
Ion Ratio Lower Upper
112 100
114 31.9 25.3 37.9









#66
 1,1,1,2-Tetrachloroethane
 Concen: 48.523 ug/l
 RT: 10.165 min Scan# 1489
 Delta R.T. 0.000 min
 Lab File: VX026571.D
 Acq: 25 Jan 2022 15:59




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion:131 Resp: 53308
 Ion Ratio Lower Upper
 131 100
 133 96.5 47.9 143.8
 119 63.9 31.7 95.1



#67
 Ethyl Benzene
 Concen: 48.121 ug/l
 RT: 10.195 min Scan# 1494
 Delta R.T. 0.000 min
 Lab File: VX026571.D
 Acq: 25 Jan 2022 15:59

Tgt Ion: 91 Resp: 258553
 Ion Ratio Lower Upper
 91 100
 106 31.3 24.7 37.1









#68
m/p-Xylenes
Concen: 97.899 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. 0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:106 Resp: 196848
Ion Ratio Lower Upper
106 100
91 204.9 165.3 247.9



#69
o-Xylene
Concen: 49.283 ug/l
RT: 10.646 min Scan# 1568
Delta R.T. 0.006 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59

Tgt Ion:106 Resp: 95619
Ion Ratio Lower Upper
106 100
91 216.1 109.7 329.0









#70
Styrene
Concen: 50.177 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. 0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:104 Resp: 158857
Ion Ratio Lower Upper
104 100
78 51.7 42.5 63.7
103 55.1 43.3 64.9



#71
Bromoform
Concen: 47.655 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59


Tgt Ion:173 Resp: 39980
Ion Ratio Lower Upper
173 100
175 47.9 24.3 72.9
254 0.0 0.0 0.0








#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:152 Resp: 65199
Ion Ratio Lower Upper
152 100
115 86.6 44.2 132.6
150 174.2 0.0 348.2




#73
Isopropylbenzene
Concen: 47.457 ug/l
RT: 10.964 min Scan# 1620
Delta R.T. 0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59




Tgt Ion:105 Resp: 253622
Ion Ratio Lower Upper
105 100
120 26.4 13.1 39.1








#74
N-amy1 acetate
Concen: 48.806 ug/l
RT: 10.848 min Scan# 1601
Delta R.T. 0.006 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 43	Resp: 100696
Ion Ratio	Lower Upper
43 100	
70 42.5	32.9 49.3
55 24.0	18.3 27.5
61 23.4	18.9 28.3



#75
1,1,2,2-Tetrachloroethane
Concen: 49.596 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. 0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59

Tgt Ion: 83	Resp: 71144
Ion Ratio	Lower Upper
83 100	
131 10.5	5.3 15.9
85 64.4	32.0 96.2









#76
1,2,3-Trichloropropane
Concen: 60.473 ug/l
RT: 11.244 min Scan# 1666
Delta R.T. 0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 75 Resp: 86811
Ion Ratio Lower Upper
75 100
77 31.1 18.3 54.8



#77
Bromobenzene
Concen: 47.902 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59

Tgt Ion: 156 Resp: 58648
Ion Ratio Lower Upper
156 100
77 168.5 84.5 253.4
158 97.8 48.1 144.4









#78
n-propylbenzene
Concen: 47.465 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 289269
Ion Ratio Lower Upper
91 100
120 23.1 11.6 34.7



#79
2-Chlorotoluene
Concen: 46.918 ug/l
RT: 11.366 min Scan# 1686
Delta R.T. 0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59

Tgt Ion: 91 Resp: 173393
Ion Ratio Lower Upper
91 100
126 33.2 16.4 49.2





#80

1,3,5-Trimethylbenzene

Concen: 47.491 ug/l

RT: 11.451 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX026571.D

Acq: 25 Jan 2022 15:59

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion:105 Resp: 209242

Ion Ratio Lower Upper

105 100

120 48.2 24.0 72.0



#81

trans-1,4-Dichloro-2-butene

Concen: 47.229 ug/l

RT: 11.018 min Scan# 1629

Delta R.T. 0.000 min

Lab File: VX026571.D

Acq: 25 Jan 2022 15:59




Tgt Ion: 75 Resp: 21354


Ion Ratio Lower Upper

75 100

53 110.4 94.0 141.0

89 46.0 37.0 55.4









#82
4-Chlorotoluene
Concen: 46.799 ug/l
RT: 11.457 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 198775
Ion Ratio Lower Upper
91 100
126 29.2 14.1 42.4



#83
tert-Butylbenzene
Concen: 48.414 ug/l
RT: 11.719 min Scan# 1744
Delta R.T. 0.006 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59


Tgt Ion: 119 Resp: 199128
Ion Ratio Lower Upper
119 100
91 61.3 31.4 94.0
134 23.8 11.8 35.3








#84
1,2,4-Trimethylbenzene
Concen: 47.802 ug/l
RT: 11.756 min Scan# 11750
Delta R.T. 0.006 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:105 Resp: 210658
Ion Ratio Lower Upper
105 100
120 44.4 21.9 65.5



#85
sec-Butylbenzene
Concen: 48.182 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. 0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59

Tgt Ion:105 Resp: 256335
Ion Ratio Lower Upper
105 100
134 20.2 10.1 30.1









#86
 p-Isopropyltoluene
 Concen: 48.790 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. 0.000 min
 Lab File: VX026571.D
 Acq: 25 Jan 2022 15:59




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion:119 Resp: 217773
 Ion Ratio Lower Upper
 119 100
 134 26.1 13.0 39.0
 91 24.5 12.6 37.6



#87
 1,3-Dichlorobenzene
 Concen: 47.545 ug/l
 RT: 11.969 min Scan# 1785
 Delta R.T. 0.000 min
 Lab File: VX026571.D
 Acq: 25 Jan 2022 15:59

Tgt Ion:146 Resp: 109111
 Ion Ratio Lower Upper
 146 100
 111 41.9 21.3 63.7
 148 63.1 31.6 95.0





#88

1,4-Dichlorobenzene

Concen: 47.514 ug/l

RT: 12.043 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX026571.D

Acq: 25 Jan 2022 15:59

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:146 Resp: 110368

Ion Ratio Lower Upper

146 100

111 40.6 20.2 60.6

148 63.4 31.4 94.0



#89

n-Butylbenzene

Concen: 48.297 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.000 min

Lab File: VX026571.D

Acq: 25 Jan 2022 15:59




Tgt Ion: 91 Resp: 184994


Ion Ratio Lower Upper

91 100

92 54.7 27.2 81.6


134 26.2 13.1 39.1








#90
Hexachloroethane
Concen: 45.495 ug/l
RT: 12.542 min Scan# 1879
Delta R.T. 0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:117 Resp: 33735
Ion Ratio Lower Upper
117 100
201 73.0 35.6 106.8




#91
1,2-Dichlorobenzene
Concen: 49.250 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59



Tgt Ion:146 Resp: 106808
Ion Ratio Lower Upper
146 100
111 42.6 21.8 65.3
148 64.3 32.0 96.2









#92
 1,2-Dibromo-3-Chloropropane
 Concen: 51.930 ug/l
 RT: 12.945 min Scan# 1945
 Delta R.T. 0.000 min
 Lab File: VX026571.D
 Acq: 25 Jan 2022 15:59




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 75 Resp: 18004
 Ion Ratio Lower Upper
 75 100
 155 81.7 40.0 119.9
 157 107.9 51.6 154.8



#93
 1,2,4-Trichlorobenzene
 Concen: 48.695 ug/l
 RT: 13.591 min Scan# 2051
 Delta R.T. 0.000 min
 Lab File: VX026571.D
 Acq: 25 Jan 2022 15:59

Tgt Ion: 180 Resp: 67706
 Ion Ratio Lower Upper
 180 100
 182 93.1 46.7 140.0
 145 32.3 16.3 48.8









#94
Hexachlorobutadiene
Concen: 46.173 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:225 Resp: 27101
Ion Ratio Lower Upper
225 100
223 62.7 32.0 96.0
227 64.2 32.3 96.9



#95
Naphthalene
Concen: 51.711 ug/l
RT: 13.780 min Scan# 2082
Delta R.T. 0.006 min
Lab File: VX026571.D
Acq: 25 Jan 2022 15:59

Tgt Ion:128 Resp: 239823
Ion Ratio Lower Upper
128 100
127 12.9 10.3 15.5
129 10.9 8.7 13.1





#96

1,2,3-Trichlorobenzene

Concen: 49.880 ug/l

RT: 13.963 min Scan# 21

Delta R.T. 0.000 min


Lab File: VX026571.D

Acq: 25 Jan 2022 15:59

Instrument :

MSVOA_X

ClientSampleId :



Tgt Ion:180 Resp: 68206

Ion Ratio Lower Upper

180 100

182 94.2 47.2 141.6

145 33.7 17.5 52.5

