

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012522\
 Data File : VX026567.D
 Acq On : 25 Jan 2022 14:26
 Operator : JC/MD
 Sample : N1252-02DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GWBR3D-B-465-493-012422-FDDL

Quant Time: Jan 26 06:12:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	123535	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	210063	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	184325	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	77234	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	83452	48.359	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.720%
35) Dibromofluoromethane	5.391	113	66483	49.010	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.020%
50) Toluene-d8	8.653	98	242395	49.587	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.180%
62) 4-Bromofluorobenzene	11.079	95	86848	50.013	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.020%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.374	62	481	0.343	ug/l	90
12) 1,1-Dichloroethene	2.325	96	2545	1.973	ug/l	92
21) trans-1,2-Dichloroethene	3.099	96	759	0.543	ug/l	96
24) 1,1-Dichloroethane	3.623	63	1363	0.524	ug/l #	81
27) cis-1,2-Dichloroethene	4.495	96	109782	68.871	ug/l	89
44) Trichloroethene	7.129	130	13643	8.123	ug/l	99
52) Toluene	8.720	92	79154	21.807	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012522\
Data File : VX026567.D
Acq On : 25 Jan 2022 14:26
Operator : JC/MD
Sample : N1252-02DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GWBR3D-B-465-493-012422-FDDL

Quant Time: Jan 26 06:12:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

