

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012522\
 Data File : VX026558.D
 Acq On : 25 Jan 2022 10:09
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 26 07:09:26 2022
 Quant Title :
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	92	0.00
2 T	Dichlorodifluoromethane	50.000	42.738	14.5	78	0.00
3 P	Chloromethane	50.000	43.536	12.9	82	0.00
4 C	Vinyl Chloride	50.000	45.361	9.3#	81	0.00
5 T	Bromomethane	50.000	42.608	14.8	79	0.00
6 T	Chloroethane	50.000	50.795	-1.6	93	0.00
7 T	Trichlorofluoromethane	50.000	45.251	9.5	80	0.00
8 T	Diethyl Ether	50.000	45.303	9.4	80	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	49.003	2.0	87	0.00
10 T	Methyl Iodide	50.000	55.153	-10.3	105	0.00
11 T	Tert butyl alcohol	250.000	209.955	16.0	73	-0.02
12 CM	1,1-Dichloroethene	50.000	47.646	4.7#	85	0.00
13 T	Acrolein	250.000	240.214	3.9	93	0.00
14 T	Allyl chloride	50.000	48.082	3.8	84	0.00
15 T	Acrylonitrile	250.000	242.078	3.2	87	0.00
16 T	Acetone	250.000	237.198	5.1	84	0.00
17 T	Carbon Disulfide	50.000	42.902	14.2	79	0.00
18 T	Methyl Acetate	50.000	47.882	4.2	88	0.00
19 T	Methyl tert-butyl Ether	50.000	46.425	7.2	83	0.00
20 T	Methylene Chloride	50.000	46.404	7.2	87	0.00
21 T	trans-1,2-Dichloroethene	50.000	47.568	4.9	85	0.00
22 T	Diisopropyl ether	50.000	47.711	4.6	85	0.00
23 T	Vinyl Acetate	250.000	241.409	3.4	83	0.00
24 P	1,1-Dichloroethane	50.000	48.157	3.7	85	0.00
25 T	2-Butanone	250.000	246.576	1.4	86	0.00
26 T	2,2-Dichloropropane	50.000	48.804	2.4	83	0.00
27 T	cis-1,2-Dichloroethene	50.000	48.104	3.8	84	0.00
28 T	Bromochloromethane	50.000	50.664	-1.3	89	0.00
29 T	Tetrahydrofuran	250.000	236.695	5.3	85	0.00
30 C	Chloroform	50.000	47.858	4.3#	85	0.00
31 T	Cyclohexane	50.000	45.821	8.4	83	0.00
32 T	1,1,1-Trichloroethane	50.000	48.592	2.8	84	0.00
33 S	1,2-Dichloroethane-d4	50.000	47.139	5.7	84	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	90	0.00
35 S	Dibromofluoromethane	50.000	49.921	0.2	86	0.00
36 T	1,1-Dichloropropene	50.000	47.842	4.3	84	0.00
37 T	Ethyl Acetate	50.000	48.597	2.8	84	0.00
38 T	Carbon Tetrachloride	50.000	49.495	1.0	84	0.00
39 T	Methylcyclohexane	50.000	48.970	2.1	85	0.00
40 TM	Benzene	50.000	48.575	2.8	85	0.00
41 T	Methacrylonitrile	50.000	50.555	-1.1	87	0.00
42 TM	1,2-Dichloroethane	50.000	47.459	5.1	83	0.00
43 T	Isopropyl Acetate	50.000	49.130	1.7	83	0.00
44 TM	Trichloroethene	50.000	50.644	-1.3	87	0.00
45 C	1,2-Dichloropropane	50.000	50.137	-0.3#	86	0.00
46 T	Dibromomethane	50.000	49.630	0.7	86	0.00
47 T	Bromodichloromethane	50.000	50.546	-1.1	85	0.00
48 T	Methyl methacrylate	50.000	48.693	2.6	84	0.00
49 T	1,4-Dioxane	1000.000	892.392	10.8	76	-0.01

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012522\
 Data File : VX026558.D
 Acq On : 25 Jan 2022 10:09
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 26 07:09:26 2022
 Quant Title :
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

50 S	Toluene-d8	50.000	49.918	0.2	87	0.00
51 T	4-Methyl-2-Pentanone	250.000	251.380	-0.6	87	0.00
52 CM	Toluene	50.000	49.376	1.2#	86	0.00
53 T	t-1,3-Dichloropropene	50.000	46.797	6.4	86	0.00
54 T	cis-1,3-Dichloropropene	50.000	52.283	-4.6	86	0.00
55 T	1,1,2-Trichloroethane	50.000	49.959	0.1	87	0.00
56 T	Ethyl methacrylate	50.000	50.478	-1.0	84	0.00
57 T	1,3-Dichloropropane	50.000	50.230	-0.5	87	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	273.931	-9.6	94	0.00
59 T	2-Hexanone	250.000	262.346	-4.9	88	0.00
60 T	Dibromochloromethane	50.000	53.230	-6.5	88	0.00
61 T	1,2-Dibromoethane	50.000	50.223	-0.4	86	0.00
62 S	4-Bromofluorobenzene	50.000	51.479	-3.0	89	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	92	0.00
64 T	Tetrachloroethene	50.000	51.386	-2.8	92	0.00
65 PM	Chlorobenzene	50.000	49.226	1.5	88	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	50.964	-1.9	89	0.00
67 C	Ethyl Benzene	50.000	49.277	1.4#	86	0.00
68 T	m/p-Xylenes	100.000	100.106	-0.1	87	0.00
69 T	o-Xylene	50.000	50.658	-1.3	88	0.00
70 T	Styrene	50.000	51.919	-3.8	88	0.00
71 P	Bromoform	50.000	48.336	3.3	91	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	95	0.00
73 T	Isopropylbenzene	50.000	47.654	4.7	88	0.00
74 T	N-amyl acetate	50.000	47.599	4.8	83	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	47.092	5.8	87	0.00
76 T	1,2,3-Trichloropropane	50.000	43.958	12.1	80	0.00
77 T	Bromobenzene	50.000	46.818	6.4	86	0.00
78 T	n-propylbenzene	50.000	48.694	2.6	89	0.00
79 T	2-Chlorotoluene	50.000	46.919	6.2	88	0.00
80 T	1,3,5-Trimethylbenzene	50.000	48.124	3.8	88	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	48.549	2.9	96	0.00
82 T	4-Chlorotoluene	50.000	47.524	5.0	88	0.00
83 T	tert-Butylbenzene	50.000	47.611	4.8	87	0.00
84 T	1,2,4-Trimethylbenzene	50.000	47.919	4.2	88	0.00
85 T	sec-Butylbenzene	50.000	49.740	0.5	90	0.00
86 T	p-Isopropyltoluene	50.000	50.292	-0.6	91	0.00
87 T	1,3-Dichlorobenzene	50.000	48.635	2.7	92	0.00
88 T	1,4-Dichlorobenzene	50.000	47.495	5.0	89	0.00
89 T	n-Butylbenzene	50.000	51.241	-2.5	91	0.00
90 T	Hexachloroethane	50.000	48.215	3.6	93	0.00
91 T	1,2-Dichlorobenzene	50.000	49.089	1.8	90	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	49.104	1.8	86	0.00
93 T	1,2,4-Trichlorobenzene	50.000	50.598	-1.2	91	0.00
94 T	Hexachlorobutadiene	50.000	50.345	-0.7	93	0.00
95 T	Naphthalene	50.000	51.514	-3.0	92	0.00
96 T	1,2,3-Trichlorobenzene	50.000	51.218	-2.4	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012522\
Data File : VX026558.D
Acq On : 25 Jan 2022 10:09
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jan 26 07:09:26 2022
Quant Title :
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 6