

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX012522\
 Data File : VX026573.D
 Acq On : 25 Jan 2022 16:45
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 26 06:14:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	93	0.00
2 T	Dichlorodifluoromethane	50.000	38.995	22.0	71	0.00
3 P	Chloromethane	50.000	40.553	18.9	77	0.00
4 C	Vinyl Chloride	50.000	42.170	15.7#	76	0.00
5 T	Bromomethane	50.000	38.576	22.8	72	0.00
6 T	Chloroethane	50.000	48.689	2.6	90	0.00
7 T	Trichlorofluoromethane	50.000	41.905	16.2	75	0.00
8 T	Diethyl Ether	50.000	43.946	12.1	78	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	44.842	10.3	80	0.00
10 T	Methyl Iodide	50.000	54.599	-9.2	105	0.00
11 T	Tert butyl alcohol	250.000	256.915	-2.8	90	-0.02
12 CM	1,1-Dichloroethene	50.000	44.682	10.6#	80	0.00
13 T	Acrolein	250.000	265.598	-6.2	103	0.00
14 T	Allyl chloride	50.000	46.325	7.3	81	0.00
15 T	Acrylonitrile	250.000	248.334	0.7	90	0.00
16 T	Acetone	250.000	201.363	19.5	72	0.00
17 T	Carbon Disulfide	50.000	39.899	20.2	74	0.00
18 T	Methyl Acetate	50.000	48.999	2.0	91	0.00
19 T	Methyl tert-butyl Ether	50.000	45.787	8.4	82	0.00
20 T	Methylene Chloride	50.000	44.688	10.6	85	0.00
21 T	trans-1,2-Dichloroethene	50.000	44.841	10.3	81	0.00
22 T	Diisopropyl ether	50.000	45.830	8.3	82	0.00
23 T	Vinyl Acetate	250.000	240.356	3.9	83	0.00
24 P	1,1-Dichloroethane	50.000	46.016	8.0	82	0.00
25 T	2-Butanone	250.000	232.282	7.1	82	0.00
26 T	2,2-Dichloropropane	50.000	45.940	8.1	79	0.00
27 T	cis-1,2-Dichloroethene	50.000	47.526	4.9	84	0.00
28 T	Bromochloromethane	50.000	51.168	-2.3	90	0.00
29 T	Tetrahydrofuran	250.000	250.922	-0.4	91	0.00
30 C	Chloroform	50.000	45.349	9.3#	81	0.00
31 T	Cyclohexane	50.000	42.828	14.3	78	0.00
32 T	1,1,1-Trichloroethane	50.000	45.826	8.3	80	0.00
33 S	1,2-Dichloroethane-d4	50.000	46.449	7.1	83	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	93	0.00
35 S	Dibromofluoromethane	50.000	49.100	1.8	86	0.00
36 T	1,1-Dichloropropene	50.000	44.524	11.0	80	0.00
37 T	Ethyl Acetate	50.000	49.391	1.2	88	0.00
38 T	Carbon Tetrachloride	50.000	45.371	9.3	78	0.00
39 T	Methylcyclohexane	50.000	44.769	10.5	79	0.00
40 TM	Benzene	50.000	46.236	7.5	83	0.00
41 T	Methacrylonitrile	50.000	50.404	-0.8	89	0.00
42 TM	1,2-Dichloroethane	50.000	44.215	11.6	79	0.00
43 T	Isopropyl Acetate	50.000	48.559	2.9	84	0.00
44 TM	Trichloroethene	50.000	47.264	5.5	83	0.00
45 C	1,2-Dichloropropane	50.000	47.680	4.6#	84	0.00
46 T	Dibromomethane	50.000	46.855	6.3	83	0.00
47 T	Bromodichloromethane	50.000	46.985	6.0	81	0.00
48 T	Methyl methacrylate	50.000	48.972	2.1	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012522\
 Data File : VX026573.D
 Acq On : 25 Jan 2022 16:45
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 26 06:14:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

49 T	1,4-Dioxane	1000.000	1045.624	-4.6	91	-0.01
50 S	Toluene-d8	50.000	48.704	2.6	87	0.00
51 T	4-Methyl-2-Pentanone	250.000	251.318	-0.5	89	0.00
52 CM	Toluene	50.000	47.023	6.0#	84	0.00
53 T	t-1,3-Dichloropropene	50.000	44.227	11.5	83	0.00
54 T	cis-1,3-Dichloropropene	50.000	49.149	1.7	82	0.00
55 T	1,1,2-Trichloroethane	50.000	48.145	3.7	86	0.00
56 T	Ethyl methacrylate	50.000	50.238	-0.5	85	0.00
57 T	1,3-Dichloropropane	50.000	47.817	4.4	85	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	245.817	1.7	86	0.00
59 T	2-Hexanone	250.000	244.027	2.4	84	0.00
60 T	Dibromochloromethane	50.000	49.180	1.6	83	0.00
61 T	1,2-Dibromoethane	50.000	48.635	2.7	86	0.00
62 S	4-Bromofluorobenzene	50.000	49.497	1.0	88	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	93	0.00
64 T	Tetrachloroethene	50.000	47.029	5.9	85	0.00
65 PM	Chlorobenzene	50.000	46.271	7.5	85	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	48.092	3.8	86	0.00
67 C	Ethyl Benzene	50.000	46.694	6.6#	83	0.00
68 T	m/p-Xylenes	100.000	94.358	5.6	84	0.00
69 T	o-Xylene	50.000	47.712	4.6	85	0.00
70 T	Styrene	50.000	48.382	3.2	84	0.00
71 P	Bromoform	50.000	44.993	10.0	86	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	92	0.00
73 T	Isopropylbenzene	50.000	46.161	7.7	83	0.00
74 T	N-amyl acetate	50.000	48.991	2.0	83	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	48.205	3.6	87	0.00
76 T	1,2,3-Trichloropropane	50.000	48.931	2.1	87	0.00
77 T	Bromobenzene	50.000	46.906	6.2	84	0.00
78 T	n-propylbenzene	50.000	46.494	7.0	83	0.00
79 T	2-Chlorotoluene	50.000	45.919	8.2	84	0.00
80 T	1,3,5-Trimethylbenzene	50.000	46.702	6.6	84	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	47.370	5.3	91	0.00
82 T	4-Chlorotoluene	50.000	45.068	9.9	82	0.00
83 T	tert-Butylbenzene	50.000	47.087	5.8	83	0.00
84 T	1,2,4-Trimethylbenzene	50.000	46.125	7.8	82	0.00
85 T	sec-Butylbenzene	50.000	47.816	4.4	84	0.00
86 T	p-Isopropyltoluene	50.000	47.748	4.5	84	0.00
87 T	1,3-Dichlorobenzene	50.000	46.211	7.6	85	0.00
88 T	1,4-Dichlorobenzene	50.000	45.715	8.6	83	0.00
89 T	n-Butylbenzene	50.000	47.696	4.6	83	0.00
90 T	Hexachloroethane	50.000	45.155	9.7	85	0.00
91 T	1,2-Dichlorobenzene	50.000	47.786	4.4	85	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	50.517	-1.0	86	0.00
93 T	1,2,4-Trichlorobenzene	50.000	48.077	3.8	85	0.00
94 T	Hexachlorobutadiene	50.000	46.845	6.3	85	0.00
95 T	Naphthalene	50.000	51.324	-2.6	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012522\
Data File : VX026573.D
Acq On : 25 Jan 2022 16:45
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jan 26 06:14:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	49.837	0.3	89	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6