

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012523\
 Data File : VX033881.D
 Acq On : 25 Jan 2023 11:09
 Operator : JC/MD
 Sample : VX0125WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0125WBL01

Quant Time: Jan 26 05:11:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X011823W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Jan 18 14:40:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.898	128	14751	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.757	114	103300	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	110590	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	34072	29.847	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	99.500%
60) 4-Bromofluorobenzene	11.079	95	54660	27.648	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	92.167%
63) Toluene-d8	8.647	98	105396	29.871	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	99.567%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012523\
Data File : VX033881.D
Acq On : 25 Jan 2023 11:09
Operator : JC/MD
Sample : VX0125WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0125WBL01

Quant Time: Jan 26 05:11:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X011823W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Jan 18 14:40:07 2023
Response via : Initial Calibration

