

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012523\
 Data File : VX033889.D
 Acq On : 25 Jan 2023 15:48
 Operator : JC/MD
 Sample : 01308-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 22-MW-6S

Quant Time: Jan 26 06:01:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	158222	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	228094	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	199034	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93301	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	74572	44.763	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.520%
35) Dibromofluoromethane	5.385	113	68825	48.594	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.180%
50) Toluene-d8	8.647	98	256843	49.771	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.540%
62) 4-Bromofluorobenzene	11.079	95	87183	45.503	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.000%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.374	62	10251	7.648	ug/l	99
27) cis-1,2-Dichloroethene	4.495	96	7709	4.455	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012523\
Data File : VX033889.D
Acq On : 25 Jan 2023 15:48
Operator : JC/MD
Sample : 01308-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
22-MW-6S

Quant Time: Jan 26 06:01:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 11 14:48:42 2023
Response via : Initial Calibration

