

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012523\
 Data File : VX033890.D
 Acq On : 25 Jan 2023 16:11
 Operator : JC/MD
 Sample : 01308-07DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 22-MW-6SDL

Quant Time: Jan 26 06:01:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	167085	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	245551	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	216086	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	103637	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	80807	45.933	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.860%
35) Dibromofluoromethane	5.385	113	74365	48.772	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.540%
50) Toluene-d8	8.647	98	274716	49.450	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.900%
62) 4-Bromofluorobenzene	11.079	95	94468	45.800	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.600%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.374	62	824	0.582	ug/l	98
27) cis-1,2-Dichloroethene	4.489	96	741	0.406	ug/l	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012523\
Data File : VX033890.D
Acq On : 25 Jan 2023 16:11
Operator : JC/MD
Sample : 01308-07DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
22-MW-6SDL

Quant Time: Jan 26 06:01:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 11 14:48:42 2023
Response via : Initial Calibration

