

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012523\
 Data File : VX033891.D
 Acq On : 25 Jan 2023 16:34
 Operator : JC/MD
 Sample : 01308-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 22-MW-2I

Quant Time: Jan 26 06:01:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	158265	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	230309	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	205459	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99751	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	76862	46.125	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.260%
35) Dibromofluoromethane	5.385	113	70572	49.348	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.700%
50) Toluene-d8	8.647	98	261704	50.225	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.460%
62) 4-Bromofluorobenzene	11.079	95	89489	46.257	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.520%
Target Compounds						
					Qvalue	
44) Trichloroethene	7.129	130	41328	24.294	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012523\
Data File : VX033891.D
Acq On : 25 Jan 2023 16:34
Operator : JC/MD
Sample : 01308-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
22-MW-2I

Quant Time: Jan 26 06:01:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 11 14:48:42 2023
Response via : Initial Calibration

