

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012523\
 Data File : VX033893.D
 Acq On : 25 Jan 2023 17:21
 Operator : JC/MD
 Sample : 01308-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LEA-12S

Quant Time: Jan 26 06:02:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	169172	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	243897	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	216707	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	104447	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	80558	45.226	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.460%
35) Dibromofluoromethane	5.385	113	74036	48.886	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.780%
50) Toluene-d8	8.647	98	275635	49.952	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.900%
62) 4-Bromofluorobenzene	11.079	95	96049	46.882	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.760%
Target Compounds						
27) cis-1,2-Dichloroethene	4.489	96	18684	10.099	ug/l	Qvalue 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012523\
Data File : VX033893.D
Acq On : 25 Jan 2023 17:21
Operator : JC/MD
Sample : 01308-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LEA-12S

Quant Time: Jan 26 06:02:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 11 14:48:42 2023
Response via : Initial Calibration

