

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012523\
 Data File : VX033900.D
 Acq On : 25 Jan 2023 20:02
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 26 06:04:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	92	0.00
2 T	Dichlorodifluoromethane	0.395	0.352	10.9	79	0.00
3 P	Chloromethane	0.443	0.425	4.1	78	0.00
4 C	Vinyl Chloride	0.424	0.439	-3.5#	82	0.00
5 T	Bromomethane	0.243	0.207	14.8	81	0.00
6 T	Chloroethane	0.262	0.277	-5.7	93	0.00
7 T	Trichlorofluoromethane	0.687	0.685	0.3	88	0.00
8 T	Diethyl Ether	0.225	0.276	-22.7	122	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.432	0.410	5.1	121	0.00
10 T	Methyl Iodide	0.555	0.595	-7.2	124	0.00
11 T	Tert butyl alcohol	0.091	0.087	4.4	89	-0.06
12 CM	1,1-Dichloroethene	0.407	0.407	0.0	125	0.00
13 T	Acrolein	0.084	0.129	-53.6#	173#	0.00
14 T	Allyl chloride	0.767	0.722	5.9	104	0.00
15 T	Acrylonitrile	0.241	0.243	-0.8	90	0.00
16 T	Acetone	0.232	0.201	13.4	109	-0.02
17 T	Carbon Disulfide	1.066	0.889	16.6	103	0.00
18 T	Methyl Acetate	0.735	0.805	-9.5	100	0.00
19 T	Methyl tert-butyl Ether	1.413	1.413	0.0	90	0.00
20 T	Methylene Chloride	0.512	0.465	9.2	88	0.00
21 T	trans-1,2-Dichloroethene	0.476	0.441	7.4	86	0.00
22 T	Diisopropyl ether	1.544	1.473	4.6	86	0.00
23 T	Vinyl Acetate	1.162	1.094	5.9	82	0.00
24 P	1,1-Dichloroethane	0.845	0.804	4.9	87	0.00
25 T	2-Butanone	0.344	0.323	6.1	86	-0.02
26 T	2,2-Dichloropropane	0.605	0.537	11.2	80	0.00
27 T	cis-1,2-Dichloroethene	0.547	0.532	2.7	88	0.00
28 T	Bromochloromethane	0.385	0.378	1.8	91	0.00
29 T	Tetrahydrofuran	0.217	0.214	1.4	88	-0.01
30 C	Chloroform	0.865	0.830	4.0#	87	0.00
31 T	Cyclohexane	0.775	0.680	12.3	77	0.00
32 T	1,1,1-Trichloroethane	0.739	0.702	5.0	85	0.00
33 S	1,2-Dichloroethane-d4	0.526	0.488	7.2	88	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	90	0.00
35 S	Dibromofluoromethane	0.310	0.321	-3.5	95	0.00
36 T	1,1-Dichloropropene	0.453	0.418	7.7	83	0.00
37 T	Ethyl Acetate	0.496	0.437	11.9	85	0.00
38 T	Carbon Tetrachloride	0.469	0.449	4.3	83	0.00
39 T	Methylcyclohexane	0.560	0.495	11.6	79	0.00
40 TM	Benzene	1.321	1.278	3.3	86	0.00
41 T	Methacrylonitrile	0.250	0.264	-5.6	90	0.00
42 TM	1,2-Dichloroethane	0.468	0.437	6.6	83	0.00
43 T	Isopropyl Acetate	0.715	0.692	3.2	85	0.00
44 TM	Trichloroethene	0.369	0.379	-2.7	91	0.00
45 C	1,2-Dichloropropane	0.331	0.332	-0.3#	88	0.00
46 T	Dibromomethane	0.231	0.232	-0.4	89	0.00
47 T	Bromodichloromethane	0.445	0.452	-1.6	87	0.00
48 T	Methyl methacrylate	0.360	0.361	-0.3	85	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012523\
 Data File : VX033900.D
 Acq On : 25 Jan 2023 20:02
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleID :
 VSTDCCC050

Quant Time: Jan 26 06:04:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.007	12.5	76	-0.06
50 S	Toluene-d8	1.131	1.121	0.9	90	0.00
51 T	4-Methyl-2-Pentanone	0.439	0.450	-2.5	87	0.00
52 CM	Toluene	0.853	0.829	2.8#	86	0.00
53 T	t-1,3-Dichloropropene	0.445	0.456	-2.5	85	0.00
54 T	cis-1,3-Dichloropropene	0.499	0.510	-2.2	85	0.00
55 T	1,1,2-Trichloroethane	0.338	0.349	-3.3	92	0.00
56 T	Ethyl methacrylate	0.475	0.506	-6.5	88	0.00
57 T	1,3-Dichloropropane	0.547	0.567	-3.7	88	0.00
58 T	2-Chloroethyl Vinyl ether	0.234	0.243	-3.8	90	0.00
59 T	2-Hexanone	0.306	0.324	-5.9	86	0.00
60 T	Dibromochloromethane	0.350	0.382	-9.1	89	0.00
61 T	1,2-Dibromoethane	0.337	0.367	-8.9	88	0.00
62 S	4-Bromofluorobenzene	0.420	0.411	2.1	88	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	86	0.00
64 T	Tetrachloroethene	0.381	0.390	-2.4	90	0.00
65 PM	Chlorobenzene	1.035	1.057	-2.1	88	0.00
66 T	1,1,1,2-Tetrachloroethane	0.381	0.410	-7.6	91	0.00
67 C	Ethyl Benzene	1.777	1.797	-1.1#	85	0.00
68 T	m/p-Xylenes	0.695	0.711	-2.3	86	0.00
69 T	o-Xylene	0.689	0.701	-1.7	85	0.00
70 T	Styrene	1.121	1.163	-3.7	84	0.00
71 P	Bromoform	0.319	0.342	-7.2	88	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	84	0.00
73 T	Isopropylbenzene	3.645	3.532	3.1	86	0.00
74 T	N-amyl acetate	1.313	1.261	4.0	80	0.00
75 P	1,1,2,2-Tetrachloroethane	1.097	1.081	1.5	89	0.00
76 T	1,2,3-Trichloropropane	0.919	0.976	-6.2	99	0.00
77 T	Bromobenzene	0.986	0.956	3.0	86	0.00
78 T	n-propylbenzene	4.080	3.906	4.3	83	0.00
79 T	2-Chlorotoluene	2.483	2.336	5.9	84	0.00
80 T	1,3,5-Trimethylbenzene	3.009	2.897	3.7	85	0.00
81 T	trans-1,4-Dichloro-2-butene	0.296	0.297	-0.3	84	0.00
82 T	4-Chlorotoluene	2.818	2.644	6.2	83	0.00
83 T	tert-Butylbenzene	2.987	3.000	-0.4	84	0.00
84 T	1,2,4-Trimethylbenzene	2.846	2.898	-1.8	84	0.00
85 T	sec-Butylbenzene	3.573	3.640	-1.9	83	0.00
86 T	p-Isopropyltoluene	3.059	3.188	-4.2	84	0.00
87 T	1,3-Dichlorobenzene	1.689	1.741	-3.1	86	0.00
88 T	1,4-Dichlorobenzene	1.721	1.742	-1.2	86	0.00
89 T	n-Butylbenzene	2.503	2.531	-1.1	81	0.00
90 T	Hexachloroethane	0.503	0.518	-3.0	82	0.00
91 T	1,2-Dichlorobenzene	1.630	1.690	-3.7	86	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.204	0.213	-4.4	86	0.00
93 T	1,2,4-Trichlorobenzene	1.098	1.196	-8.9	86	0.00
94 T	Hexachlorobutadiene	0.538	0.541	-0.6	83	0.00
95 T	Naphthalene	3.159	3.515	-11.3	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012523\
Data File : VX033900.D
Acq On : 25 Jan 2023 20:02
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jan 26 06:04:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 11 14:48:42 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.099	1.201	-9.3	88	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 5