

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012524\
 Data File : VX039957.D
 Acq On : 25 Jan 2024 16:27
 Operator : JC/MD
 Sample : P1235-06MEDL 4X
 Misc : 5.05g/10mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 137DL

Quant Time: Jan 26 01:36:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	87836	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	159374	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	153318	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	81122	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	71010	48.261	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.520%
35) Dibromofluoromethane	5.391	113	49705	46.304	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.600%
50) Toluene-d8	8.647	98	195721	48.343	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.680%
62) 4-Bromofluorobenzene	11.079	95	78659	49.684	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.360%
Target Compounds						
16) Acetone	2.386	43	86284	104.271	ug/l	92
31) Cyclohexane	5.489	56	5066	2.609	ug/l #	43
52) Toluene	8.714	92	219582	77.701	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012524\
Data File : VX039957.D
Acq On : 25 Jan 2024 16:27
Operator : JC/MD
Sample : P1235-06MEDL 4X
Misc : 5.05g/10mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
137DL

Quant Time: Jan 26 01:36:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

