

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012524\
 Data File : VX039958.D
 Acq On : 25 Jan 2024 16:50
 Operator : JC/MD
 Sample : P1235-07DL 5X
 Misc : 5.07g/10mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 131DL

Manual Integrations
 APPROVED

Reviewed By : John Carlone 01/26/2024
 Supervised By : Semsettin Yesilyurt 01/26/2024

Quant Time: Jan 26 01:36:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	82618	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	151769	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	140080	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	78244	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	65188	47.102	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	94.200%		
35) Dibromofluoromethane	5.385	113	46769	45.752	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	91.500%		
50) Toluene-d8	8.647	98	186004	48.245	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	96.500%		
62) 4-Bromofluorobenzene	11.079	95	78409	52.008	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	104.020%		
Target Compounds						
16) Acetone	2.386	43	2193	2.818	ug/l #	85
31) Cyclohexane	5.501	56	6116m	3.348	ug/l	
39) Methylcyclohexane	7.373	83	3209	1.680	ug/l #	88
52) Toluene	8.720	92	322541	119.853	ug/l	95
67) Ethyl Benzene	10.195	91	5929	1.095	ug/l	96
68) m/p-Xylenes	10.299	106	7894	3.963	ug/l	89
69) o-Xylene	10.640	106	3645	1.848	ug/l	94
80) 1,3,5-Trimethylbenzene	11.451	105	6227	1.177	ug/l	93
84) 1,2,4-Trimethylbenzene	11.750	105	23739	4.497	ug/l	99
89) n-Butylbenzene	12.335	91	5191	1.026	ug/l #	64

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012524\
Data File : VX039958.D
Acq On : 25 Jan 2024 16:50
Operator : JC/MD
Sample : P1235-07DL 5X
Misc : 5.07g/10mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
131DL

Manual Integrations
APPROVED

Reviewed By : John Carlone 01/26/2024
Supervised By : Semsettin Yesilyurt 01/26/2024

Quant Time: Jan 26 01:36:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

