

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012524\
 Data File : VX039962.D
 Acq On : 25 Jan 2024 18:21
 Operator : JC/MD
 Sample : PB158617TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB158617TB

Quant Time: Jan 26 01:38:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	79416	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	142428	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	141356	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	62487	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	63003	47.359	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.720%
35) Dibromofluoromethane	5.391	113	47294	49.300	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.600%
50) Toluene-d8	8.647	98	170333	47.078	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.160%
62) 4-Bromofluorobenzene	11.079	95	66443	46.961	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.920%
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	12188	16.290	ug/l #	86
20) Methylene Chloride	2.788	84	3941	3.482	ug/l	96
25) 2-Butanone	4.568	43	6743	6.272	ug/l	95
37) Ethyl Acetate	4.733	43	2531	1.271	ug/l #	80
43) Isopropyl Acetate	6.342	43	74709	25.566	ug/l	99
52) Toluene	8.720	92	5087	2.014	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012524\
Data File : VX039962.D
Acq On : 25 Jan 2024 18:21
Operator : JC/MD
Sample : PB158617TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB158617TB

Quant Time: Jan 26 01:38:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

