

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012524\
 Data File : VX039963.D
 Acq On : 25 Jan 2024 18:44
 Operator : JC/MD
 Sample : PB158617ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB158617ZHE#01

Quant Time: Jan 26 01:39:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	79621	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	146272	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	143406	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	67396	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	63887	47.900	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.800%
35) Dibromofluoromethane	5.385	113	46873	47.577	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.160%
50) Toluene-d8	8.647	98	174064	46.845	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.680%
62) 4-Bromofluorobenzene	11.079	95	74311	51.142	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.280%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	14026	18.699	ug/l	91
20) Methylene Chloride	2.794	84	4518	3.982	ug/l	94
25) 2-Butanone	4.574	43	6927	6.426	ug/l	98
37) Ethyl Acetate	4.733	43	2472	1.208	ug/l #	75
43) Isopropyl Acetate	6.342	43	74662	24.878	ug/l	99
52) Toluene	8.726	92	2930	1.130	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012524\
Data File : VX039963.D
Acq On : 25 Jan 2024 18:44
Operator : JC/MD
Sample : PB158617ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB158617ZHE#01

Quant Time: Jan 26 01:39:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

