

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012524\
 Data File : VX039965.D
 Acq On : 25 Jan 2024 19:30
 Operator : JC/MD
 Sample : PB158617ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB158617ZHE#03

Quant Time: Jan 26 01:40:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	78862	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	148160	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	147550	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	70223	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	60380	45.706	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.420%
35) Dibromofluoromethane	5.385	113	47954	48.054	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.100%
50) Toluene-d8	8.647	98	188356	50.045	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.100%
62) 4-Bromofluorobenzene	11.079	95	71063	48.283	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.560%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.953	59	1775	5.162	ug/l #	93
16) Acetone	2.379	43	12198	16.418	ug/l	95
20) Methylene Chloride	2.788	84	3434	3.056	ug/l	92
25) 2-Butanone	4.568	43	8494	7.956	ug/l	94
43) Isopropyl Acetate	6.336	43	65224	21.456	ug/l	99
52) Toluene	8.720	92	4269	1.625	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012524\
Data File : VX039965.D
Acq On : 25 Jan 2024 19:30
Operator : JC/MD
Sample : PB158617ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB158617ZHE#03

Quant Time: Jan 26 01:40:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

