

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012524\
 Data File : VX039966.D
 Acq On : 25 Jan 2024 19:53
 Operator : JC/MD
 Sample : PB158617ZHE#05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB158617ZHE#05

Quant Time: Jan 26 01:40:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	76239	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	150191	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	150920	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	70758	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	60354	47.258	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.520%
35) Dibromofluoromethane	5.391	113	45822	45.296	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.600%
50) Toluene-d8	8.647	98	184752	48.424	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.840%
62) 4-Bromofluorobenzene	11.079	95	70361	47.160	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.320%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.953	59	1666	5.011	ug/l #	90
16) Acetone	2.379	43	14477	20.156	ug/l	92
20) Methylene Chloride	2.788	84	4478	4.122	ug/l	96
25) 2-Butanone	4.568	43	8539	8.273	ug/l	92
43) Isopropyl Acetate	6.342	43	67902	22.035	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012524\
Data File : VX039966.D
Acq On : 25 Jan 2024 19:53
Operator : JC/MD
Sample : PB158617ZHE#05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB158617ZHE#05

Quant Time: Jan 26 01:40:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

