

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012524\
 Data File : VX039967.D
 Acq On : 25 Jan 2024 20:16
 Operator : JC/MD
 Sample : PB158617ZHE#06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB158617ZHE#06

Quant Time: Jan 26 01:41:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	77263	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	145756	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	145766	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	71320	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	61162	47.256	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	94.520%	
35) Dibromofluoromethane	5.391	113	46972	47.846	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	95.700%	
50) Toluene-d8	8.647	98	187474	50.632	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	101.260%	
62) 4-Bromofluorobenzene	11.079	95	73088	50.478	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	100.960%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.947	59	1702	5.052	ug/l	96
16) Acetone	2.374	43	12707	17.457	ug/l	97
20) Methylene Chloride	2.788	84	3588	3.259	ug/l	94
25) 2-Butanone	4.562	43	8415	8.045	ug/l	91
43) Isopropyl Acetate	6.336	43	64792	21.666	ug/l	99
52) Toluene	8.720	92	2630	1.018	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX012524\
Data File : VX039967.D
Acq On : 25 Jan 2024 20:16
Operator : JC/MD
Sample : PB158617ZHE#06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB158617ZHE#06

Quant Time: Jan 26 01:41:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

