

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX012619\
 Data File : VX007251.D
 Acq On : 26 Jan 2019 01:08
 Operator : JC/SP
 Sample : K1238-20
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Jan 28 00:21:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	526512	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	820012	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	789211	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	374675	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	285689	51.92	ug/l	0.00
Spiked Amount			Recovery	=	103.84%	
35) Dibromofluoromethane	5.51	113	252725	48.13	ug/l	0.00
Spiked Amount			Recovery	=	96.26%	
50) Toluene-d8	8.71	98	1004983	52.28	ug/l	0.00
Spiked Amount			Recovery	=	104.56%	
62) 4-Bromofluorobenzene	11.13	95	347786	51.90	ug/l	0.00
Spiked Amount			Recovery	=	103.80%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	3.07	59	2799	2.156	ug/l	98
13) Acrolein	2.29	56	352	3.705	ug/l	88
16) Acetone	2.45	43	64776	23.812	ug/l	98
18) Methyl Acetate	2.78	43	25838	3.280	ug/l	94
20) Methylene Chloride	2.86	84	27889	4.880	ug/l	97
25) 2-Butanone	4.70	43	14481	3.720	ug/l	90
29) Tetrahydrofuran	5.16	42	16112	6.370	ug/l	95
30) Chloroform	5.21	83	1882	0.206	ug/l	99
36) 1,1-Dichloropropene	5.67	75	33657	4.721	ug/l #	49
37) Ethyl Acetate	4.70	43	14481	1.895	ug/l #	68
38) Carbon Tetrachloride	5.67	117	46435	6.474	ug/l #	15
43) Isopropyl Acetate	6.46	43	27367	2.270	ug/l	96
48) Methyl methacrylate	7.68	41	12383	2.000	ug/l #	12
49) 1,4-Dioxane	7.77	88	127	Below Cal	#	1
51) 4-Methyl-2-Pentanone	8.62	43	136852	17.663	ug/l #	50
54) cis-1,3-Dichloropropene	8.62	75	66223	8.266	ug/l #	63
57) 1,3-Dichloropropane	9.54	76	16022	1.748	ug/l #	46
59) 2-Hexanone	9.54	43	216198	36.616	ug/l #	51
74) N-amyl acetate	10.69	43	21584	2.029	ug/l #	54
76) 1,2,3-Trichloropropane	11.13	75	169375	22.308	ug/l #	40
80) 1,3,5-Trimethylbenzene	11.51	105	7893	0.358	ug/l	92
81) trans-1,4-Dichloro-2-buten	11.13	75	169375	66.011	ug/l #	10
84) 1,2,4-Trimethylbenzene	11.80	105	8124	0.366	ug/l	100
88) 1,4-Dichlorobenzene	12.08	146	727	0.203	ug/l #	1
90) Hexachloroethane	12.58	117	500	3.034	ug/l #	6
95) Naphthalene	13.83	128	26811	0.993	ug/l	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX012619\
Data File : VX007251.D
Acq On : 26 Jan 2019 01:08
Operator : JC/SP
Sample : K1238-20
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Jan 28 00:21:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.67 min Scan# 740

Delta R.T. -0.00 min

Lab File: VX007251.D

Acq: 26 Jan 2019 01:08

Instrument :

MSVOA_X


ClientSampleId :

Tot Ion:168 Resp: 526512

Ion Ratio Lower Upper

168 100

99 51.0 41.8 62.8



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.67

200000



150000

100000

50000

0

Time-->



#11

Tert butyl alcohol

Concen: 2.156 ug/l

RT: 3.07 min Scan# 314

Delta R.T. -0.01 min

Lab File: VX007251.D


Acq: 26 Jan 2019 01:08

Tgt Ion: 59 Resp: 2799

Ion Ratio Lower Upper

59 100

57 9.9 8.5 12.7



Abundance Ion 59.00 (58.70 to 59.70): VX0

Ion 57.00 (56.70 to 57.70): VX0

1200

1000

800


600


400

200

0

Time-->









#13
Acrolein
Concen: 3.705 ug/l
RT: 2.29 min Scan# 187
Delta R.T. -0.01 min
Lab File: VX007251.D
Acq: 26 Jan 2019 01:08




Instrument :
MSVOA_X
ClientSampleId :


Tot Ion: 56 Resp: 352
Ion Ratio Lower Upper
56 100
55 83.0 58.2 87.4



#16
Acetone
Concen: 23.812 ug/l
RT: 2.45 min Scan# 212
Delta R.T. -0.00 min
Lab File: VX007251.D
Acq: 26 Jan 2019 01:08

Tgt Ion: 43 Resp: 64776
Ion Ratio Lower Upper
43 100
58 30.2 25.0 37.6





#18

Methyl Acetate

Concen: 3.280 ug/l

RT: 2.78 min Scan# 267

Delta R.T. -0.00 min

Lab File: VX007251.D

Acq: 26 Jan 2019 01:08

Instrument :

MSVOA_X





ClientSampleId :

Tot Ion: 43 Resp: 25838

Ion Ratio Lower Upper

43 100

74 26.0 18.6 28.0



#20

Methylene Chloride

Concen: 4.880 ug/l

RT: 2.86 min Scan# 280

Delta R.T. 0.00 min

Lab File: VX007251.D

Acq: 26 Jan 2019 01:08

Tgt Ion: 84 Resp: 27889




Ion Ratio Lower Upper


84 100

49 115.1 91.8 137.6

51 30.9 28.5 42.7

86 59.0 50.2 75.2





#25

2-Butanone

Concen: 3.720 ug/l

RT: 4.70 min Scan# 582

Delta R.T. -0.00 min

Lab File: VX007251.D

Acq: 26 Jan 2019 01:08





Instrument :
MSVOA_X
ClientSampled :

Tot Ion: 43 Resp: 14481

Ion Ratio Lower Upper

43 100

72 22.5 22.4 33.6



#29

Tetrahydrofuran

Concen: 6.370 ug/l

RT: 5.16 min Scan# 657

Delta R.T. 0.01 min

Lab File: VX007251.D

Acq: 26 Jan 2019 01:08




Tgt Ion: 42 Resp: 16112


Ion Ratio Lower Upper

42 100

72 45.1 38.9 58.3

71 48.7 36.2 54.4





#30

Chloroform

Concen: 0.206 ug/l

RT: 5.21 min Scan# 666

Delta R.T. -0.00 min

Lab File: VX007251.D

Acq: 26 Jan 2019 01:08

Instrument :

MSVOA_X





ClientSampleId :

Tot Ion: 83 Resp: 1882

Ion Ratio Lower Upper

83 100

85 63.3 51.2 76.8



#33

1,2-Dichloroethane-d4

Concen: 51.920 ug/l

RT: 6.07 min Scan# 806

Delta R.T. -0.00 min

Lab File: VX007251.D




Acq: 26 Jan 2019 01:08


Tgt Ion: 65 Resp: 285689

Ion Ratio Lower Upper

65 100

67 53.6 0.0 105.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.86 min Scan# 936

Delta R.T. -0.00 min

Lab File: VX007251.D

Acq: 26 Jan 2019 01:08

Instrument :

MSVOA_X

ClientSampleId :


Tot Ion:114 Resp: 820012

Ion Ratio Lower Upper

114 100

63 18.4 0.0 34.8




88 14.4 0.0 28.8



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 48.126 ug/l

RT: 5.51 min Scan# 714

Delta R.T. 0.01 min

Lab File: VX007251.D

Acq: 26 Jan 2019 01:08


Tgt Ion:113 Resp: 252725

Ion Ratio Lower Upper

113 100

111 102.3 81.6 122.4



192 22.3 16.7 25.1




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V









#36
 1,1-Dichloropropene
 Concen: 4.721 ug/l
 RT: 5.67 min Scan# 740
 Delta R.T. -0.13 min
 Lab File: VX007251.D
 Acq: 26 Jan 2019 01:08




Instrument :
 MSVOA_X
 ClientSampleId :


Tot Ion	Ratio	Lower	Upper
75	100		
110	10.6	20.2	60.5#
77	0.1	24.8	37.2#



#37
 Ethyl Acetate
 Concen: 1.895 ug/l
 RT: 4.70 min Scan# 582
 Delta R.T. -0.15 min
 Lab File: VX007251.D
 Acq: 26 Jan 2019 01:08

Tgt Ion	Ratio	Lower	Upper
43	100		
61	0.5	11.9	17.9#
70	1.2	9.7	14.5#





#38

Carbon Tetrachloride

Concen: 6.474 ug/l

RT: 5.67 min Scan# 740

Delta R.T. -0.12 min

Lab File: VX007251.D

Acq: 26 Jan 2019 01:08

Instrument :

MSVOA_X

ClientSampleId :


Tot Ion:117 Resp: 46435

Ion Ratio Lower Upper

117 100

119 5.9 79.4 119.2#




121 0.0 24.0 36.0#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V



#43

Isopropyl Acetate

Concen: 2.270 ug/l

RT: 6.46 min Scan# 871

Delta R.T. 0.01 min

Lab File: VX007251.D

Acq: 26 Jan 2019 01:08


Tgt Ion: 43 Resp: 27367

Ion Ratio Lower Upper

43 100

61 23.0 16.8 25.2



87 17.3 12.4 18.6




Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 61.00 (60.70 to 61.70): VX0

Ion 87.00 (86.70 to 87.70): VX0









#48
Methvl methacrylate
Concen: 2.000 ug/l
RT: 7.68 min Scan# 1071
Delta R.T. -0.09 min
Lab File: VX007251.D
Acq: 26 Jan 2019 01:08




Instrument :
MSVOA_X
ClientSampleId :


Tot Ion: 41 Resp: 12383
Ion Ratio Lower Upper
41 100
69 0.8 72.4 108.6#
39 0.0 45.8 68.8#



#49
1,4-Dioxane
Concen: Below Cal
RT: 7.77 min Scan# 1086
Delta R.T. 0.02 min
Lab File: VX007251.D
Acq: 26 Jan 2019 01:08

Tgt Ion: 88 Resp: 127
Ion Ratio Lower Upper
88 100
43 0.0 22.2 33.4#
58 607.1 51.0 76.4#





#50

Toluene-d8

Concen: 52.285 ug/l

RT: 8.71 min Scan# 1240

Delta R.T. -0.00 min

Lab File: VX007251.D

Acq: 26 Jan 2019 01:08

Instrument :

MSVOA_X


ClientSampleId :

Tot Ion: 98 Resp: 1004983

Ion Ratio Lower Upper

98 100

100 64.6 52.0 78.0



Abundance Ion 98.00 (97.70 to 98.70): VX006951.D (-1233) (-)

Ion 100.00 (99.70 to 100.70): VX006951.D (-1233) (-)

8.71

600000



400000

200000

0

8.60 8.70 8.80 8.90

Time-->



#51

4-Methyl-2-Pentanone

Concen: 17.663 ug/l

RT: 8.62 min Scan# 1224

Delta R.T. -0.03 min

Lab File: VX007251.D


Acq: 26 Jan 2019 01:08

Tgt Ion: 43 Resp: 136852

Ion Ratio Lower Upper

43 100

58 9.2 31.8 47.6#



Abundance Ion 43.00 (42.70 to 43.70): VX006951.D (-1218) (-)

Ion 58.00 (57.70 to 58.70): VX006951.D (-1218) (-)

8.62

100000

80000

60000


40000


20000

0

8.50 8.55 8.60 8.65 8.70

Time-->





#54

cis-1,3-Dichloropropene

Concen: 8.266 ug/l

RT: 8.62 min Scan# 1224

Delta R.T. 0.18 min

Lab File: VX007251.D

Acq: 26 Jan 2019 01:08

Instrument :
MSVOA_X
ClientSampleId :





Tot Ion: 75 Resp: 66223

Ion Ratio Lower Upper

75 100

77 0.8 26.2 39.2#

39 31.3 36.7 55.1#



#57

1,3-Dichloropropane

Concen: 1.748 ug/l

RT: 9.54 min Scan# 1376

Delta R.T. 0.17 min

Lab File: VX007251.D




Acq: 26 Jan 2019 01:08


Tgt Ion: 76 Resp: 16022

Ion Ratio Lower Upper

76 100

78 2.4 26.0 39.0#









#59
2-Hexanone
Concen: 36.616 ug/l
RT: 9.54 min Scan# 1376
Delta R.T. 0.05 min
Lab File: VX007251.D
Acq: 26 Jan 2019 01:08




Instrument :
MSVOA_X
ClientSampleId :


Tot Ion: 43 Resp: 216198
Ion Ratio Lower Upper
43 100
58 19.8 27.7 83.1#



#62
4-Bromofluorobenzene
Concen: 51.900 ug/l
RT: 11.13 min Scan# 1637
Delta R.T. -0.00 min
Lab File: VX007251.D
Acq: 26 Jan 2019 01:08

Tgt Ion: 95 Resp: 347786
Ion Ratio Lower Upper
95 100
174 93.8 0.0 182.2
176 91.7 0.0 178.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1469
Delta R.T. -0.01 min
Lab File: VX007251.D
Acq: 26 Jan 2019 01:08

Instrument :
MSVOA_X
ClientSampleId :

Tot Ion:117 Resp: 789211
Ion Ratio Lower Upper
117 100
82 52.4 41.0 61.6
119 32.3 25.4 38.0




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V






#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.08 min Scan# 1792
Delta R.T. -0.00 min
Lab File: VX007251.D
Acq: 26 Jan 2019 01:08

Tgt Ion:152 Resp: 374675
Ion Ratio Lower Upper
152 100
115 54.8 39.1 117.2
150 157.6 0.0 344.8



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V





#74

N-amvl acetate

Concen: 2.029 ug/l

RT: 10.69 min Scan# 1565

Delta R.T. -0.20 min

Lab File: VX007251.D

Acq: 26 Jan 2019 01:08

Instrument :

MSVOA_X

ClientSampleId :

Tot Ion: 43 Resp: 21584


Ion Ratio Lower Upper

43 100

70 3.9 35.8 53.8#

55 5.2 19.4 29.0#

61 13.2 20.1 30.1#






Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 70.00 (69.70 to 70.70): VX0

Ion 55.00 (54.70 to 55.70): VX0

Ion 61.00 (60.70 to 61.70): VX0



#76

1,2,3-Trichloropropane

Concen: 22.308 ug/l

RT: 11.13 min Scan# 1637

Delta R.T. -0.16 min

Lab File: VX007251.D


Acq: 26 Jan 2019 01:08

Tgt Ion: 75 Resp: 169375

Ion Ratio Lower Upper



75 100


77 1.3 18.5 55.5#



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 77.00 (76.70 to 77.70): VX0





#80

1,3,5-Trimethylbenzene

Concen: 0.358 ug/l

RT: 11.51 min Scan# 1698

Delta R.T. -0.00 min

Lab File: VX007251.D

Acq: 26 Jan 2019 01:08

Instrument :

MSVOA_X


ClientSampleId :

Tot Ion:105 Resp: 7893

Ion Ratio Lower Upper




105 100

120 44.5 25.1 75.3



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V



#81

trans-1,4-Dichloro-2-butene

Concen: 66.011 ug/l

RT: 11.13 min Scan# 1637

Delta R.T. 0.06 min

Lab File: VX007251.D

Acq: 26 Jan 2019 01:08


Tgt Ion: 75 Resp: 169375

Ion Ratio Lower Upper

75 100

53 0.0 79.7 119.5#



89 0.0 38.0 57.0#




Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 53.00 (52.70 to 53.70): VX0

Ion 88.90 (88.60 to 89.60): VX0









#84
 1,2,4-Trimethylbenzene
 Concen: 0.366 ug/l
 RT: 11.80 min Scan# 1747
 Delta R.T. -0.00 min
 Lab File: VX007251.D
 Acq: 26 Jan 2019 01:08




Instrument :
 MSVOA_X
 ClientSampleId :


Tot Ion:105 Resp: 8124
 Ion Ratio Lower Upper
 105 100
 120 46.3 23.2 69.6



#88
 1,4-Dichlorobenzene
 Concen: 0.203 ug/l
 RT: 12.08 min Scan# 1793
 Delta R.T. -0.01 min
 Lab File: VX007251.D
 Acq: 26 Jan 2019 01:08

Tgt Ion:146 Resp: 727
 Ion Ratio Lower Upper
 146 100
 111 655.0 18.1 54.1#
 148 278.4 31.9 95.9#





#90

Hexachloroethane

Concen: 3.034 ug/l

RT: 12.58 min Scan# 1874

Delta R.T. -0.02 min

Lab File: VX007251.D

Acq: 26 Jan 2019 01:08

Instrument :

MSVOA_X


ClientSampled :

Tot Ion:117 Resp: 500

Ion Ratio Lower Upper




117 100

201 0.0 42.9 128.7#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 200.70 (200.40 to 201.40): V



#95

Naphthalene

Concen: 0.993 ug/l

RT: 13.83 min Scan# 2080

Delta R.T. -0.00 min

Lab File: VX007251.D

Acq: 26 Jan 2019 01:08


Tgt Ion:128 Resp: 26811

Ion Ratio Lower Upper

128 100

127 13.7 10.1 15.1

129 12.8 8.6 13.0



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V

