

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX012619\
 Data File : VX007231.D
 Acq On : 25 Jan 2019 16:15
 Operator : JC/SP
 Sample : VX0126WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0126WBL01

Quant Time: Jan 28 00:31:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	551207	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	845762	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	780587	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	378735	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	295150	51.24	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.48%	
35) Dibromofluoromethane	5.50	113	264422	48.82	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.64%	
50) Toluene-d8	8.71	98	995776	50.23	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.46%	
62) 4-Bromofluorobenzene	11.13	95	351390	50.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.68%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX012619\
Data File : VX007231.D
Acq On : 25 Jan 2019 16:15
Operator : JC/SP
Sample : VX0126WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0126WBL01

Quant Time: Jan 28 00:31:03 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

